

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
 Data File : BN009802.D
 Acq On : 21 Feb 2020 05:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 06:58:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	176#	-0.02
2	1,4-Dioxane	0.410	0.393	4.1	170#	-0.02
3	n-Nitrosodimethylamine	0.673	0.642	4.6	191#	-0.03
4 S	2-Fluorophenol	0.889	1.028	-15.6	211#	-0.02
5 S	Phenol-d6	1.082	1.221	-12.8	209#	-0.02
6	bis(2-Chloroethyl)ether	1.052	1.017	3.3	177#	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	187#	-0.03
8 S	Nitrobenzene-d5	0.333	0.339	-1.8	202#	-0.04
9	Naphthalene	1.091	1.067	2.2	188#	-0.03
10	Hexachlorobutadiene	0.240	0.228	5.0	176#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.774	0.733	5.3	183#	-0.03
12	2-Methylnaphthalene	0.748	0.725	3.1	188#	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	168#	-0.02
14 S	2,4,6-Tribromophenol	0.154	0.174	-13.0	212#	-0.02
15 S	2-Fluorobiphenyl	1.510	1.529	-1.3	177#	-0.03
16	Acenaphthylene	1.705	3.199	-87.6#	350#	-0.02
17	Acenaphthene	1.199	1.357	-13.2	198#	-0.02
18	Fluorene	1.582	1.563	1.2	171#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	150	-0.02
20	4-Bromophenyl-phenylether	0.279	0.154	44.8#	87	-0.01
21	Hexachlorobenzene	0.249	0.256	-2.8	162#	-0.02
22	Pentachlorophenol	0.064	0.082	-28.1#	258#	-0.04
23	Phenanthrene	1.190	1.133	4.8	154#	-0.02
24	Anthracene	1.006	1.067	-6.1	177#	-0.02
25 SURR	Fluoranthene-d10	1.384	1.296	6.4	150#	-0.02
26	Fluoranthene	1.425	1.367	4.1	155#	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	167#	-0.02
28	Pvrene	1.519	1.485	2.2	168#	-0.02
29 S	Terphenyl-d14	0.956	1.245	-30.2#	223#	-0.02
30	Benzo(a)anthracene	1.327	1.394	-5.0	197#	-0.03
31	Chrysene	1.536	1.345	12.4	150	-0.02
32	Bis(2-ethylhexyl)phthalate	0.648	0.866	-33.6#	248#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.599	1.476	7.7	154#	-0.04
34 I	Perylene-d12	1.000	1.000	0.0	164#	-0.03
35	Benzo(b)fluoranthene	1.462	1.344	8.1	165#	-0.03
36	Benzo(k)fluoranthene	1.592	1.363	14.4	148	-0.03
37 C	Benzo(a)pyrene	1.334	1.304	2.2	165#	-0.03
38	Dibenzo(a,h)anthracene	1.314	1.144	12.9	152#	-0.04
39	Benzo(a,h,i)perylene	1.400	1.298	7.3	162#	-0.05

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
Data File : BN009802.D
Acq On : 21 Feb 2020 05:09
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 21 06:58:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			