

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
 Data File : BN009802.D
 Acq On : 21 Feb 2020 05:09
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 06:58:43 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	176	-0.02
2	1,4-Dioxane	0.400	0.383	4.3	170	-0.02
3	n-Nitrosodimethylamine	0.400	0.382	4.5	191	-0.03
4 S	2-Fluorophenol	0.400	0.463	-15.8	211	-0.02
5 S	Phenol-d6	0.400	0.451	-12.7	209	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.386	3.5	177	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	187	-0.03
8 S	Nitrobenzene-d5	0.400	0.407	-1.7	202	-0.04
9	Naphthalene	0.400	0.391	2.3	188	-0.03
10	Hexachlorobutadiene	0.400	0.379	5.3	176	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	183	-0.03
12	2-Methylnaphthalene	0.400	0.388	3.0	188	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	168	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.452	-13.0	212	-0.02
15 S	2-Fluorobiphenyl	0.400	0.405	-1.3	177	-0.03
16	Acenaphthylene	0.400	0.750	-87.5#	350	-0.02
17	Acenaphthene	0.400	0.453	-13.2	198	-0.02
18	Fluorene	0.400	0.395	1.3	171	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	150	-0.02
20	4-Bromophenyl-phenylether	0.400	0.158	60.5#	87	-0.01
21	Hexachlorobenzene	0.400	0.412	-3.0	162	-0.02
22	Pentachlorophenol	0.400	0.571	-42.7#	258	-0.04
23	Phenanthrene	0.400	0.381	4.8	154	-0.02
24	Anthracene	0.400	0.424	-6.0	177	-0.02
25 SURR	Fluoranthene-d10	0.400	0.375	6.3	150	-0.02
26	Fluoranthene	0.400	0.384	4.0	155	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	167	-0.02
28	Pvrene	0.400	0.391	2.3	168	-0.02
29 S	Terphenyl-d14	0.400	0.521	-30.3#	223	-0.02
30	Benzo(a)anthracene	0.400	0.420	-5.0	197	-0.03
31	Chrysene	0.400	0.350	12.5	150	-0.02
32	Bis(2-ethylhexyl)phthalate	0.400	0.535	-33.8#	248	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	154	-0.04
34 I	Perylene-d12	0.400	0.400	0.0	164	-0.03
35	Benzo(b)fluoranthene	0.400	0.368	8.0	165	-0.03
36	Benzo(k)fluoranthene	0.400	0.342	14.5	148	-0.03
37 C	Benzo(a)pyrene	0.400	0.391	2.3	165	-0.03
38	Dibenzo(a,h)anthracene	0.400	0.348	13.0	152	-0.04
39	Benzo(a,h,i)perylene	0.400	0.371	7.3	162	-0.05

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
Data File : BN009802.D
Acq On : 21 Feb 2020 05:09
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 21 06:58:43 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				