

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
 Data File : BN009800.D
 Acq On : 21 Feb 2020 03:57
 Operator : CG/JU
 Sample : L1482-04MSD 5X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TP6-EP2-2MSD

Manual Integrations
 APPROVED

mohammad
 2/21/2020 2:19:59 PM

Quant Time: Feb 21 04:29:26 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.41	152	1779	0.40	ng	0.00
7) Naphthalene-d8	10.15	136	6346	0.40	ng	-0.03
13) Acenaphthene-d10	14.03	164	3747	0.40	ng	-0.02
19) Phenanthrene-d10	16.79	188	7572	0.40	ng	-0.02
27) Chrysene-d12	21.00	240	8031	0.40	ng	-0.03
34) Perylene-d12	23.11	264	8314	0.40	ng	-0.03

System Monitoring Compounds						
4) 2-Fluorophenol	5.07	112	332	0.08	ng	-0.02
5) Phenol-d6	6.62	99	381	0.08	ng	-0.02
8) Nitrobenzene-d5	8.57	82	396	0.07	ng	-0.03
11) 2-Methylnaphthalene-d10	11.75	152	920	0.07	ng	-0.04
14) 2,4,6-Tribromophenol	15.55	330	127	0.09	ng	-0.03
15) 2-Fluorobiphenyl	12.65	172	1168	0.08	ng	-0.03
25) Fluoranthene-d10	18.83	212	1403	0.05	ng	-0.03
29) Terphenyl-d14	19.45	244	4316	0.22	ng	-0.03

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.09	88	118	0.065	ng	# 63
3) n-Nitrosodimethylamine	3.42	42	215	0.072	ng	# 17
6) bis(2-Chloroethyl)ether	6.86	93	401	0.086	ng	90
9) Naphthalene	10.19	128	154227	8.911	ng	# 94
10) Hexachlorobutadiene	10.48	225	245	0.064	ng	# 95
12) 2-Methylnaphthalene	11.82	142	9639	0.812	ng	98
16) Acenaphthylene	13.76	152	106708	6.680	ng	99
17) Acenaphthene	14.10	154	15301	1.362	ng	98
18) Fluorene	15.09	166	28216	1.904	ng	95
21) Hexachlorobenzene	16.11	284	337	0.071	ng	# 91
22) Pentachlorophenol	16.48	266	258	0.342	ng	96
23) Phenanthrene	16.84	178	296385	13.152	ng	99
24) Anthracene	16.92	178	96406	5.061	ng	99
26) Fluoranthene	18.86	202	516965	19.159	ng	99
28) Pyrene	19.23	202	471450	15.459	ng	98
30) Benzo(a)anthracene	20.99	228	317765	11.923	ng	93
31) Chrysene	21.05	228	256176	8.308	ng	97
32) Bis(2-ethylhexyl)phthalate	20.95	149	1786	0.137	ng	# 99
33) Indeno(1,2,3-cd)pyrene	25.15	276	203450	6.336	ng	99
35) Benzo(b)fluoranthene	22.50	252	370676	12.196	ng	# 86
36) Benzo(k)fluoranthene	22.53	252	137163m	4.145	ng	
37) Benzo(a)pyrene	23.02	252	303208	10.933	ng	# 80
38) Dibenzo(a,h)anthracene	25.15	278	58243	2.133	ng	97
39) Benzo(g,h,i)perylene	25.77	276	205463	7.063	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022020\
Data File : BN009800.D
Acq On : 21 Feb 2020 03:57
Operator : CG/JU
Sample : L1482-04MSD 5X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM

Quant Time: Feb 21 04:29:26 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.41 min Scan# 569
Delta R.T. -0.01 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD




Tot Ion	Ratio	Lower	Upper
152	100		
150	148.9	120.6	181.0
115	60.9	51.2	76.8

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM




Abundance Ion 152.00 (151.70 to 152.70): E
2000 Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2

1,4-Dioxane
Concen: 0.065 ng
RT: 3.09 min Scan# 32
Delta R.T. -0.01 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
88	100		
58	100.8	97.8	146.8
43	0.0	45.1	67.7#



Abundance Ion 88.00 (87.70 to 88.70): BNC
1200 Ion 58.00 (57.70 to 58.70): BNC
Ion 43.00 (42.70 to 43.70): BNC





#3





n-Nitrosodimethylamine
 Concen: 0.072 ng
 RT: 3.42 min Scan# 73
 Delta R.T. -0.02 min
 Lab File: BN009800.D
 Acq: 21 Feb 2020 03:57

Instrument :
 BNA_N
 ClientSampleId :
 TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
42	100		
74	0.0	51.2	76.8#
44	35.3	2.6	4.0#

Manual Integrations
 APPROVED




mohammad
 2/21/2020 2:19:59 PM



#4

2-Fluorophenol
 Concen: 0.084 ng
 RT: 5.07 min Scan# 278
 Delta R.T. -0.02 min
 Lab File: BN009800.D
 Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
112	100		
64	87.3	65.8	98.6
63	36.4	38.7	58.1#





#5

Phenol-d6

Concen: 0.079 ng

RT: 6.62 min Scan# 471

Delta R.T. -0.02 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Instrument :

BNA_N





ClientSampleId :

TP6-EP2-2MSD

Tot Ion:	99	Resp:	381
Ion	Ratio	Lower	Upper
99	100		
42	34.1	30.8	46.2
71	40.7	34.9	52.3

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.086 ng




RT: 6.86 min Scan# 501


Delta R.T. -0.02 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Tgt Ion:	93	Resp:	401
Ion	Ratio	Lower	Upper
93	100		
63	80.5	74.0	111.0
95	31.2	26.7	40.1





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.15 min Scan# 826

Delta R.T. -0.03 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Instrument :

BNA_N


ClientSampleId :

TP6-EP2-2MSD

Tot Ion	136	Resp:	6346
Ion Ratio	Lower	Upper	
136	100		
137	11.7	11.3	16.9
54	10.4	13.4	20.2#
68	5.8	7.0	10.4#

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM




Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E


Ion 54.00 (53.70 to 54.70): BNC

Ion 68.00 (67.70 to 68.70): BNC

Time-->



Time-->



#8

Nitrobenzene-d5

Concen: 0.075 ng


RT: 8.57 min Scan# 701

Delta R.T. -0.03 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Tgt Ion	82	Resp:	396
Ion Ratio	Lower	Upper	
82	100		
128	57.1	32.5	48.7#
54	86.7	73.5	110.3




Abundance Ion 82.00 (81.70 to 82.70): BNC


Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BNC

Time-->



Time-->



#9

Naphthalene

Concen: 8.911 ng

RT: 10.19 min Scan# 829

Delta R.T. -0.04 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Instrument :

BNA_N





ClientSampleId :

TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
128	100		
129	11.0	11.1	16.7#
127	13.4	12.6	19.0

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



#10

Hexachlorobutadiene

Concen: 0.064 ng




RT: 10.48 min Scan# 852

Delta R.T. -0.01 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.4	53.4	80.2









#11
2-Methylnaphthalene-d10
Concen: 0.075 ng
RT: 11.75 min Scan# 987
Delta R.T. -0.04 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion:152 Resp: 920
Ion Ratio Lower Upper
152 100
151 14.8 15.1 22.7#




Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



#12
2-Methylnaphthalene
Concen: 0.812 ng
RT: 11.82 min Scan# 1004
Delta R.T. -0.04 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion:142 Resp: 9639
Ion Ratio Lower Upper
142 100
141 87.9 69.8 104.6
115 39.5 34.6 52.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.03 min Scan# 1295
Delta R.T. -0.02 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
164	100		
162	101.4	79.9	119.9
160	47.5	38.2	57.2




Manual Integrations
APPROVED



mohammad
2/21/2020 2:19:59 PM



#14
2,4,6-Tribromophenol
Concen: 0.088 ng
RT: 15.55 min Scan# 1431
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	80.3	36.8	55.2









#15
2-Fluorobiphenyl
Concen: 0.083 ng
RT: 12.65 min Scan# 1171
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tot Ion	Ratio	Lower	Upper
172	100		
171	38.3	28.7	43.1
170	26.1	20.1	30.1

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD


Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



#16
Acenaphthylene
Concen: 6.680 ng
RT: 13.76 min Scan# 1270
Delta R.T. -0.02 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.5	15.8	23.6
153	13.2	10.2	15.4









#17
Acenaphthene
Concen: 1.362 ng
RT: 14.10 min Scan# 1301
Delta R.T. -0.02 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
154	100		
153	113.1	88.9	133.3
152	53.7	44.2	66.2




Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



#18
Fluorene
Concen: 1.904 ng
RT: 15.09 min Scan# 1390
Delta R.T. -0.02 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
166	100		
165	101.6	77.2	115.8
167	14.7	10.5	15.7





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.79 min Scan# 1546

Delta R.T. -0.02 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Instrument :

BNA_N


ClientSampleId :

TP6-EP2-2MSD

Tot Ion:	188	Resp:	7572
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	8.9	8.3	12.5

Manual Integrations
APPROVED




mohammad
2/21/2020 2:19:59 PM



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BNC

Ion 80.00 (79.70 to 80.70): BNC



#21

Hexachlorobenzene

Concen: 0.071 ng


RT: 16.11 min Scan# 1482

Delta R.T. -0.02 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57



Tgt Ion:	284	Resp:	337
Ion Ratio	Lower	Upper	
284	100		
142	50.4	33.3	49.9#
249	33.8	25.7	38.5






Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E









#22
 Pentachlorophenol
 Concen: 0.342 ng
 RT: 16.48 min Scan# 1516
 Delta R.T. -0.04 min
 Lab File: BN009800.D
 Acq: 21 Feb 2020 03:57

Tot Ion:	266	Resp:	258
Ion Ratio	Lower	Upper	
266	100		
264	73.0	58.1	87.1
268	76.0	65.7	98.5

Instrument :
 BNA_N
 ClientSampleId :
 TP6-EP2-2MSD


Manual Integrations
 APPROVED


mohammad
 2/21/2020 2:19:59 PM



#23
 Phenanthrene
 Concen: 13.152 ng
 RT: 16.84 min Scan# 1550
 Delta R.T. -0.02 min
 Lab File: BN009800.D
 Acq: 21 Feb 2020 03:57

Tgt Ion:	178	Resp:	296385
Ion Ratio	Lower	Upper	
178	100		
176	18.6	15.3	22.9
179	15.3	12.4	18.6





#24

Anthracene

Concen: 5.061 ng

RT: 16.92 min Scan# 1558

Delta R.T. -0.03 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Instrument :

BNA_N





ClientSampleId :

TP6-EP2-2MSD

Tot Ion:178	Resp:	96406
Ion Ratio	Lower	Upper
178	100	
176	18.1	14.3 21.5
179	16.3	12.4 18.6

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



#25

Fluoranthene-d10

Concen: 0.054 ng




RT: 18.83 min Scan# 1827


Delta R.T. -0.03 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Tgt Ion:212	Resp:	1403
Ion Ratio	Lower	Upper
212	100	
106	11.8	8.4 12.6
104	13.5	4.8 7.2#





#26

Fluoranthene

Concen: 19.159 ng

RT: 18.86 min Scan# 1833

Delta R.T. -0.03 min

Lab File: BN009800.D

Acq: 21 Feb 2020 03:57

Instrument :

BNA_N


ClientSampleId :

TP6-EP2-2MSD




Tot Ion:	202	Resp:	516965
Ion Ratio	Lower	Upper	
202	100		
101	9.9	7.1	10.7
203	17.5	13.8	20.6

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#27

Chrysene-d12

Concen: 0.400 ng


RT: 21.00 min Scan# 2157

Delta R.T. -0.03 min



Lab File: BN009800.D


Acq: 21 Feb 2020 03:57

Tgt Ion:	240	Resp:	8031
Ion Ratio	Lower	Upper	
240	100		
120	17.7	7.9	11.9#
236	30.6	22.0	33.0



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#28
Pvrene
Concen: 15.459 ng
RT: 19.23 min Scan# 1907
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
202	100		
200	20.4	16.4	24.6
203	19.1	13.9	20.9




Manual Integrations
APPROVED


mohammad
2/21/2020 2:19:59 PM



#29
Terphenyl-d14
Concen: 0.225 ng
RT: 19.45 min Scan# 1951
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.5	8.0	12.0
122	16.2	10.0	15.0









#30
Benzo(a)anthracene
Concen: 11.923 ng
RT: 20.99 min Scan# 2156
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
228	100		
226	32.7	21.8	32.8
229	20.5	17.1	25.7




Manual Integrations
APPROVED


mohammad
2/21/2020 2:19:59 PM



#31
Chrysene
Concen: 8.308 ng
RT: 21.05 min Scan# 2161
Delta R.T. -0.02 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.3	24.8	37.2
229	22.0	16.4	24.6









#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.137 ng
 RT: 20.95 min Scan# 2152
 Delta R.T. -0.01 min
 Lab File: BN009800.D
 Acq: 21 Feb 2020 03:57

Instrument :
 BNA_N
 ClientSampleId :
 TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
149	100		
167	28.3	22.7	34.1
279	4.9	3.1	4.7




Manual Integrations
 APPROVED


mohammad
 2/21/2020 2:19:59 PM



#33
 Indeno(1,2,3-cd)pyrene
 Concen: 6.336 ng
 RT: 25.15 min Scan# 3091
 Delta R.T. -0.05 min
 Lab File: BN009800.D
 Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
276	100		
138	19.5	15.8	23.6
277	24.9	19.3	28.9









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.11 min Scan# 2495
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
264	100		
260	27.6	21.1	31.7
265	67.3	57.5	86.3




Manual Integrations
APPROVED


mohammad
2/21/2020 2:19:59 PM



#35
Benzo(b)fluoranthene
Concen: 12.196 ng
RT: 22.50 min Scan# 2316
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.4	23.0	34.6#
125	9.9	14.4	21.6#









#36
Benzo(k)fluoranthene
Concen: 4.145 ng m
RT: 22.53 min Scan# 2326
Delta R.T. -0.03 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.5	22.2	33.2
125	9.7	10.5	15.7




Manual Integrations
APPROVED


mohammad
2/21/2020 2:19:59 PM



#37
Benzo(a)pyrene
Concen: 10.933 ng
RT: 23.02 min Scan# 2469
Delta R.T. -0.04 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	25.3	37.9
125	10.1	17.7	26.5









#38
Dibenzo(a,h)anthracene
Concen: 2.133 ng
RT: 25.15 min Scan# 3091
Delta R.T. -0.06 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Instrument :
BNA_N
ClientSampleId :
TP6-EP2-2MSD

Tot Ion	Ratio	Lower	Upper
278	100		
139	17.4	14.5	21.7
279	28.2	24.3	36.5

Manual Integrations
APPROVED

mohammad
2/21/2020 2:19:59 PM



#39
Benzo(g,h,i)perylene
Concen: 7.063 ng
RT: 25.77 min Scan# 3272
Delta R.T. -0.06 min
Lab File: BN009800.D
Acq: 21 Feb 2020 03:57

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.9	21.4	32.0
138	19.4	16.6	24.8

