

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022120\
 Data File : BN009810.D
 Acq On : 21 Feb 2020 16:39
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 17:09:38 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	-0.02
2	1,4-Dioxane	0.410	0.408	0.5	116	-0.02
3	n-Nitrosodimethylamine	0.673	0.696	-3.4	136	-0.02
4 S	2-Fluorophenol	0.889	0.911	-2.5	123	-0.02
5 S	Phenol-d6	1.082	0.985	9.0	111	-0.02
6	bis(2-Chloroethyl)ether	1.052	1.108	-5.3	126	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	122	-0.03
8 S	Nitrobenzene-d5	0.333	0.293	12.0	114	-0.03
9	Naphthalene	1.091	1.062	2.7	122	-0.03
10	Hexachlorobutadiene	0.240	0.238	0.8	120	-0.01
11 SURR	2-Methylnaphthalene-d10	0.774	0.755	2.5	123	-0.03
12	2-Methylnaphthalene	0.748	0.735	1.7	124	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	126	-0.02
14 S	2,4,6-Tribromophenol	0.154	0.137	11.0	125	-0.02
15 S	2-Fluorobiphenyl	1.510	1.434	5.0	124	-0.03
16	Acenaphthylene	1.705	1.566	8.2	128	-0.02
17	Acenaphthene	1.199	1.147	4.3	125	-0.02
18	Fluorene	1.582	1.469	7.1	120	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	-0.02
20	4-Bromophenyl-phenylether	0.279	0.149	46.6#	67	-0.01
21	Hexachlorobenzene	0.249	0.242	2.8	120	-0.02
22	Pentachlorophenol	0.064	0.049	23.4	122	-0.02
23	Phenanthrene	1.190	1.096	7.9	118	-0.02
24	Anthracene	1.006	0.896	10.9	117	-0.01
25 SURR	Fluoranthene-d10	1.384	1.262	8.8	116	-0.02
26	Fluoranthene	1.425	1.285	9.8	115	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	117	-0.02
28	Pvrene	1.519	1.487	2.1	118	-0.02
29 S	Terphenyl-d14	0.956	0.901	5.8	113	-0.02
30	Benzo(a)anthracene	1.327	1.228	7.5	121	-0.02
31	Chrysene	1.536	1.491	2.9	116	-0.01
32	Bis(2-ethylhexyl)phthalate	0.648	0.771	-19.0	154#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.599	1.588	0.7	116	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	122	-0.01
35	Benzo(b)fluoranthene	1.462	1.275	12.8	116	-0.01
36	Benzo(k)fluoranthene	1.592	1.447	9.1	116	-0.01
37 C	Benzo(a)pyrene	1.334	1.283	3.8	120	-0.02
38	Dibenzo(a,h)anthracene	1.314	1.118	14.9	110	-0.02
39	Benzo(a,h,i)perylene	1.400	1.303	6.9	120	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022120\
Data File : BN009810.D
Acq On : 21 Feb 2020 16:39
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 21 17:09:38 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			