

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
 Data File : BN030003.D
 Acq On : 24 Feb 2024 21:32
 Operator : MA/JU
 Sample : P1520-09
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE125D1-20240216

Quant Time: Feb 26 02:45:19 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	4868	0.400	ng	0.00
7) Naphthalene-d8	10.647	136	12274	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	5484	0.400	ng	0.01
19) Phenanthrene-d10	17.255	188	10698	0.400	ng	# 0.00
29) Chrysene-d12	21.465	240	6802	0.400	ng	0.00
35) Perylene-d12	23.829	264	6276	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.435	112	1726	0.148	ng	0.00
5) Phenol-d6	7.038	99	1484	0.110	ng	0.01
8) Nitrobenzene-d5	9.025	82	4666	0.457	ng	0.01
11) 2-Methylnaphthalene-d10	12.246	152	7223	0.432	ng	0.00
14) 2,4,6-Tribromophenol	16.014	330	157	0.089	ng	0.01
15) 2-Fluorobiphenyl	13.132	172	10657	0.453	ng	0.01
27) Fluoranthene-d10	19.289	212	12849	0.496	ng	0.00
31) Terphenyl-d14	19.898	244	10194	0.604	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.355	88	37728	5.505	ng	96
34) Bis(2-ethylhexyl)phtha...	21.384	149	1108	0.070	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
Data File : BN030003.D
Acq On : 24 Feb 2024 21:32
Operator : MA/JU
Sample : P1520-09
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE125D1-20240216

Quant Time: Feb 26 02:45:19 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

