

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
 Data File : BN030006.D
 Acq On : 24 Feb 2024 23:20
 Operator : MA/JU
 Sample : P1507-12
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D2-20240214

Quant Time: Feb 26 02:45:55 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3926	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	10147	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	4840	0.400	ng	0.01
19) Phenanthrene-d10	17.255	188	8761	0.400	ng	0.00
29) Chrysene-d12	21.456	240	7446	0.400	ng	0.00
35) Perylene-d12	23.824	264	6572	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.435	112	1257	0.133	ng	0.00
5) Phenol-d6	7.031	99	1104	0.101	ng	0.00
8) Nitrobenzene-d5	9.025	82	3390	0.401	ng	0.01
11) 2-Methylnaphthalene-d10	12.246	152	5053	0.366	ng	0.00
14) 2,4,6-Tribromophenol	16.002	330	660	0.425	ng	0.00
15) 2-Fluorobiphenyl	13.121	172	8319	0.400	ng	0.00
27) Fluoranthene-d10	19.290	212	11129	0.524	ng	0.00
31) Terphenyl-d14	19.898	244	8921	0.483	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.348	88	31013	5.611	ng	# 95
25) Phenanthrene	17.293	178	604	0.023	ng	96
28) Fluoranthene	19.317	202	662	0.023	ng	# 66
30) Pyrene	19.684	202	848	0.025	ng	94
34) Bis(2-ethylhexyl)phtha...	21.385	149	12361	0.713	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
Data File : BN030006.D
Acq On : 24 Feb 2024 23:20
Operator : MA/JU
Sample : P1507-12
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D2-20240214

Quant Time: Feb 26 02:45:55 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

