

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
 Data File : BN030020.D
 Acq On : 25 Feb 2024 08:55
 Operator : MA/JU
 Sample : P1532-09
 Misc :
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW11-MW01I-20240219

Quant Time: Feb 26 02:49:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3964	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	10308	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	4715	0.400	ng	0.01
19) Phenanthrene-d10	17.255	188	9687	0.400	ng	0.00
29) Chrysene-d12	21.456	240	7010	0.400	ng	0.00
35) Perylene-d12	23.824	264	6604	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.435	112	902	0.095	ng	0.00
5) Phenol-d6	7.031	99	492	0.045	ng	0.00
8) Nitrobenzene-d5	9.025	82	2755	0.321	ng	0.01
11) 2-Methylnaphthalene-d10	12.250	152	4580	0.326	ng	0.00
14) 2,4,6-Tribromophenol	16.014	330	368	0.243	ng	0.01
15) 2-Fluorobiphenyl	13.132	172	6756	0.334	ng	0.01
27) Fluoranthene-d10	19.290	212	9983	0.425	ng	0.00
31) Terphenyl-d14	19.893	244	9069	0.521	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.385	149	477	0.029	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
Data File : BN030020.D
Acq On : 25 Feb 2024 08:55
Operator : MA/JU
Sample : P1532-09
Misc :
ALS Vial : 70 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW11-MW01I-20240219

Quant Time: Feb 26 02:49:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

