

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
 Data File : BN030022.D
 Acq On : 25 Feb 2024 10:07
 Operator : MA/JU
 Sample : P1533-02
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW8-MW01D1-20240217

Quant Time: Feb 26 02:49:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.854	152	3878	0.400	ng	0.00
7) Naphthalene-d8	10.647	136	10460	0.400	ng	# 0.00
13) Acenaphthene-d10	14.500	164	4585	0.400	ng	0.01
19) Phenanthrene-d10	17.255	188	9541	0.400	ng	# 0.00
29) Chrysene-d12	21.456	240	6973	0.400	ng	0.00
35) Perylene-d12	23.823	264	6468	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.435	112	1105	0.119	ng	0.00
5) Phenol-d6	7.031	99	612	0.057	ng	0.00
8) Nitrobenzene-d5	9.025	82	2380	0.273	ng	0.01
11) 2-Methylnaphthalene-d10	12.253	152	3766	0.264	ng	0.00
14) 2,4,6-Tribromophenol	16.014	330	487	0.331	ng	0.01
15) 2-Fluorobiphenyl	13.132	172	5430	0.276	ng	0.01
27) Fluoranthene-d10	19.289	212	9268	0.401	ng	0.00
31) Terphenyl-d14	19.898	244	8535	0.493	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.355	88	985	0.180	ng	# 34
34) Bis(2-ethylhexyl)phtha...	21.384	149	372	0.023	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
Data File : BN030022.D
Acq On : 25 Feb 2024 10:07
Operator : MA/JU
Sample : P1533-02
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW8-MW01D1-20240217

Quant Time: Feb 26 02:49:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 26 02:39:45 2024
Response via : Initial Calibration

