

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022424\
 Data File : BN030010.D
 Acq On : 25 Feb 2024 02:20
 Operator : MA/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 26 02:46:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN012924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 26 02:39:45 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	129	0.00
2	1,4-Dioxane	0.400	0.368	8.0	118	0.00
3	n-Nitrosodimethylamine	0.400	0.416	-4.0	138	0.00
4 S	2-Fluorophenol	0.400	0.436	-9.0	148	0.00
5 S	Phenol-d6	0.400	0.389	2.8	137	0.00
6	bis(2-Chloroethyl)ether	0.400	0.391	2.3	135	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	121	0.00
8 S	Nitrobenzene-d5	0.400	0.394	1.5	128	0.01
9	Naphthalene	0.400	0.395	1.3	122	0.00
10	Hexachlorobutadiene	0.400	0.431	-7.7	132	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	129	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	123	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	122	0.01
14 S	2,4,6-Tribromophenol	0.400	0.444	-11.0	150	0.00
15 S	2-Fluorobiphenyl	0.400	0.375	6.3	118	0.01
16	Acenaphthylene	0.400	0.354	11.5	116	0.01
17	Acenaphthene	0.400	0.397	0.8	126	0.01
18	Fluorene	0.400	0.381	4.8	122	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	134	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.284	29.0#	108	0.00
21	4-Bromophenyl-phenylether	0.400	0.365	8.8	128	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	136	0.00
23	Atrazine	0.400	0.396	1.0	146	0.00
24	Pentachlorophenol	0.400	0.409	-2.2	157	0.00
25	Phenanthrene	0.400	0.366	8.5	129	0.00
26	Anthracene	0.400	0.368	8.0	134	0.00
27 SURR	Fluoranthene-d10	0.400	0.437	-9.2	159	0.00
28	Fluoranthene	0.400	0.411	-2.7	154	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	165	0.00
30	Pyrene	0.400	0.370	7.5	150	0.00
31 S	Terphenyl-d14	0.400	0.374	6.5	155	0.00
32	Benzo(a)anthracene	0.400	0.359	10.3	157	0.00
33	Chrysene	0.400	0.411	-2.7	173	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.321	19.8	142	0.00
35 I	Perylene-d12	0.400	0.400	0.0	166	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.423	-5.7	189	0.00
37	Benzo(b)fluoranthene	0.400	0.409	-2.2	181	0.00
38	Benzo(k)fluoranthene	0.400	0.399	0.3	166	0.00
39 C	Benzo(a)pyrene	0.400	0.398	0.5	170	0.00
40	Dibenzo(a,h)anthracene	0.400	0.412	-3.0	184	0.00
41	Benzo(g,h,i)perylene	0.400	0.402	-0.5	171	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4