

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022619\
 Data File : BN004515.D
 Acq On : 27 Feb 2019 00:51
 Operator : JU/SJ
 Sample : K1553-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BF5X0

Quant Time: Feb 27 05:42:00 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 27 05:37:09 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	32753	20.00	ng/ul	0.00
18) Naphthalene-d8	10.52	136	158341	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.37	164	100370	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.12	188	245547	20.00	ng/ul	-0.01
78) Chrysene-d12	21.32	240	348563	20.00	ng/ul	-0.01
86) Perylene-d12	23.59	264	406363	20.00	ng/ul	-0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.24	96	3337	6.13	ng/uL	0.00
5) Phenol-d5	6.89	99	42158	14.44	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	7.08	67	21854	14.92	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.26	132	37857	15.99	ng/ul	-0.01
13) 4-Methylphenol-d8	8.43	113	36316	14.47	ng/ul	-0.01
19) Nitrobenzene-d5	8.90	128	18503	17.65	ng/ul	0.00
22) 2-Nitrophenol-d4	9.61	143	21146	19.33	ng/ul	-0.01
26) 2,4-Dichlorophenol-d3	10.14	165	42148	16.99	ng/ul	0.00
29) 4-Chloroaniline-d4	10.66	131	47223	17.21	ng/ul	-0.01
44) Dimethylphthalate-d6	13.79	166	142941	16.79	ng/ul	-0.01
47) Acenaphthylene-d8	14.07	160	166286	16.49	ng/ul	-0.01
52) 4-Nitrophenol-d4	14.57	143	19080	11.71	ng/ul	-0.01
58) Fluorene-d10	15.37	176	128502	17.15	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.49	200	12640	8.83	ng/ul	-0.01
71) Anthracene-d10	17.22	188	200206	17.01	ng/ul	-0.01
79) Pyrene-d10	19.52	212	247711	15.86	ng/ul	-0.01
90) Benzo(a)pyrene-d12	23.44	264	347102	16.28	ng/ul	-0.02

Target Compounds					Ovalue	
45) Dimethylphthalate	13.83	163	28348	3.369	ng/ul	99
70) Phenanthrene	17.16	178	14926	1.096	ng/ul	99
77) Fluoranthene	19.19	202	35029	2.089	ng/ul#	96
80) Pyrene	19.55	202	32398	1.681	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN022619\
Data File : BN004515.D
Acq On : 27 Feb 2019 00:51
Operator : JU/SJ
Sample : K1553-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BF5X0

Quant Time: Feb 27 05:42:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Feb 27 05:37:09 2019
Response via : Initial Calibration

