


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022619\
 Data File : BN004499.D
 Acq On : 26 Feb 2019 14:10
 Operator : JU/SJ
 Sample : DFTPP09
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP09

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
 Title : SVOA CALIBRATION
 Last Update : Wed Feb 27 05:37:09 2019



AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2470

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.7	12832	PASS
68	69	0.00	2	0.8	117	PASS
69	198	0.00	100	28.1	15215	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.5	22493	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	54221	PASS
199	198	5	9	6.5	3529	PASS
275	198	10	60	31.7	17172	PASS
365	198	1	100	4.3	2334	PASS
441	198	0.01	100	14.9	8105	PASS
442	198	50	100	92.5	50165	PASS
443	442	15	24	18.7	9395	PASS

m/z	Abundance
39.00	973.0
50.00	3773.0
51.00	13557.0
52.00	701.0
56.00	466.0
57.00	995.0
63.00	612.0
65.00	306.0
68.10	353.0
69.00	15648.0
74.00	1633.0
75.00	2707.0
76.00	906.0
77.00	17624.0
78.00	1298.0
79.00	1396.0
80.00	1011.0
81.00	1343.0
82.00	304.0
85.90	410.0
91.00	332.0
92.00	397.0
93.00	2216.0
98.00	1893.0
99.00	1746.0
101.00	801.0
103.90	563.0
104.90	419.0
107.00	6247.0
108.00	954.0
110.00	11113.0
111.00	1785.0
116.00	371.0
117.00	5477.0
117.90	450.0
121.90	427.0
123.00	670.0
127.00	22336.0
128.00	1714.0
129.00	8913.0
129.90	774.0
134.90	752.0
137.00	541.0
141.00	1255.0
141.90	353.0
147.00	610.0
148.00	1391.0
149.00	343.0
152.90	368.0
154.00	306.0
155.00	716.0
156.00	1071.0
159.90	413.0
160.90	557.0
164.90	512.0
166.00	368.0
167.00	2369.0
168.00	1785.0
174.00	553.0
175.00	1070.0
177.00	465.0
179.00	1749.0
180.00	1349.0
181.00	550.0
185.00	961.0
186.00	6695.0
187.00	2077.0
188.00	303.0
188.90	518.0
191.90	548.0
193.00	664.0
196.00	1386.0
198.00	50296.0
199.00	3430.0
203.00	373.0
204.00	1792.0
205.00	3183.0
206.00	13738.0
207.00	2187.0
208.00	545.0
211.00	547.0
217.00	3652.0
218.00	435.0
221.00	2038.0
223.00	860.0
224.00	7579.0
225.00	2104.0
227.00	3255.0
228.00	503.0
229.00	701.0
236.90	376.0

Instrument :
BNA_N
ClientSampleId :
DFTPP09

242.00	454.0
243.00	451.0
244.00	6044.0
245.00	827.0
246.00	1072.0
255.00	30568.0
256.00	4834.0
257.00	463.0
258.00	1915.0
265.00	676.0
273.00	899.0
274.00	2448.0
275.00	14868.0
276.00	1950.0
277.00	1234.0
296.00	3831.0
297.00	620.0
303.00	451.0
315.00	516.0
323.00	1272.0
334.00	838.0
351.90	368.0
354.00	391.0
365.00	1902.0
372.00	635.0
403.00	364.0
420.90	310.0
422.00	316.0
423.00	2399.0
424.00	846.0
441.00	5220.0
442.00	33992.0
443.00	6948.0
444.00	664.0

Instrument :
BNA_N
ClientSampleId :
DFTPP09

m/z	Abundance
39.00	1102.0
50.00	4038.0
51.00	15388.0
52.00	842.0
55.90	521.0
57.00	1095.0
63.00	673.0
65.00	379.0
69.00	18168.0
74.00	1953.0
75.00	3314.0
76.00	1065.0
77.00	20424.0
78.00	1443.0
79.00	1523.0
79.90	1268.0
81.00	1668.0
81.90	381.0
82.90	365.0
85.90	445.0
90.90	349.0
92.00	440.0
93.00	2766.0
98.00	2102.0
99.00	1860.0
101.00	934.0
104.00	687.0
104.90	611.0
107.00	7417.0
108.00	1079.0
110.00	13662.0
111.00	2106.0
116.00	488.0
117.00	6072.0
118.00	535.0
121.90	573.0
122.90	816.0
123.90	417.0
127.00	26824.0
128.00	2038.0
129.00	10768.0
130.00	919.0
133.90	368.0
135.00	917.0
136.00	329.0
137.00	526.0
141.00	1457.0
142.00	527.0
142.90	345.0
147.00	790.0
148.00	1699.0
148.90	472.0
153.00	484.0
154.00	346.0
155.00	814.0
156.00	1292.0
160.00	510.0
160.90	735.0
164.90	600.0
166.00	556.0
167.00	3189.0
168.00	1717.0
172.00	320.0
173.00	375.0
174.00	551.0
175.00	1185.0
175.90	377.0
177.00	677.0
179.00	2239.0
180.00	1635.0
181.00	784.0
185.00	1262.0
186.00	8994.0
187.00	2694.0
189.00	592.0
192.00	695.0
193.00	916.0
196.00	1827.0
198.00	64472.0
199.00	4189.0
199.90	383.0
201.40	333.0
203.00	465.0
204.00	2478.0
205.00	4078.0
206.00	17408.0
207.00	2568.0
207.90	589.0
210.20	334.0
210.90	726.0
216.00	397.0

Instrument :
BNA_N
ClientSampleId :
DFTPP09

217.00	4825.0
218.00	568.0
221.00	2919.0
221.80	697.0
223.00	1110.0
224.00	10124.0
225.00	2713.0
226.00	310.0
227.00	4134.0
228.00	703.0
229.00	878.0
231.00	392.0
235.00	369.0
237.00	422.0
242.00	508.0
243.00	573.0
244.00	8182.0
245.00	1102.0
246.00	1715.0
246.90	344.0
249.00	308.0
255.00	43520.0
256.00	6382.0
257.00	559.0
258.00	2520.0
258.90	422.0
265.00	1019.0
273.00	1371.0
274.00	3568.0
275.00	20096.0
276.00	2663.0
277.00	1786.0
293.00	340.0
296.00	5478.0
297.00	749.0
303.00	739.0
315.00	615.0
316.00	418.0
323.00	1997.0
326.90	337.0
334.00	1231.0
335.00	306.0
346.00	375.0
352.10	591.0
352.90	396.0
354.00	645.0
365.00	2816.0
366.00	368.0
372.00	1023.0
402.00	443.0
403.00	587.0
421.00	582.0
421.90	444.0
423.00	3671.0
424.00	1133.0
441.10	9823.0
442.00	59392.0
443.00	11085.0
444.10	1024.0

Instrument :
BNA_N
ClientSampleId :
DFTPP09

m/z	Abundance
39.00	715.0
50.00	2480.0
51.00	9551.0
52.00	510.0
55.90	340.0
57.00	702.0
63.00	391.0
69.00	11831.0
74.00	1273.0
75.00	1971.0
76.00	610.0
77.00	13341.0
78.00	1028.0
79.00	1080.0
80.00	756.0
81.00	1036.0
86.00	343.0
93.00	1752.0
98.00	1498.0
99.00	1117.0
100.90	637.0
103.90	374.0
104.90	357.0
107.00	4865.0
107.90	760.0
110.00	9161.0
111.00	1355.0
116.00	308.0
117.00	4309.0
122.00	395.0
122.90	606.0
127.00	18320.0
128.00	1416.0
129.00	7558.0
130.00	603.0
134.90	600.0
136.90	386.0
141.00	1086.0
142.00	347.0
147.00	527.0
148.00	1203.0
152.90	352.0
155.00	594.0
156.00	753.0
160.00	376.0
161.00	532.0
164.90	400.0
166.10	372.0
167.00	2229.0
168.00	1271.0
174.00	515.0
175.00	851.0
176.90	483.0
178.90	1666.0
180.00	1141.0
181.00	513.0
185.00	883.0
186.00	6438.0
187.00	1892.0
188.90	420.0
192.00	485.0
193.00	685.0
196.00	1517.0
198.00	47896.0
199.00	2969.0
203.00	352.0
204.00	1875.0
205.00	2961.0
206.00	12764.0
207.00	2031.0
208.00	502.0
211.00	479.0
216.90	3653.0
217.90	399.0
221.00	2185.0
221.80	494.0
223.00	879.0
224.00	7914.0
225.00	1962.0
227.00	3237.0
228.00	461.0
228.90	730.0
242.00	421.0
243.00	458.0
244.00	6102.0
245.00	772.0
246.00	1202.0
255.00	34528.0
256.00	4733.0
257.00	426.0
258.00	2076.0

Instrument :
BNA_N
ClientSampleId :
DFTPP09

258.90	313.0
265.00	721.0
273.00	973.0
274.00	2678.0
275.00	16552.0
276.00	2115.0
277.00	1410.0
293.00	391.0
296.00	4960.0
296.90	610.0
303.00	649.0
315.00	700.0
316.00	345.0
323.00	1645.0
324.00	320.0
327.00	319.0
334.00	1059.0
345.90	417.0
352.00	526.0
352.90	411.0
354.00	546.0
364.90	2286.0
366.00	362.0
372.00	915.0
402.00	436.0
403.00	547.0
421.00	534.0
422.00	534.0
423.00	3583.0
424.00	877.0
441.00	9273.0
442.00	57112.0
443.10	10154.0
444.00	1037.0

Instrument :
BNA_N
ClientSampleId :
DFTPP09

m/z	Abundance
207.00	314.0

Instrument :
BNA_N
ClientSampleId :
DFTPP09