

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022719\
 Data File : BN004541.D
 Acq On : 28 Feb 2019 00:00
 Operator : JU/SJ
 Sample : K1600-08
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BF7Z0

Quant Time: Feb 28 06:37:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 28 06:33:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	53185	20.00	ng/ul	-0.01
18) Naphthalene-d8	10.52	136	248697	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.37	164	156145	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.12	188	387729	20.00	ng/ul	-0.01
78) Chrysene-d12	21.32	240	496240	20.00	ng/ul	0.00
86) Perylene-d12	23.59	264	583008	20.00	ng/ul	-0.01

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.24	96	5348	6.05	ng/uL	0.00
5) Phenol-d5	6.89	99	33724	7.12	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	7.08	67	68017	28.60	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.26	132	98523	25.63	ng/ul	-0.01
13) 4-Methylphenol-d8	8.43	113	65124	15.98	ng/ul	-0.01
19) Nitrobenzene-d5	8.89	128	58184	35.34	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.61	143	67024	39.00	ng/ul	-0.01
26) 2,4-Dichlorophenol-d3	10.13	165	118195	30.34	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.66	131	9511	2.21	ng/ul	-0.01
44) Dimethylphthalate-d6	13.79	166	418723	31.62	ng/ul	-0.01
47) Acenaphthylene-d8	14.07	160	507362	32.34	ng/ul	-0.01
52) 4-Nitrophenol-d4	14.57	143	17625	6.95	ng/ul	-0.01
58) Fluorene-d10	15.37	176	365668	31.37	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.49	200	74175	32.80	ng/ul	-0.01
71) Anthracene-d10	17.22	188	593393	31.93	ng/ul	-0.01
79) Pyrene-d10	19.52	212	720378	32.39	ng/ul	-0.01
90) Benzo(a)pyrene-d12	23.45	264	939125	30.69	ng/ul	-0.01

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022719\
Data File : BN004541.D
Acq On : 28 Feb 2019 00:00
Operator : JU/SJ
Sample : K1600-08
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BF7Z0

Quant Time: Feb 28 06:37:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 28 06:33:25 2019
Response via : Initial Calibration

