

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022719\
 Data File : BN004549.D
 Acq On : 28 Feb 2019 04:44
 Operator : JU/SJ
 Sample : K1602-02
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BF5D4

Quant Time: Feb 28 06:37:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 28 06:33:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	52796	20.00	ng/ul	-0.01
18) Naphthalene-d8	10.52	136	248578	20.00	ng/ul	-0.02
36) Acenaphthene-d10	14.37	164	150377	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.12	188	364702	20.00	ng/ul	-0.01
78) Chrysene-d12	21.32	240	476966	20.00	ng/ul	0.00
86) Perylene-d12	23.59	264	556880	20.00	ng/ul	-0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.23	96	5008	5.70	ng/uL	-0.01
5) Phenol-d5	6.89	99	24650	5.24	ng/ul	-0.01
7) Bis-(2-Chloroethyl)ether-d	7.07	67	46324	19.62	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.26	132	70433	18.46	ng/ul	-0.01
13) 4-Methylphenol-d8	8.43	113	48578	12.01	ng/ul	-0.01
19) Nitrobenzene-d5	8.89	128	40111	24.38	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.61	143	48069	27.98	ng/ul	-0.01
26) 2,4-Dichlorophenol-d3	10.13	165	87752	22.53	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.66	131	7505	1.74	ng/ul	-0.01
44) Dimethylphthalate-d6	13.78	166	307340	24.10	ng/ul	-0.02
47) Acenaphthylene-d8	14.07	160	362777	24.01	ng/ul	-0.01
52) 4-Nitrophenol-d4	14.57	143	11356	4.65	ng/ul	-0.01
58) Fluorene-d10	15.37	176	264737	23.58	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.49	200	49630	23.33	ng/ul	-0.01
71) Anthracene-d10	17.22	188	435718	24.93	ng/ul	-0.01
79) Pyrene-d10	19.52	212	524256	24.52	ng/ul	-0.01
90) Benzo(a)pyrene-d12	23.44	264	666589	22.81	ng/ul	-0.02

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN022719\
Data File : BN004549.D
Acq On : 28 Feb 2019 04:44
Operator : JU/SJ
Sample : K1602-02
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BF5D4

Quant Time: Feb 28 06:37:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN020719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 28 06:33:25 2019
Response via : Initial Calibration

