

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022723\
 Data File : BN024105.D
 Acq On : 27 Feb 2023 12:47
 Operator : CG/JU
 Sample : SST02073
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST020240

Quant Time: Feb 27 14:07:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN022723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 27 14:05:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.063	152	233526	20.000	ng/u1	0.00
20) Naphthalene-d8	10.887	136	923087	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.704	164	544891	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	1189064	20.000	ng/u1	0.00
79) Chrysene-d12	21.639	240	1142277	20.000	ng/u1	0.00
88) Perylene-d12	24.157	264	1304845	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	48163	8.340	ng/uL	0.00
4) Pyridine-d5	3.834	84	323613	20.201	ng/u1	0.00
7) Phenol-d5	7.205	99	398182	19.946	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.387	67	247010	19.607	ng/u1	0.00
11) 2-Chlorophenol-d4	7.587	132	330410	21.280	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	313365	20.093	ng/u1	0.00
21) Nitrobenzene-d5	9.234	128	135417	18.376	ng/u1	0.00
24) 2-Nitrophenol-d4	9.957	143	147175	17.787	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.499	165	314927	20.424	ng/u1	0.00
31) 4-Chloroaniline-d4	11.016	131	424924	20.390	ng/u1	0.00
46) Dimethylphthalate-d6	14.110	166	891807	21.176	ng/u1	0.00
49) Acenaphthylene-d8	14.398	160	1061824	22.759	ng/u1	0.00
54) 4-Nitrophenol-d4	14.875	143	153598	21.202	ng/u1	0.00
60) Fluorene-d10	15.692	176	763532	21.532	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	105229	12.809	ng/u1	0.00
73) Anthracene-d10	17.545	188	1162845	21.468	ng/u1	0.00
81) Pyrene-d10	19.833	212	1376023	19.998	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.004	264	1460685	21.881	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	51414	8.417	ng/uL	100
5) Pyridine	3.858	79	327431	20.331	ng/u1	100
6) Benzaldehyde	7.193	77	238401	27.562	ng/u1	100
8) Phenol	7.234	94	407896	20.242	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.481	93	328020	20.027	ng/u1	100
12) 2-Chlorophenol	7.622	128	334207	21.170	ng/u1	100
13) 2-Methylphenol	8.499	108	310065	20.540	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.605	45	453260	22.215	ng/u1	100
16) Acetophenone	8.893	105	487735	20.047	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.875	70	240174	18.381	ng/u1	100
18) 4-Methylphenol	8.828	108	335808	20.649	ng/u1	100
19) Hexachloroethane	9.158	117	132388	20.070	ng/u1	100
22) Nitrobenzene	9.275	77	340285	18.201	ng/u1	100
23) Isophorone	9.805	82	711402	20.678	ng/u1	100
25) 2-Nitrophenol	9.993	139	166668	19.254	ng/u1	100
26) 2,4-Dimethylphenol	10.046	107	358180	20.767	ng/u1	100
27) Bis(2-Chloroethoxy)met...	10.293	93	423725	19.828	ng/u1	100
29) 2,4-Dichlorophenol	10.522	162	307563	20.568	ng/u1	100
30) Naphthalene	10.940	128	1054247	21.298	ng/u1	100
32) 4-Chloroaniline	11.040	127	428667	20.605	ng/u1	100
33) Hexachlorobutadiene	11.228	225	206637	19.270	ng/u1	100
34) Caprolactam	11.793	113	94182	22.149	ng/u1	100
35) 4-Chloro-3-methylphenol	12.151	107	322214	20.507	ng/u1	100
36) 2-Methylnaphthalene	12.540	142	688088	20.646	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022723\
 Data File : BN024105.D
 Acq On : 27 Feb 2023 12:47
 Operator : CG/JU
 Sample : SST02073
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SST020240


Quant Time: Feb 27 14:07:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN022723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Feb 27 14:05:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.757	142	684279	20.181	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.904	216	381351	19.977	ng/ul	100
40) Hexachlorocyclopentadiene	12.887	237	183308	13.252	ng/ul	100
41) 2,4,6-Trichlorophenol	13.134	196	251485	20.759	ng/ul	100
42) 2,4,5-Trichlorophenol	13.204	196	268350	20.455	ng/ul	100
43) 1,1'-Biphenyl	13.540	154	914722	21.084	ng/ul	100
44) 2-Chloronaphthalene	13.581	162	714422	21.183	ng/ul	100
45) 2-Nitroaniline	13.781	65	134733	13.997	ng/ul	100
47) Dimethylphthalate	14.157	163	877652	20.921	ng/ul	100
48) 2,6-Dinitrotoluene	14.275	165	119125	14.528	ng/ul	100
50) Acenaphthylene	14.428	152	1128109	22.712	ng/ul	100
51) 3-Nitroaniline	14.598	138	143891	19.926	ng/ul	100
52) Acenaphthene	14.763	153	762968	21.989	ng/ul	100
53) 2,4-Dinitrophenol	14.798	184	69125	12.538	ng/ul	100
55) 4-Nitrophenol	14.887	109	121090	18.885	ng/ul	100
56) Dibenzofuran	15.098	168	1058097	21.555	ng/ul	100
57) 2,4-Dinitrotoluene	15.051	165	183809	16.152	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	15.316	232	225456	19.759	ng/ul	100
59) Diethylphthalate	15.516	149	858736	20.965	ng/ul	100
61) Fluorene	15.745	166	850182	21.132	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.740	204	424229	20.043	ng/ul	100
63) 4-Nitroaniline	15.757	138	139110	22.484	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.810	198	111383	13.941	ng/ul	100
67) N-Nitrosodiphenylamine	15.951	169	730710	21.324	ng/ul	100
68) 4-Bromophenyl-phenylether	16.634	248	272620	19.567	ng/ul	100
69) Hexachlorobenzene	16.751	284	319170	20.050	ng/ul	100
70) Atrazine	16.898	200	252410	18.855	ng/ul	100
71) Pentachlorophenol	17.092	266	195154	19.300	ng/ul	100
72) Phenanthrene	17.486	178	1365518	21.574	ng/ul	100
74) Anthracene	17.581	178	1383081	21.823	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.504	216	391564	20.529	ng/ul	100
76) Pentachlorobenzene	15.016	250	367572	19.417	ng/ul	100
77) Carbazole	17.845	167	1260521	23.078	ng/ul	100
78) Di-n-butylphthalate	18.416	149	1400975	15.358	ng/ul	100
80) Fluoranthene	19.498	202	1629279	19.901	ng/ul	100
82) Pyrene	19.863	202	1693748	20.324	ng/ul	100
83) Butylbenzylphthalate	20.763	149	247205	7.571	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.551	252	514228	19.138	ng/ul	100
85) Benzo(a)anthracene	21.622	228	1709365	21.596	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.551	149	472663	9.756	ng/ul	100
87) Chrysene	21.674	228	1589263	21.697	ng/ul	100
89) Di-n-octyl phthalate	22.527	149	958156	10.621	ng/ul	100
90) Benzo(b)fluoranthene	23.380	252	1784928	20.762	ng/ul	100
91) Benzo(k)fluoranthene	23.433	252	1667270	20.355	ng/ul	100
93) Benzo(a)pyrene	24.051	252	1582427	21.681	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.821	276	2115622	22.458	ng/ul	100
95) Dibenzo(a,h)anthracene	26.845	278	1739920	22.258	ng/ul	100
96) Benzo(g,h,i)perylene	27.639	276	1733297	22.750	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Feb 27 14:07:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN022723.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Feb 27 14:05:23 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.417 ng/uL
RT: 3.452 min Scan# 62
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion: 88 Resp: 51414
Ion Ratio Lower Upper
88 100
43 29.6 23.7 35.5
58 73.4 58.7 88.1



#5
Pyridine
Concen: 20.331 ng/uL
RT: 3.858 min Scan# 131
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion: 79 Resp: 327431
Ion Ratio Lower Upper
79 100
52 75.2 60.2 90.2
51 36.6 29.3 43.9









#6
Benzaldehyde
Concen: 27.562 ng/ul
RT: 7.193 min Scan# 698
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion: 77 Resp: 238401
Ion Ratio Lower Upper
77 100
105 96.4 77.1 115.7
106 95.3 76.2 114.4



#8
Phenol
Concen: 20.242 ng/ul
RT: 7.234 min Scan# 705
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion: 94 Resp: 407896
Ion Ratio Lower Upper
94 100
65 28.9 23.1 34.7
66 38.3 30.6 46.0









#10
Bis(2-Chloroethyl)ether
Concen: 20.027 ng/ul
RT: 7.481 min Scan# 747
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion: 93 Resp: 328020
Ion Ratio Lower Upper
93 100
63 76.1 60.9 91.3
95 33.5 26.8 40.2



#12
2-Chlorophenol
Concen: 21.170 ng/ul
RT: 7.622 min Scan# 771
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:128 Resp: 334207
Ion Ratio Lower Upper
128 100
64 44.6 35.7 53.5
130 32.0 25.6 38.4





#13

2-Methylphenol

Concen: 20.540 ng/ul

RT: 8.499 min Scan# 91

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :





SSTD020240

Tgt Ion:108 Resp: 310065

Ion Ratio Lower Upper

108 100

107 89.6 71.7 107.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.215 ng/ul

RT: 8.605 min Scan# 938

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion: 45 Resp: 453260


Ion Ratio Lower Upper

45 100

77 14.0 11.2 16.8

79 10.6 8.5 12.7









#16
Acetophenone
Concen: 20.047 ng/ul
RT: 8.893 min Scan# 987
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:105 Resp: 487735
Ion Ratio Lower Upper
105 100
77 78.9 63.1 94.7
51 27.0 21.6 32.4



#17
N-Nitroso-di-n-propylamine
Concen: 18.381 ng/ul
RT: 8.875 min Scan# 984
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion: 70 Resp: 240174
Ion Ratio Lower Upper
70 100
42 59.0 47.2 70.8
101 9.2 7.4 11.0
130 19.1 15.3 22.9









#18
4-Methylphenol
Concen: 20.649 ng/ul
RT: 8.828 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:108 Resp: 335808
Ion Ratio Lower Upper
108 100
107 111.9 89.5 134.3



#19
Hexachloroethane
Concen: 20.070 ng/ul
RT: 9.158 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:117 Resp: 132388
Ion Ratio Lower Upper
117 100
201 107.4 85.9 128.9
199 67.4 53.9 80.9









#22
Nitrobenzene
Concen: 18.201 ng/ul
RT: 9.275 min Scan# 1052
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion: 77 Resp: 340285
Ion Ratio Lower Upper
77 100
123 44.9 35.9 53.9
65 13.3 10.6 16.0



#23
Isophorone
Concen: 20.678 ng/ul
RT: 9.805 min Scan# 1142
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion: 82 Resp: 711402
Ion Ratio Lower Upper
82 100
95 6.7 5.4 8.0
138 17.9 14.3 21.5





#25

2-Nitrophenol

Concen: 19.254 ng/ul

RT: 9.993 min Scan# 1174

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:139 Resp: 166668

Ion Ratio Lower Upper

139 100

65 47.2 37.8 56.6

109 28.4 22.7 34.1



#26

2,4-Dimethylphenol

Concen: 20.767 ng/ul

RT: 10.046 min Scan# 1183

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:107 Resp: 358180


Ion Ratio Lower Upper

107 100

121 51.3 41.0 61.6

122 86.5 69.2 103.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.828 ng/ul
 RT: 10.293 min Scan# 1225
 Delta R.T. 0.000 min
 Lab File: BN024105.D
 Acq: 27 Feb 2023 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020240


Tgt Ion: 93 Resp: 423725
 Ion Ratio Lower Upper
 93 100
 95 31.7 25.4 38.0
 123 11.8 9.4 14.2



#29
 2,4-Dichlorophenol
 Concen: 20.568 ng/ul
 RT: 10.522 min Scan# 1264
 Delta R.T. 0.000 min
 Lab File: BN024105.D
 Acq: 27 Feb 2023 12:47

Tgt Ion: 162 Resp: 307563
 Ion Ratio Lower Upper
 162 100
 164 64.0 51.2 76.8
 98 34.8 27.8 41.8









#30
Naphthalene
Concen: 21.298 ng/ul
RT: 10.940 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:128 Resp: 1054247
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 12.7 10.2 15.2



#32
4-Chloroaniline
Concen: 20.605 ng/ul
RT: 11.040 min Scan# 1352
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:127 Resp: 428667
Ion Ratio Lower Upper
127 100
129 32.1 25.7 38.5





#33

Hexachlorobutadiene

Concen: 19.270 ng/ul

RT: 11.228 min Scan# 1384

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:225 Resp: 206637

Ion Ratio Lower Upper

225 100

223 64.3 51.4 77.2

227 59.7 47.8 71.6



#34

Caprolactam

Concen: 22.149 ng/ul

RT: 11.793 min Scan# 1480

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:113 Resp: 94182


Ion Ratio Lower Upper

113 100

55 161.6 129.3 193.9

56 122.5 98.0 147.0









#35
4-Chloro-3-methylphenol
Concen: 20.507 ng/ul
RT: 12.151 min Scan# 1541
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:107 Resp: 322214
Ion Ratio Lower Upper
107 100
144 27.0 21.6 32.4
142 81.2 65.0 97.4



#36
2-Methylnaphthalene
Concen: 20.646 ng/ul
RT: 12.540 min Scan# 1607
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:142 Resp: 688088
Ion Ratio Lower Upper
142 100
141 89.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 20.181 ng/ul

RT: 12.757 min Scan# 1644

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:142 Resp: 684279

Ion Ratio Lower Upper

142 100

141 89.2 71.4 107.0

116 3.7 3.0 4.4



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.977 ng/ul

RT: 12.904 min Scan# 1669

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Tgt Ion:216 Resp: 381351




Ion Ratio Lower Upper


216 100

214 76.5 61.2 91.8

179 18.9 15.1 22.7

108 16.0 12.8 19.2





#40

Hexachlorocyclopentadiene

Concen: 13.252 ng/ul

RT: 12.887 min Scan# 1666

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:237 Resp: 183308

Ion Ratio Lower Upper

237 100

235 59.7 47.8 71.6

272 14.1 11.3 16.9



#41

2,4,6-Trichlorophenol

Concen: 20.759 ng/ul

RT: 13.134 min Scan# 1708

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:196 Resp: 251485


Ion Ratio Lower Upper

196 100

198 88.6 70.9 106.3

200 29.0 23.2 34.8









#42
2,4,5-Trichlorophenol
Concen: 20.455 ng/ul
RT: 13.204 min Scan# 1720
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:196 Resp: 268350
Ion Ratio Lower Upper
196 100
198 100.2 80.2 120.2
200 31.5 25.2 37.8



#43
1,1'-Biphenyl
Concen: 21.084 ng/ul
RT: 13.540 min Scan# 1777
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:154 Resp: 914722
Ion Ratio Lower Upper
154 100
153 39.2 31.4 47.0
76 12.4 9.9 14.9





#44

2-Chloronaphthalene

Concen: 21.183 ng/ul

RT: 13.581 min Scan# 1784

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:162 Resp: 714422

Ion Ratio Lower Upper

162 100

164 32.0 25.6 38.4

127 39.5 31.6 47.4



#45

2-Nitroaniline

Concen: 13.997 ng/ul

RT: 13.781 min Scan# 1818

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion: 65 Resp: 134733


Ion Ratio Lower Upper

65 100

92 70.6 56.5 84.7

138 114.5 91.6 137.4









#47
Dimethylphthalate
Concen: 20.921 ng/ul
RT: 14.157 min Scan# 1882
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:163 Resp: 877652
Ion Ratio Lower Upper
163 100
194 4.2 3.4 5.0
164 10.8 8.6 13.0



#48
2,6-Dinitrotoluene
Concen: 14.528 ng/ul
RT: 14.275 min Scan# 1902
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:165 Resp: 119125
Ion Ratio Lower Upper
165 100
89 55.8 44.6 67.0
121 21.5 17.2 25.8





#50

Acenaphthylene

Concen: 22.712 ng/ul

RT: 14.428 min Scan# 1928

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:152 Resp: 1128109

Ion Ratio Lower Upper

152 100

151 19.5 15.6 23.4

153 12.9 10.3 15.5



#51

3-Nitroaniline

Concen: 19.926 ng/ul

RT: 14.598 min Scan# 1957

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:138 Resp: 143891


Ion Ratio Lower Upper

138 100

108 10.2 8.2 12.2

92 111.7 89.4 134.0





#52

Acenaphthene

Concen: 21.989 ng/ul

RT: 14.763 min Scan# 1985

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:153 Resp: 762968

Ion Ratio Lower Upper

153 100

152 48.0 38.4 57.6

154 86.1 68.9 103.3



#53

2,4-Dinitrophenol

Concen: 12.538 ng/ul

RT: 14.798 min Scan# 1991

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:184 Resp: 69125


Ion Ratio Lower Upper

184 100

63 61.4 49.1 73.7

154 63.1 50.5 75.7









#55
4-Nitrophenol
Concen: 18.885 ng/ul
RT: 14.887 min Scan# 2006
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:109 Resp: 121090
Ion Ratio Lower Upper
109 100
139 132.0 105.6 158.4
65 128.5 102.8 154.2



#56
Dibenzofuran
Concen: 21.555 ng/ul
RT: 15.098 min Scan# 2042
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:168 Resp: 1058097
Ion Ratio Lower Upper
168 100
139 37.7 30.2 45.2





#57

2,4-Dinitrotoluene

Concen: 16.152 ng/ul

RT: 15.051 min Scan# 2034

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:165 Resp: 183809

Ion Ratio Lower Upper

165 100

63 45.1 36.1 54.1

182 2.6 2.1 3.1



#58

2,3,4,6-Tetrachlorophenol

Concen: 19.759 ng/ul

RT: 15.316 min Scan# 2079

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Tgt Ion:232 Resp: 225456




Ion Ratio Lower Upper


232 100

131 45.5 36.4 54.6

130 2.6 2.1 3.1

166 31.6 25.3 37.9





#59

Diethylphthalate

Concen: 20.965 ng/ul

RT: 15.516 min Scan# 2113

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:149 Resp: 858736

Ion Ratio Lower Upper

149 100

177 21.8 17.4 26.2

150 12.8 10.2 15.4



#61

Fluorene

Concen: 21.132 ng/ul

RT: 15.745 min Scan# 2152

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:166 Resp: 850182


Ion Ratio Lower Upper

166 100

165 95.6 76.5 114.7

167 13.8 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 20.043 ng/ul
RT: 15.740 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:204 Resp: 424229
Ion Ratio Lower Upper
204 100
206 34.3 27.4 41.2
141 57.7 46.2 69.2



#63
4-Nitroaniline
Concen: 22.484 ng/ul
RT: 15.757 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:138 Resp: 139110
Ion Ratio Lower Upper
138 100
92 49.9 39.9 59.9
108 72.8 58.2 87.4









#66
4,6-Dinitro-2-methylphenol
Concen: 13.941 ng/ul
RT: 15.810 min Scan# 2163
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:198 Resp: 111383
Ion Ratio Lower Upper
198 100
182 4.9 3.9 5.9
77 25.5 20.4 30.6



#67
N-Nitrosodiphenylamine
Concen: 21.324 ng/ul
RT: 15.951 min Scan# 2187
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:169 Resp: 730710
Ion Ratio Lower Upper
169 100
168 64.3 51.4 77.2
167 35.8 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 19.567 ng/ul
RT: 16.634 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:248 Resp: 272620
Ion Ratio Lower Upper
248 100
250 95.0 76.0 114.0
141 70.6 56.5 84.7



#69
Hexachlorobenzene
Concen: 20.050 ng/ul
RT: 16.751 min Scan# 2323
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:284 Resp: 319170
Ion Ratio Lower Upper
284 100
142 33.1 26.5 39.7
249 31.9 25.5 38.3









#70
Atrazine
Concen: 18.855 ng/ul
RT: 16.898 min Scan# 2348
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:200 Resp: 252410
Ion Ratio Lower Upper
200 100
173 25.8 20.6 31.0
215 49.4 39.5 59.3



#71
Pentachlorophenol
Concen: 19.300 ng/ul
RT: 17.092 min Scan# 2381
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:266 Resp: 195154
Ion Ratio Lower Upper
266 100
264 62.9 50.3 75.5
268 64.4 51.5 77.3





#72

Phenanthrene

Concen: 21.574 ng/ul

RT: 17.486 min Scan# 2448

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:178 Resp: 1365518

Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.5 14.8 22.2



#74

Anthracene

Concen: 21.823 ng/ul

RT: 17.581 min Scan# 2464

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:178 Resp: 1383081


Ion Ratio Lower Upper

178 100

179 15.6 12.5 18.7

176 18.8 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.529 ng/uL

RT: 13.504 min Scan# 1771

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:216 Resp: 391564

Ion Ratio Lower Upper

216 100

214 79.7 63.8 95.6

218 47.8 38.2 57.4



#76

Pentachlorobenzene

Concen: 19.417 ng/uL

RT: 15.016 min Scan# 2028

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:250 Resp: 367572


Ion Ratio Lower Upper

250 100

248 66.6 53.3 79.9

252 64.4 51.5 77.3









#77
 Carbazole
 Concen: 23.078 ng/ul
 RT: 17.845 min Scan# 2509
 Delta R.T. 0.000 min
 Lab File: BN024105.D
 Acq: 27 Feb 2023 12:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020240


Tgt Ion:167 Resp: 1260521
 Ion Ratio Lower Upper
 167 100
 166 21.2 17.0 25.4
 139 13.1 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 15.358 ng/ul
 RT: 18.416 min Scan# 2606
 Delta R.T. 0.000 min
 Lab File: BN024105.D
 Acq: 27 Feb 2023 12:47

Tgt Ion:149 Resp: 1400975
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 5.2 4.2 6.2









#80
Fluoranthene
Concen: 19.901 ng/ul
RT: 19.498 min Scan# 2790
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:202 Resp: 1629279
Ion Ratio Lower Upper
202 100
101 10.8 8.6 13.0
100 8.4 6.7 10.1



#82
Pyrene
Concen: 20.324 ng/ul
RT: 19.863 min Scan# 2852
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:202 Resp: 1693748
Ion Ratio Lower Upper
202 100
101 12.5 10.0 15.0
100 10.2 8.2 12.2





#83

Butylbenzylphthalate

Concen: 7.571 ng/ul

RT: 20.763 min Scan# 3005

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:149 Resp: 247205

Ion Ratio Lower Upper

149 100

91 71.2 57.0 85.4

206 20.8 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 19.138 ng/ul

RT: 21.551 min Scan# 3139

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:252 Resp: 514228


Ion Ratio Lower Upper

252 100

254 62.7 50.2 75.2

126 10.5 8.4 12.6





#85

Benzo(a)anthracene

Concen: 21.596 ng/ul

RT: 21.622 min Scan# 3151

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:228 Resp: 1709365

Ion Ratio Lower Upper

228 100

229 19.1 15.3 22.9

226 26.3 21.0 31.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.756 ng/ul

RT: 21.551 min Scan# 3139

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:149 Resp: 472663


Ion Ratio Lower Upper

149 100

167 28.5 22.8 34.2

279 4.2 3.4 5.0





#87

Chrysene

Concen: 21.697 ng/ul

RT: 21.674 min Scan# 3160

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:228 Resp: 1589263

Ion Ratio Lower Upper

228 100

226 30.2 24.2 36.2

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 10.621 ng/ul




RT: 22.527 min Scan# 3305


Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Tgt Ion:149 Resp: 958156









#90
Benzo(b)fluoranthene
Concen: 20.762 ng/ul
RT: 23.380 min Scan# 3450
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47




Instrument :
BNA_N
ClientSampleId :
SSTD020240


Tgt Ion:252 Resp: 1784928
Ion Ratio Lower Upper
252 100
253 21.0 16.8 25.2
125 9.9 7.9 11.9



#91
Benzo(k)fluoranthene
Concen: 20.355 ng/ul
RT: 23.433 min Scan# 3459
Delta R.T. 0.000 min
Lab File: BN024105.D
Acq: 27 Feb 2023 12:47

Tgt Ion:252 Resp: 1667270
Ion Ratio Lower Upper
252 100
253 23.3 18.6 28.0
125 10.3 8.2 12.4





#93

Benzo(a)pyrene

Concen: 21.681 ng/ul

RT: 24.051 min Scan# 3564

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:252 Resp: 1582427

Ion Ratio Lower Upper

252 100

253 21.4 17.1 25.7

125 10.6 8.5 12.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 22.458 ng/ul

RT: 26.821 min Scan# 4035

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47




Tgt Ion:276 Resp: 2115622


Ion Ratio Lower Upper

276 100

138 22.5 18.0 27.0

277 25.0 20.0 30.0





#95

Dibenzo(a,h)anthracene

Concen: 22.258 ng/ul

RT: 26.845 min Scan# 4039

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Instrument :

BNA_N

ClientSampleId :

SSTD020240





Tgt Ion:278 Resp: 1739920

Ion Ratio Lower Upper

278 100

139 15.3 12.2 18.4

279 22.8 18.2 27.4



#96

Benzo(g,h,i)perylene

Concen: 22.750 ng/ul

RT: 27.639 min Scan# 4174

Delta R.T. 0.000 min

Lab File: BN024105.D

Acq: 27 Feb 2023 12:47

Tgt Ion:276 Resp: 1733297

Ion Ratio Lower Upper

276 100

138 22.4 17.9 26.9

277 25.3 20.2 30.4

