

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030220\
 Data File : BN010036.D
 Acq On : 02 Mar 2020 12:22
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 02 15:35:46 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 02 15:27:13 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	91	-0.03
2	1,4-Dioxane	0.400	0.406	-1.5	93	-0.03
3	n-Nitrosodimethylamine	0.400	0.357	10.8	93	-0.03
4 S	2-Fluorophenol	0.400	0.436	-9.0	103	-0.03
5 S	Phenol-d6	0.400	0.400	0.0	96	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.428	-7.0	102	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	97	-0.04
8 S	Nitrobenzene-d5	0.400	0.376	6.0	97	-0.03
9	Naphthalene	0.400	0.392	2.0	98	-0.04
10	Hexachlorobutadiene	0.400	0.381	4.8	92	-0.04
11 SURR	2-Methylnaphthalene-d10	0.400	0.388	3.0	98	-0.04
12	2-Methylnaphthalene	0.400	0.390	2.5	99	-0.04
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	-0.04
14 S	2,4,6-Tribromophenol	0.400	0.440	-10.0	126	-0.03
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	101	-0.04
16	Acenaphthylene	0.400	0.385	3.8	110	-0.04
17	Acenaphthene	0.400	0.386	3.5	103	-0.04
18	Fluorene	0.400	0.389	2.8	103	-0.03
19 I	Phenanthrene-d10	0.400	0.400	0.0	104	-0.04
20	4-Bromophenyl-phenylether	0.400	0.062	84.5#	42	0.00
21	Hexachlorobenzene	0.400	0.374	6.5	102	-0.04
22	Pentachlorophenol	0.400	0.436	-9.0	118	-0.04
23	Phenanthrene	0.400	0.374	6.5	106	-0.04
24	Anthracene	0.400	0.382	4.5	111	-0.03
25 SURR	Fluoranthene-d10	0.400	0.380	5.0	106	-0.03
26	Fluoranthene	0.400	0.386	3.5	109	-0.03
27 I	Chrysene-d12	0.400	0.400	0.0	104	-0.03
28	Pvrene	0.400	0.414	-3.5	111	-0.03
29 S	Terphenyl-d14	0.400	0.397	0.8	105	-0.03
30	Benzo(a)anthracene	0.400	0.357	10.8	104	-0.03
31	Chrysene	0.400	0.385	3.8	102	-0.03
32	Bis(2-ethylhexyl)phthalate	0.400	0.620	-55.0#	179	-0.03
33	Indeno(1,2,3-cd)pyrene	0.400	0.381	4.8	99	-0.06
34 I	Perylene-d12	0.400	0.400	0.0	108	-0.04
35	Benzo(b)fluoranthene	0.400	0.356	11.0	106	-0.04
36	Benzo(k)fluoranthene	0.400	0.359	10.3	102	-0.04
37 C	Benzo(a)pyrene	0.400	0.381	4.8	106	-0.04
38	Dibenzo(a,h)anthracene	0.400	0.345	13.8	100	-0.06
39	Benzo(a,h,i)perylene	0.400	0.359	10.3	103	-0.06

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030220\
Data File : BN010036.D
Acq On : 02 Mar 2020 12:22
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 02 15:35:46 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 02 15:27:13 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				