

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030220\
 Data File : BN010042.D
 Acq On : 02 Mar 2020 16:21
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 02 18:23:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 02 18:15:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	-0.03
2	1,4-Dioxane	0.410	0.398	2.9	104	-0.03
3	n-Nitrosodimethylamine	0.673	0.532	21.0	96	-0.02
4 S	2-Fluorophenol	0.889	0.906	-1.9	113	-0.03
5 S	Phenol-d6	1.082	1.109	-2.5	115	-0.02
6	bis(2-Chloroethyl)ether	1.052	1.023	2.8	108	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	119	-0.04
8 S	Nitrobenzene-d5	0.333	0.307	7.8	117	-0.03
9	Naphthalene	1.091	1.075	1.5	121	-0.04
10	Hexachlorobutadiene	0.240	0.229	4.6	113	-0.04
11 SURR	2-Methylnaphthalene-d10	0.774	0.783	-1.2	124	-0.04
12	2-Methylnaphthalene	0.748	0.763	-2.0	126	-0.04
13 I	Acenaphthene-d10	1.000	1.000	0.0	127	-0.04
14 S	2,4,6-Tribromophenol	0.154	0.164	-6.5	151#	-0.04
15 S	2-Fluorobiphenyl	1.510	1.398	7.4	122	-0.04
16	Acenaphthylene	1.705	1.754	-2.9	146	-0.05
17	Acenaphthene	1.199	1.176	1.9	130	-0.04
18	Fluorene	1.582	1.498	5.3	124	-0.04
19 I	Phenanthrene-d10	1.000	1.000	0.0	122	-0.04
20	4-Bromophenyl-phenylether	0.279	0.096	65.6#	44#	0.00
21	Hexachlorobenzene	0.249	0.243	2.4	125	-0.04
22	Pentachlorophenol	0.064	0.056	12.5	143	-0.04
23	Phenanthrene	1.190	1.109	6.8	123	-0.04
24	Anthracene	1.006	0.964	4.2	130	-0.04
25 SURR	Fluoranthene-d10	1.384	1.293	6.6	122	-0.04
26	Fluoranthene	1.425	1.446	-1.5	134	-0.04
27 I	Chrysene-d12	1.000	1.000	0.0	124	-0.04
28	Pvrene	1.519	1.577	-3.8	133	-0.04
29 S	Terphenyl-d14	0.956	0.916	4.2	122	-0.04
30	Benzo(a)anthracene	1.327	1.298	2.2	137	-0.04
31	Chrysene	1.536	1.445	5.9	120	-0.03
32	Bis(2-ethylhexyl)phthalate	0.648	0.891	-37.5#	190#	-0.03
33	Indeno(1,2,3-cd)pyrene	1.599	1.488	6.9	116	-0.07
34 I	Perylene-d12	1.000	1.000	0.0	124	-0.05
35	Benzo(b)fluoranthene	1.462	1.338	8.5	125	-0.04
36	Benzo(k)fluoranthene	1.592	1.384	13.1	114	-0.04
37 C	Benzo(a)pyrene	1.334	1.295	2.9	124	-0.05
38	Dibenzo(a,h)anthracene	1.314	1.135	13.6	114	-0.07
39	Benzo(a,h,i)perylene	1.400	1.211	13.5	114	-0.07

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030220\
Data File : BN010042.D
Acq On : 02 Mar 2020 16:21
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 02 18:23:18 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 02 18:15:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			