

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030320\
 Data File : BN010054.D
 Acq On : 02 Mar 2020 23:38
 Operator : CG/JU
 Sample : L1615-08DL 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleID :
 DBDN2DL

Manual Integrations
 APPROVED

mohammad
 3/4/2020 7:50:58 AM

Quant Time: Mar 03 03:04:06 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 03 01:05:31 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.37	152	71819	20.00	ng/ul	0.00
18) Naphthalene-d8	10.12	136	317887	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.01	164	209188	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.77	188	438872	20.00	ng/ul	0.00
78) Chrysene-d12	20.99	240	414773	20.00	ng/ul	0.00
86) Perylene-d12	23.07	264	437247	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.03	96	416	0.24	ng/uL	0.02
5) Phenol-d5	6.62	99	10502m	1.72	ng/ul	0.04
7) Bis-(2-Chloroethyl)ether-d	6.75	67	8281m	1.96	ng/ul	0.02
9) 2-Chlorophenol-d4	6.93	132	8123	1.77	ng/ul	0.02
13) 4-Methylphenol-d8	8.12	113	8492m	1.75	ng/ul	0.04
19) Nitrobenzene-d5	8.54	128	3448m	1.48	ng/ul	0.03
22) 2-Nitrophenol-d4	9.25	143	2974m	1.17	ng/ul	0.02
26) 2,4-Dichlorophenol-d3	9.83	165	8504m	1.74	ng/ul	0.08
29) 4-Chloroaniline-d4	10.35	131	3410m	0.62	ng/ul	0.08
44) Dimethylphthalate-d6	13.45	166	33361	2.14	ng/ul	0.02
47) Acenaphthylene-d8	13.70	160	34512	1.88	ng/ul	0.00
52) 4-Nitrophenol-d4	14.36	143	1821m	0.64	ng/ul	0.09
58) Fluorene-d10	15.02	176	28789	2.13	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.20	200	1477m	0.54	ng/ul	0.02
71) Anthracene-d10	16.87	188	46398	2.31	ng/ul	0.00
79) Pyrene-d10	19.17	212	50662	2.34	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.94	264	52015	2.38	ng/ul	0.00

Target Compounds

					Ovalue
48) Acenaphthylene	13.73	152	21330	1.071	ng/ul 97
70) Phenanthrene	16.81	178	147548	5.891	ng/ul 98
72) Anthracene	16.90	178	184794	7.217	ng/ul 99
77) Fluoranthene	18.84	202	834186	28.433	ng/ul 98
80) Pyrene	19.20	202	1164377	38.938	ng/ul 98
83) Benzo(a)anthracene	20.97	228	370878	13.012	ng/ul 95
85) Chrysene	21.02	228	501640	17.877	ng/ul 99
88) Benzo(b)fluoranthene	22.47	252	550564	19.421	ng/ul 97
89) Benzo(k)fluoranthene	22.50	252	190895	7.111	ng/ul 98
91) Benzo(a)pyrene	22.99	252	264218	10.362	ng/ul 98
92) Indeno(1,2,3-cd)pyrene	25.12	276	115081	3.802	ng/ul 99
93) Dibenzo(a,h)anthracene	25.10	278	30993	1.197	ng/ul# 89
94) Benzo(g,h,i)perylene	25.73	276	66159	2.606	ng/ul 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030320\
Data File : BN010054.D
Acq On : 02 Mar 2020 23:38
Operator : CG/JU
Sample : L1615-08DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DBDN2DL

Manual Integrations
APPROVED

mohammad
3/4/2020 7:50:58 AM

Quant Time: Mar 03 03:04:06 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Mar 03 01:05:31 2020
Response via : Initial Calibration

