


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN030320\
 Data File : BN010056.D
 Acq On : 03 Mar 2020 11:10
 Operator : CG/JU
 Sample : DFTPP68
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP68

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Title : SVOA CALIBRATION
 Last Update : Tue Mar 03 05:36:52 2020



AutoFind: Scans 2422, 2423, 2424; Background Corrected with Scan 2417

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	69.1	135402	PASS
68	69	0.00	2	1.6	1634	PASS
69	198	0.00	100	51.1	100098	PASS
70	69	0.00	2	0.3	326	PASS
127	198	10	80	50.5	98986	PASS
197	198	0.00	2	0.6	1170	PASS
198	198	100	100	100.0	195840	PASS
199	198	5	9	6.6	12997	PASS
275	198	10	60	24.8	48512	PASS
365	198	1	100	3.5	6916	PASS
441	198	0.01	100	9.4	18416	PASS
442	198	50	100	56.9	111485	PASS
443	442	15	24	23.0	25659	PASS

Scan 2417 (17.204 min): BN010056.D
DFTPP68

m/z	Abundance
39.90	916.0
44.00	2335.0
48.90	377.0
73.00	508.0
207.00	540.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

m/z	Abundance
37.00	546.0
38.00	2235.0
39.00	11204.0
40.00	1504.0
44.00	2052.0
44.90	383.0
49.10	1108.0
50.00	34928.0
51.10	135680.0
52.00	6458.0
52.90	318.0
55.00	775.0
56.00	3798.0
57.00	8784.0
58.10	440.0
61.00	2275.0
62.00	1867.0
63.00	4847.0
64.00	755.0
65.00	2102.0
68.00	1533.0
69.00	98280.0
70.00	365.0
73.00	1053.0
74.00	11190.0
75.00	15341.0
76.00	5410.0
77.00	105880.0
78.10	7676.0
79.00	7758.0
80.00	5353.0
81.00	7636.0
81.90	2048.0
82.90	1803.0
84.90	1720.0
86.00	1998.0
87.00	1162.0
91.00	2242.0
92.00	2427.0
93.00	14057.0
94.00	633.0
96.00	501.0
98.00	10469.0
99.00	7248.0
100.00	753.0
101.00	4041.0
103.00	1067.0
103.90	2711.0
104.90	2888.0
106.00	711.0
107.00	32184.0
108.00	4762.0
109.10	1451.0
110.00	49152.0
110.90	7934.0
112.00	870.0
116.00	1737.0
117.00	29552.0
117.90	2377.0
121.90	2350.0
122.90	2874.0
123.90	1748.0
124.90	1546.0
126.10	518.0
127.00	92072.0
128.00	8066.0
129.00	38872.0
130.00	3109.0
130.90	814.0
132.00	361.0
133.90	1338.0
135.00	3705.0
135.90	1584.0
137.00	2196.0
140.00	390.0
141.00	5247.0
141.90	1415.0
142.90	921.0
145.90	1149.0
146.90	2484.0
147.90	6649.0
148.90	1632.0
150.00	414.0
151.00	710.0
151.80	501.0
152.90	1949.0
153.90	1220.0
154.90	2473.0
156.00	3795.0
157.00	636.0
157.90	877.0
159.00	905.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

159.90	1817.0
161.00	2280.0
161.90	784.0
164.90	1638.0
166.00	1698.0
166.90	9955.0
168.00	6026.0
169.10	1197.0
170.00	337.0
171.00	388.0
171.90	1097.0
173.00	1059.0
173.90	1988.0
174.90	3610.0
176.00	1291.0
176.80	1760.0
178.10	477.0
178.90	6088.0
180.00	5551.0
181.00	2106.0
182.00	312.0
183.90	835.0
185.00	3761.0
186.00	24904.0
187.00	7687.0
188.00	1039.0
188.90	1794.0
190.00	315.0
190.90	897.0
191.90	1868.0
193.00	2611.0
194.00	412.0
194.90	333.0
195.90	4476.0
197.10	958.0
197.90	165312.0
198.90	11009.0
200.10	1062.0
200.90	376.0
201.60	656.0
203.00	1522.0
204.00	6769.0
205.00	10748.0
206.00	46376.0
207.00	6115.0
207.90	1746.0
209.00	773.0
210.00	686.0
211.00	1977.0
211.80	472.0
215.00	609.0
215.90	930.0
216.90	12507.0
218.00	2015.0
221.00	5985.0
221.80	1622.0
223.00	2341.0
224.00	23696.0
225.00	6758.0
226.00	621.0
227.00	12083.0
227.90	1668.0
229.00	2723.0
229.90	439.0
230.90	834.0
233.90	485.0
234.90	896.0
235.90	560.0
236.90	1251.0
240.80	514.0
241.90	1308.0
243.00	1301.0
244.00	17344.0
245.10	1999.0
245.90	3754.0
246.90	1072.0
248.90	603.0
253.00	579.0
253.90	581.0
255.00	95904.0
256.00	14236.0
256.90	1738.0
257.90	5913.0
259.00	1089.0
264.90	2396.0
265.90	404.0
272.00	373.0
272.90	2678.0
274.00	7065.0
275.00	36936.0
276.00	5644.0
276.90	3725.0
277.90	665.0
282.90	436.0
283.90	472.0
284.90	465.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

293.10	913.0
296.00	12434.0
296.90	1348.0
301.90	460.0
303.00	1298.0
303.90	507.0
313.80	433.0
314.30	367.0
315.00	994.0
316.00	805.0
321.00	373.0
323.00	2773.0
323.90	475.0
327.00	831.0
327.90	303.0
332.00	315.0
333.00	335.0
333.90	1749.0
335.00	682.0
340.90	304.0
345.90	341.0
351.90	973.0
352.90	550.0
354.00	867.0
365.00	4030.0
365.90	703.0
372.00	1362.0
373.00	629.0
383.10	486.0
402.00	822.0
402.90	967.0
420.90	664.0
422.00	740.0
423.00	4500.0
423.90	1893.0
424.80	387.0
441.00	12186.0
442.00	74840.0
443.00	14361.0
443.90	1741.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

m/z	Abundance
37.00	698.0
38.00	1939.0
39.10	12249.0
40.00	1252.0
40.90	372.0
44.00	2483.0
45.00	389.0
49.10	1542.0
50.00	40480.0
51.00	160320.0
52.00	8513.0
55.00	885.0
56.00	4394.0
57.00	11544.0
57.90	328.0
61.00	1989.0
62.00	2103.0
63.00	5784.0
64.00	1081.0
65.00	2460.0
68.10	1896.0
69.00	122344.0
70.00	615.0
72.90	1533.0
74.00	13294.0
75.00	19776.0
76.10	6343.0
77.00	132160.0
78.00	9790.0
79.00	9784.0
80.00	7910.0
81.00	9723.0
81.90	2189.0
83.00	2180.0
83.90	307.0
84.90	1712.0
86.00	3289.0
87.00	1433.0
88.00	493.0
91.00	2388.0
92.00	2678.0
93.00	17096.0
94.00	1092.0
96.00	516.0
98.00	12266.0
99.00	9548.0
100.00	1041.0
101.00	4935.0
103.10	1634.0
104.00	3791.0
104.90	2797.0
106.00	979.0
107.00	38280.0
108.00	5538.0
109.00	1188.0
110.00	64752.0
111.00	10437.0
112.00	1486.0
112.90	485.0
116.10	2446.0
117.00	37584.0
118.00	2548.0
119.80	463.0
122.00	3091.0
123.00	3779.0
124.00	1963.0
125.00	1495.0
126.10	524.0
127.00	118472.0
128.00	9128.0
129.00	50712.0
130.00	3339.0
131.00	981.0
131.80	327.0
133.90	1197.0
135.00	4050.0
135.90	1575.0
137.10	2398.0
137.90	410.0
138.90	340.0
140.00	650.0
141.00	6565.0
142.00	1397.0
143.00	1434.0
144.00	439.0
144.90	444.0
146.00	1263.0
147.00	3675.0
147.90	9182.0
149.00	1977.0
149.90	540.0
151.10	845.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

151.80	481.0
152.80	1789.0
154.00	1422.0
155.00	3453.0
156.00	5245.0
156.90	856.0
157.90	862.0
158.90	998.0
160.00	1929.0
161.00	3195.0
162.00	754.0
163.00	366.0
164.00	393.0
165.00	2396.0
165.90	1870.0
167.00	13537.0
168.00	7784.0
168.90	916.0
170.00	519.0
170.90	480.0
172.00	1317.0
172.90	1250.0
173.90	2202.0
175.00	4433.0
176.00	1163.0
176.90	2808.0
177.90	556.0
179.00	9149.0
180.00	6503.0
180.90	3509.0
182.00	627.0
182.90	362.0
183.90	957.0
184.90	5218.0
186.00	36032.0
187.00	10612.0
187.90	1129.0
188.90	2745.0
189.90	539.0
191.00	1221.0
192.00	3687.0
193.00	3993.0
193.90	705.0
196.00	7281.0
197.00	1337.0
197.90	239808.0
198.90	15188.0
200.00	1447.0
200.80	415.0
201.40	1093.0
202.00	317.0
203.00	1715.0
204.00	8514.0
205.00	17544.0
206.00	65000.0
207.00	9307.0
208.00	2663.0
209.00	1095.0
210.00	1293.0
211.00	2172.0
212.00	429.0
214.90	839.0
215.90	1683.0
216.90	18328.0
218.00	1893.0
220.00	333.0
220.90	6875.0
222.00	2011.0
223.00	3295.0
224.00	37072.0
225.00	8613.0
225.90	1174.0
227.00	17112.0
228.00	2132.0
229.00	3776.0
231.00	1581.0
233.90	1025.0
234.90	1295.0
235.90	763.0
236.90	1550.0
238.90	845.0
239.90	554.0
240.90	984.0
242.10	1828.0
243.00	1927.0
243.90	26600.0
245.10	3784.0
246.00	5815.0
247.00	1183.0
249.00	997.0
250.90	555.0
253.00	643.0
255.00	140288.0
256.00	18872.0
257.00	1594.0
258.00	7575.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

259.00	1481.0
265.00	3462.0
266.00	584.0
270.80	374.0
271.90	460.0
273.00	4365.0
273.90	10068.0
275.00	60640.0
276.00	8150.0
277.00	5091.0
278.00	1116.0
282.90	639.0
284.00	385.0
284.90	1210.0
292.00	356.0
292.90	1199.0
296.00	18344.0
296.90	2727.0
301.00	386.0
303.00	1807.0
304.10	417.0
307.90	347.0
314.00	1046.0
315.00	2242.0
316.00	846.0
320.80	774.0
323.00	5258.0
324.00	974.0
327.00	1003.0
328.00	385.0
331.90	502.0
332.80	386.0
334.00	3726.0
335.10	778.0
341.00	701.0
342.00	337.0
346.00	1062.0
346.80	359.0
352.00	1546.0
353.10	1041.0
354.00	2289.0
365.00	8343.0
365.90	1161.0
370.90	360.0
372.00	2548.0
373.00	679.0
382.90	602.0
402.00	980.0
402.90	1643.0
404.00	747.0
421.00	1344.0
421.90	1193.0
422.90	7944.0
423.90	2794.0
441.00	21440.0
442.00	132032.0
443.00	31136.0
443.90	2324.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

m/z	Abundance
37.00	591.0
38.00	1322.0
39.10	8788.0
40.00	1257.0
41.00	399.0
44.00	1733.0
45.00	491.0
48.90	914.0
50.00	28832.0
51.00	110208.0
52.00	5694.0
53.00	316.0
55.00	585.0
55.90	3263.0
57.00	6762.0
61.00	1435.0
62.00	1220.0
63.00	3124.0
64.00	554.0
65.00	1671.0
68.00	1474.0
69.00	79672.0
73.00	1333.0
74.00	9756.0
75.00	12957.0
76.10	4599.0
77.00	88368.0
78.00	4928.0
79.00	7188.0
80.00	5088.0
81.00	6196.0
82.00	1566.0
83.00	1174.0
85.10	1347.0
86.00	1970.0
87.00	924.0
88.00	327.0
90.90	1525.0
92.00	2023.0
93.00	11323.0
93.90	808.0
95.90	501.0
98.00	8501.0
99.00	7205.0
100.00	509.0
101.00	3453.0
102.00	303.0
103.00	722.0
103.90	2030.0
105.00	2204.0
106.10	499.0
107.00	27728.0
108.00	4339.0
109.10	883.0
110.00	45736.0
110.90	7631.0
111.90	967.0
116.00	1820.0
117.00	25840.0
118.00	1904.0
119.90	415.0
122.00	2122.0
123.00	2548.0
123.90	1377.0
124.90	1443.0
127.00	86416.0
128.00	6906.0
129.00	36224.0
130.10	3345.0
130.90	636.0
131.90	355.0
132.60	323.0
133.90	872.0
135.00	3215.0
136.00	1158.0
137.00	1813.0
137.80	516.0
139.90	604.0
140.90	4933.0
142.00	1778.0
143.00	812.0
146.10	857.0
147.00	2310.0
148.00	6087.0
148.90	955.0
149.80	499.0
151.10	684.0
153.00	1701.0
154.00	1379.0
155.00	2892.0
156.00	3971.0
156.90	648.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

157.90	1213.0
159.00	777.0
159.90	1601.0
160.90	1834.0
162.10	594.0
165.00	1682.0
166.00	1622.0
167.00	10547.0
168.00	5920.0
168.90	1236.0
170.20	429.0
171.00	613.0
172.00	910.0
172.90	1103.0
173.90	2438.0
175.00	4290.0
175.90	924.0
177.00	2278.0
177.90	810.0
178.90	7639.0
180.00	5355.0
180.90	2192.0
182.10	301.0
183.80	722.0
185.00	3705.0
186.00	27104.0
187.00	8143.0
188.00	863.0
188.90	1885.0
189.90	358.0
191.00	838.0
191.90	2098.0
193.00	2411.0
194.00	725.0
194.90	461.0
195.90	5340.0
197.00	1217.0
197.90	182400.0
198.90	12794.0
200.00	1011.0
201.00	558.0
201.60	819.0
202.90	1695.0
204.00	7624.0
205.00	12098.0
206.00	50120.0
207.00	7592.0
208.00	1841.0
208.90	550.0
210.00	845.0
211.00	2033.0
212.00	309.0
215.00	757.0
215.90	999.0
216.90	14997.0
217.90	1659.0
221.00	7244.0
221.90	1383.0
223.00	2801.0
224.00	29200.0
225.00	8281.0
226.00	948.0
227.00	13404.0
227.90	2125.0
228.90	3020.0
230.00	365.0
231.00	1140.0
232.90	302.0
234.00	1131.0
234.90	1199.0
235.90	766.0
236.90	1199.0
239.00	649.0
239.90	408.0
240.80	628.0
242.00	1766.0
242.90	1585.0
244.00	22080.0
245.00	3007.0
246.00	4991.0
246.90	877.0
248.80	347.0
250.80	322.0
252.90	493.0
255.00	115128.0
256.00	18064.0
257.00	1203.0
258.00	7407.0
259.00	1023.0
261.00	378.0
265.00	2969.0
266.00	445.0
271.00	353.0
273.00	3632.0
274.00	9819.0
275.00	47960.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68

276.00	6655.0
276.90	4715.0
278.00	752.0
282.90	568.0
284.00	427.0
285.00	856.0
292.90	1159.0
295.90	16656.0
296.90	2208.0
303.00	1658.0
303.90	404.0
307.90	400.0
313.90	811.0
315.00	1441.0
316.00	905.0
317.00	361.0
320.90	606.0
323.00	3633.0
324.00	1020.0
326.90	856.0
327.90	418.0
332.00	400.0
332.90	505.0
334.00	3489.0
335.00	895.0
341.00	346.0
345.90	1270.0
352.00	1356.0
352.90	1184.0
354.00	1459.0
355.00	366.0
364.90	8376.0
365.80	1040.0
371.00	402.0
371.90	2799.0
373.00	665.0
383.00	755.0
390.00	528.0
401.90	992.0
403.00	1518.0
403.80	499.0
421.00	1470.0
421.90	1047.0
423.00	9031.0
423.90	2017.0
440.90	21624.0
442.00	127584.0
443.00	31480.0
444.10	2240.0

Instrument :
BNA_N
ClientSampleId :
DFTPP68