

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030323\
 Data File : BN024120.D
 Acq On : 03 Mar 2023 09:50
 Operator : CG/JU
 Sample : SSTD01078
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246

Quant Time: Mar 03 23:06:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 03 23:03:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.057	152	258608	20.000	ng/u1	0.00
20) Naphthalene-d8	10.881	136	1099802	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.698	164	673272	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1440331	20.000	ng/u1	0.00
79) Chrysene-d12	21.633	240	1337771	20.000	ng/u1	0.00
88) Perylene-d12	24.163	264	1502914	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	26874	4.240	ng/uL	0.00
4) Pyridine-d5	3.834	84	178575	10.677	ng/u1	0.00
7) Phenol-d5	7.199	99	217371	10.445	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.381	67	147311	11.538	ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	169217	9.750	ng/u1	0.00
15) 4-Methylphenol-d8	8.757	113	169146	10.409	ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	60504	7.716	ng/u1	0.00
24) 2-Nitrophenol-d4	9.952	143	65135	7.679	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	158525	8.883	ng/u1	0.00
31) 4-Chloroaniline-d4	11.010	131	241596	10.193	ng/u1	0.00
46) Dimethylphthalate-d6	14.098	166	515039	10.067	ng/u1	0.00
49) Acenaphthylene-d8	14.392	160	599322	9.666	ng/u1	0.00
54) 4-Nitrophenol-d4	14.869	143	84200	9.439	ng/u1	0.00
60) Fluorene-d10	15.687	176	429633	9.665	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.792	200	52171	7.471	ng/u1	0.00
73) Anthracene-d10	17.539	188	677668	9.928	ng/u1	0.00
81) Pyrene-d10	19.827	212	767265	9.760	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.986	264	697911	8.850	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	27922	4.049	ng/uL	94
5) Pyridine	3.858	79	188888	11.053	ng/u1	99
6) Benzaldehyde	7.187	77	134349	12.933	ng/u1	98
8) Phenol	7.228	94	231874	10.934	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.475	93	193191	11.289	ng/u1	99
12) 2-Chlorophenol	7.616	128	178960	10.191	ng/u1	99
13) 2-Methylphenol	8.493	108	167099	10.434	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.593	45	305544	13.083	ng/u1	99
16) Acetophenone	8.887	105	292246	11.921	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.869	70	135416	10.990	ng/u1	98
18) 4-Methylphenol	8.822	108	185570	10.899	ng/u1	96
19) Hexachloroethane	9.152	117	69041	9.738	ng/u1	97
22) Nitrobenzene	9.269	77	182942	9.470	ng/u1	99
23) Isophorone	9.793	82	414805	10.485	ng/u1	99
25) 2-Nitrophenol	9.987	139	81735	8.707	ng/u1	98
26) 2,4-Dimethylphenol	10.040	107	198783	9.834	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.281	93	257261	10.658	ng/u1	100
29) 2,4-Dichlorophenol	10.516	162	163661	9.325	ng/u1	99
30) Naphthalene	10.928	128	597697	9.965	ng/u1	100
32) 4-Chloroaniline	11.034	127	248341	10.514	ng/u1	97
33) Hexachlorobutadiene	11.222	225	102818	8.536	ng/u1	91
34) Caprolactam	11.793	113	49335	9.466	ng/u1	99
35) 4-Chloro-3-methylphenol	12.146	107	181027	10.144	ng/u1	99
36) 2-Methylnaphthalene	12.534	142	394764	10.178	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030323\
 Data File : BN024120.D
 Acq On : 03 Mar 2023 09:50
 Operator : CG/JU
 Sample : SST01078
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SST010246

Quant Time: Mar 03 23:06:27 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-BN030323.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 03 23:03:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.751	142	399767	10.424	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.893	216	201856	8.744	ng/ul	98
40) Hexachlorocyclopentadiene	12.881	237	84808	6.938	ng/ul	97
41) 2,4,6-Trichlorophenol	13.128	196	129238	8.709	ng/ul	97
42) 2,4,5-Trichlorophenol	13.193	196	136192	8.592	ng/ul	96
43) 1,1'-Biphenyl	13.534	154	531397	9.696	ng/ul	99
44) 2-Chloronaphthalene	13.575	162	413538	9.544	ng/ul	98
45) 2-Nitroaniline	13.775	65	70482	8.804	ng/ul	92
47) Dimethylphthalate	14.145	163	519079	10.198	ng/ul	99
48) 2,6-Dinitrotoluene	14.269	165	59751	8.584	ng/ul	92
50) Acenaphthylene	14.416	152	656499	10.008	ng/ul	99
51) 3-Nitroaniline	14.592	138	75210	9.043	ng/ul	99
52) Acenaphthene	14.757	153	450579	10.103	ng/ul	99
53) 2,4-Dinitrophenol	14.792	184	30018	6.248	ng/ul	97
55) 4-Nitrophenol	14.881	109	67850	9.672	ng/ul	96
56) Dibenzofuran	15.092	168	634884	10.249	ng/ul	95
57) 2,4-Dinitrotoluene	15.045	165	102123	9.916	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.310	232	120399	9.131	ng/ul	96
59) Diethylphthalate	15.510	149	510275	10.398	ng/ul	99
61) Fluorene	15.739	166	511040	10.432	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.734	204	242729	9.776	ng/ul	97
63) 4-Nitroaniline	15.751	138	78456	10.580	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.804	198	58619	8.229	ng/ul#	94
67) N-Nitrosodiphenylamine	15.945	169	420578	9.782	ng/ul	98
68) 4-Bromophenyl-phenylether	16.628	248	138840	8.598	ng/ul	99
69) Hexachlorobenzene	16.745	284	172979	9.239	ng/ul	99
70) Atrazine	16.892	200	130046	8.961	ng/ul	98
71) Pentachlorophenol	17.081	266	93772	8.009	ng/ul	95
72) Phenanthrene	17.481	178	794938	9.964	ng/ul	98
74) Anthracene	17.575	178	811626	10.091	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.498	216	204465	8.352	ng/uL	91
76) Pentachlorobenzene	15.010	250	204334	8.992	ng/uL	91
77) Carbazole	17.839	167	734691	10.591	ng/ul	99
78) Di-n-butylphthalate	18.410	149	794040	10.013	ng/ul	100
80) Fluoranthene	19.492	202	867021	9.332	ng/ul	99
82) Pyrene	19.857	202	958837	10.016	ng/ul	98
83) Butylbenzylphthalate	20.751	149	186708	12.683	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.545	252	211835	7.616	ng/ul	98
85) Benzo(a)anthracene	21.616	228	940641	10.012	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.539	149	327762	12.103	ng/ul	99
87) Chrysene	21.674	228	896521	10.013	ng/ul	93
89) Di-n-octyl phthalate	22.510	149	540383	8.643	ng/ul	100
90) Benzo(b)fluoranthene	23.386	252	912397	9.481	ng/ul	99
91) Benzo(k)fluoranthene	23.439	252	910535	9.713	ng/ul	97
93) Benzo(a)pyrene	24.039	252	811589	9.382	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	26.815	276	992047	8.647	ng/ul	98
95) Dibenzo(a,h)anthracene	26.839	278	793988	8.446	ng/ul	99
96) Benzo(g,h,i)perylene	27.621	276	424242	4.480	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 4.049 ng/uL

RT: 3.446 min Scan# 61

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion: 88 Resp: 27922

Ion Ratio Lower Upper

88 100

43 33.1 26.2 39.4

58 74.2 64.9 97.3



#5

Pyridine

Concen: 11.053 ng/uL

RT: 3.858 min Scan# 131

Delta R.T. 0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion: 79 Resp: 188888


Ion Ratio Lower Upper

79 100

52 83.3 65.1 97.7

51 39.2 31.4 47.0









#6
Benzaldehyde
Concen: 12.933 ng/ul
RT: 7.187 min Scan# 697
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion: 77 Resp: 134349
Ion Ratio Lower Upper
77 100
105 96.6 79.2 118.8
106 95.6 75.9 113.9



#8
Phenol
Concen: 10.934 ng/ul
RT: 7.228 min Scan# 704
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion: 94 Resp: 231874
Ion Ratio Lower Upper
94 100
65 27.8 22.9 34.3
66 36.0 29.0 43.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 11.289 ng/ul
 RT: 7.475 min Scan# 746
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50

Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246

Tgt Ion: 93 Resp: 193191




Ion	Ratio	Lower	Upper
93	100		
63	78.7	62.6	93.8
95	30.7	25.8	38.6




#12
 2-Chlorophenol
 Concen: 10.191 ng/ul
 RT: 7.616 min Scan# 770
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50

Tgt Ion: 128 Resp: 178960

Ion	Ratio	Lower	Upper
128	100		
64	49.3	39.3	58.9
130	31.9	25.1	37.7





#13

2-Methylphenol

Concen: 10.434 ng/ul

RT: 8.493 min Scan# 91

Delta R.T. -0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :





SSTD010246

Tgt Ion:108 Resp: 167099

Ion Ratio Lower Upper

108 100

107 87.7 72.3 108.5



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.083 ng/ul

RT: 8.593 min Scan# 936

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion: 45 Resp: 305544


Ion Ratio Lower Upper

45 100

77 12.4 9.8 14.6

79 8.9 7.4 11.0









#16
Acetophenone
Concen: 11.921 ng/ul
RT: 8.887 min Scan# 986
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:105 Resp: 292246
Ion Ratio Lower Upper
105 100
77 79.3 61.4 92.0
51 31.0 23.8 35.8



#17
N-Nitroso-di-n-propylamine
Concen: 10.990 ng/ul
RT: 8.869 min Scan# 983
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion: 70 Resp: 135416
Ion Ratio Lower Upper
70 100
42 64.1 50.8 76.2
101 8.9 7.2 10.8
130 17.4 15.8 23.8









#18
4-Methylphenol
Concen: 10.899 ng/ul
RT: 8.822 min Scan# 91
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:108 Resp: 185570
Ion Ratio Lower Upper
108 100
107 113.6 87.8 131.6



#19
Hexachloroethane
Concen: 9.738 ng/ul
RT: 9.152 min Scan# 1031
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:117 Resp: 69041
Ion Ratio Lower Upper
117 100
201 90.7 73.8 110.6
199 63.5 47.2 70.8









#22
Nitrobenzene
Concen: 9.470 ng/ul
RT: 9.269 min Scan# 110
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion: 77 Resp: 182942
Ion Ratio Lower Upper
77 100
123 43.4 34.8 52.2
65 12.9 11.0 16.4



#23
Isophorone
Concen: 10.485 ng/ul
RT: 9.793 min Scan# 1140
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion: 82 Resp: 414805
Ion Ratio Lower Upper
82 100
95 6.6 5.0 7.6
138 16.6 13.8 20.6









#25
2-Nitrophenol
Concen: 8.707 ng/ul
RT: 9.987 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:139 Resp: 81735
Ion Ratio Lower Upper
139 100
65 53.2 44.5 66.7
109 28.9 23.4 35.2



#26
2,4-Dimethylphenol
Concen: 9.834 ng/ul
RT: 10.040 min Scan# 1182
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:107 Resp: 198783
Ion Ratio Lower Upper
107 100
121 49.5 40.2 60.2
122 87.8 71.8 107.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 10.658 ng/ul
 RT: 10.281 min Scan# 1223
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246


Tgt Ion: 93 Resp: 257261
 Ion Ratio Lower Upper
 93 100
 95 31.7 25.4 38.2
 123 11.0 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 9.325 ng/ul
 RT: 10.516 min Scan# 1263
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50

Tgt Ion: 162 Resp: 163661
 Ion Ratio Lower Upper
 162 100
 164 65.1 52.6 78.8
 98 36.4 29.8 44.8









#30
Naphthalene
Concen: 9.965 ng/ul
RT: 10.928 min Scan# 1333
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:128 Resp: 597697
Ion Ratio Lower Upper
128 100
129 10.9 8.6 12.8
127 12.7 10.1 15.1



#32
4-Chloroaniline
Concen: 10.514 ng/ul
RT: 11.034 min Scan# 1351
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:127 Resp: 248341
Ion Ratio Lower Upper
127 100
129 33.2 25.1 37.7





#33

Hexachlorobutadiene

Concen: 8.536 ng/ul

RT: 11.222 min Scan# 1383

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:225 Resp: 102818

Ion Ratio Lower Upper

225 100

223 55.9 53.6 80.4

227 63.1 53.0 79.4



#34

Caprolactam

Concen: 9.466 ng/ul

RT: 11.793 min Scan# 1480

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion:113 Resp: 49335


Ion Ratio Lower Upper

113 100

55 196.1 155.4 233.2

56 139.3 111.0 166.6









#35
 4-Chloro-3-methylphenol
 Concen: 10.144 ng/ul
 RT: 12.146 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50




Instrument :
 BNA_N
 ClientSampleId :
 SST010246


Tgt Ion:107 Resp: 181027
 Ion Ratio Lower Upper
 107 100
 144 26.4 21.0 31.4
 142 82.7 65.7 98.5



#36
 2-Methylnaphthalene
 Concen: 10.178 ng/ul
 RT: 12.534 min Scan# 1606
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50

Tgt Ion:142 Resp: 394764
 Ion Ratio Lower Upper
 142 100
 141 85.3 68.2 102.4









#37
1-Methylnaphthalene
Concen: 10.424 ng/ul
RT: 12.751 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:142 Resp: 399767
Ion Ratio Lower Upper
142 100
141 91.6 75.5 113.3
116 3.6 3.1 4.7



#39
1,2,4,5-Tetrachlorobenzene
Concen: 8.744 ng/ul
RT: 12.893 min Scan# 1667
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:216 Resp: 201856
Ion Ratio Lower Upper
216 100
214 73.2 60.0 90.0
179 20.7 17.0 25.6
108 19.5 14.3 21.5









#40
Hexachlorocyclopentadiene
Concen: 6.938 ng/ul
RT: 12.881 min Scan# 1665
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:237 Resp: 84808
Ion Ratio Lower Upper
237 100
235 65.5 50.1 75.1
272 13.1 11.1 16.7



#41
2,4,6-Trichlorophenol
Concen: 8.709 ng/ul
RT: 13.128 min Scan# 1707
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:196 Resp: 129238
Ion Ratio Lower Upper
196 100
198 102.1 80.6 121.0
200 28.9 26.1 39.1





#42

2,4,5-Trichlorophenol

Concen: 8.592 ng/ul

RT: 13.193 min Scan# 1718

Delta R.T. -0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:196 Resp: 136192

Ion Ratio Lower Upper

196 100

198 95.8 79.7 119.5

200 29.6 26.1 39.1



#43

1,1'-Biphenyl

Concen: 9.696 ng/ul

RT: 13.534 min Scan# 1776

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion:154 Resp: 531397


Ion Ratio Lower Upper

154 100

153 39.1 30.8 46.2

76 13.8 11.1 16.7





#44

2-Chloronaphthalene

Concen: 9.544 ng/ul

RT: 13.575 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:162 Resp: 413538

Ion Ratio Lower Upper

162 100

164 33.4 25.8 38.6

127 39.5 30.8 46.2



#45

2-Nitroaniline

Concen: 8.804 ng/ul

RT: 13.775 min Scan# 1817

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion: 65 Resp: 70482


Ion Ratio Lower Upper

65 100

92 64.1 53.8 80.8

138 94.9 84.1 126.1









#47
Dimethylphthalate
Concen: 10.198 ng/ul
RT: 14.145 min Scan# 1880
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:163 Resp: 519079
Ion Ratio Lower Upper
163 100
194 3.8 3.1 4.7
164 10.5 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 8.584 ng/ul
RT: 14.269 min Scan# 1901
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:165 Resp: 59751
Ion Ratio Lower Upper
165 100
89 57.6 52.6 78.8
121 22.9 19.4 29.2









#50
Acenaphthylene
Concen: 10.008 ng/ul
RT: 14.416 min Scan# 1926
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:152 Resp: 656499
Ion Ratio Lower Upper
152 100
151 19.7 15.6 23.4
153 13.0 10.1 15.1



#51
3-Nitroaniline
Concen: 9.043 ng/ul
RT: 14.592 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:138 Resp: 75210
Ion Ratio Lower Upper
138 100
108 12.4 9.6 14.4
92 117.5 95.3 142.9









#52
Acenaphthene
Concen: 10.103 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:153 Resp: 450579
Ion Ratio Lower Upper
153 100
152 49.8 39.7 59.5
154 89.0 70.6 106.0



#53
2,4-Dinitrophenol
Concen: 6.248 ng/ul
RT: 14.792 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:184 Resp: 30018
Ion Ratio Lower Upper
184 100
63 74.6 61.6 92.4
154 65.3 50.2 75.4









#55
4-Nitrophenol
Concen: 9.672 ng/ul
RT: 14.881 min Scan# 2005
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:109 Resp: 67850
Ion Ratio Lower Upper
109 100
139 135.7 115.0 172.4
65 143.7 115.7 173.5



#56
Dibenzofuran
Concen: 10.249 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:168 Resp: 634884
Ion Ratio Lower Upper
168 100
139 35.2 30.7 46.1









#57
2,4-Dinitrotoluene
Concen: 9.916 ng/ul
RT: 15.045 min Scan# 2034
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:165 Resp: 102123
Ion Ratio Lower Upper
165 100
63 50.5 38.5 57.7
182 2.8 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.131 ng/ul
RT: 15.310 min Scan# 2078
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:232 Resp: 120399
Ion Ratio Lower Upper
232 100
131 45.6 38.5 57.7
130 2.2 2.0 3.0
166 29.9 26.0 39.0









#59
Diethylphthalate
Concen: 10.398 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:149 Resp: 510275
Ion Ratio Lower Upper
149 100
177 21.6 16.9 25.3
150 12.8 9.7 14.5



#61
Fluorene
Concen: 10.432 ng/ul
RT: 15.739 min Scan# 2151
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:166 Resp: 511040
Ion Ratio Lower Upper
166 100
165 95.9 75.6 113.4
167 12.9 11.2 16.8





#62

4-Chlorophenyl-phenylether

Concen: 9.776 ng/ul

RT: 15.734 min Scan# 2150

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:204 Resp: 242729

Ion Ratio Lower Upper

204 100

206 32.3 27.1 40.7

141 61.3 47.4 71.0



#63

4-Nitroaniline

Concen: 10.580 ng/ul

RT: 15.751 min Scan# 2153

Delta R.T. -0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion:138 Resp: 78456


Ion Ratio Lower Upper

138 100

92 51.6 40.0 60.0

108 78.4 60.4 90.6









#66
4,6-Dinitro-2-methylphenol
Concen: 8.229 ng/ul
RT: 15.804 min Scan# 2163
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:198 Resp: 58619
Ion Ratio Lower Upper
198 100
182 5.1 3.4 5.0#
77 29.0 20.4 30.6



#67
N-Nitrosodiphenylamine
Concen: 9.782 ng/ul
RT: 15.945 min Scan# 2186
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:169 Resp: 420578
Ion Ratio Lower Upper
169 100
168 66.1 54.5 81.7
167 35.4 28.3 42.5









#68
4-Bromophenyl-phenylether
Concen: 8.598 ng/ul
RT: 16.628 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:248 Resp: 138840
Ion Ratio Lower Upper
248 100
250 95.7 76.6 115.0
141 74.6 61.4 92.0



#69
Hexachlorobenzene
Concen: 9.239 ng/ul
RT: 16.745 min Scan# 2322
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:284 Resp: 172979
Ion Ratio Lower Upper
284 100
142 35.0 29.0 43.6
249 31.0 25.0 37.6





#70

Atrazine

Concen: 8.961 ng/ul

RT: 16.892 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:200 Resp: 130046

Ion Ratio Lower Upper

200 100

173 28.0 20.7 31.1

215 49.6 38.8 58.2



#71

Pentachlorophenol

Concen: 8.009 ng/ul

RT: 17.081 min Scan# 2379

Delta R.T. -0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion:266 Resp: 93772


Ion Ratio Lower Upper

266 100

264 67.8 49.5 74.3

268 66.7 51.3 76.9









#72
Phenanthrene
Concen: 9.964 ng/ul
RT: 17.481 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:178 Resp: 794938
Ion Ratio Lower Upper
178 100
179 15.9 11.6 17.4
176 19.2 14.9 22.3



#74
Anthracene
Concen: 10.091 ng/ul
RT: 17.575 min Scan# 2463
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:178 Resp: 811626
Ion Ratio Lower Upper
178 100
179 15.2 11.9 17.9
176 18.8 14.6 21.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 8.352 ng/uL
RT: 13.498 min Scan# 1770
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:216 Resp: 204465
Ion Ratio Lower Upper
216 100
214 83.7 59.6 89.4
218 51.0 37.0 55.4



#76
Pentachlorobenzene
Concen: 8.992 ng/uL
RT: 15.010 min Scan# 2027
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:250 Resp: 204334
Ion Ratio Lower Upper
250 100
248 58.5 53.0 79.4
252 62.3 54.6 81.8









#77
 Carbazole
 Concen: 10.591 ng/ul
 RT: 17.839 min Scan# 2508
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246


Tgt Ion:167 Resp: 734691
 Ion Ratio Lower Upper
 167 100
 166 20.7 17.3 25.9
 139 13.3 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 10.013 ng/ul
 RT: 18.410 min Scan# 2605
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50

Tgt Ion:149 Resp: 794040
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.0
 104 5.0 4.0 6.0









#80
 Fluoranthene
 Concen: 9.332 ng/u1
 RT: 19.492 min Scan# 2789
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50




Instrument :
 BNA_N
 ClientSampleId :
 SSTD010246


Tgt Ion:202 Resp: 867021
 Ion Ratio Lower Upper
 202 100
 101 13.0 10.0 15.0
 100 10.0 7.6 11.4



#82
 Pyrene
 Concen: 10.016 ng/u1
 RT: 19.857 min Scan# 2851
 Delta R.T. 0.000 min
 Lab File: BN024120.D
 Acq: 03 Mar 2023 09:50

Tgt Ion:202 Resp: 958837
 Ion Ratio Lower Upper
 202 100
 101 14.4 11.9 17.9
 100 11.3 9.9 14.9









#83
Butylbenzylphthalate
Concen: 12.683 ng/ul
RT: 20.751 min Scan# 3003
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:149 Resp: 186708
Ion Ratio Lower Upper
149 100
91 81.0 62.2 93.2
206 19.8 16.4 24.6



#84
3,3'-Dichlorobenzidine
Concen: 7.616 ng/ul
RT: 21.545 min Scan# 3138
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:252 Resp: 211835
Ion Ratio Lower Upper
252 100
254 67.0 54.6 81.8
126 12.1 11.0 16.6









#85
Benzo(a)anthracene
Concen: 10.012 ng/ul
RT: 21.616 min Scan# 3150
Delta R.T. 0.000 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50




Instrument :
BNA_N
ClientSampleId :
SSTD010246


Tgt Ion:228 Resp: 940641
Ion Ratio Lower Upper
228 100
229 17.4 16.3 24.5
226 26.8 22.6 34.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 12.103 ng/ul
RT: 21.539 min Scan# 3137
Delta R.T. -0.006 min
Lab File: BN024120.D
Acq: 03 Mar 2023 09:50

Tgt Ion:149 Resp: 327762
Ion Ratio Lower Upper
149 100
167 27.2 21.6 32.4
279 3.9 3.4 5.2





#87

Chrysene

Concen: 10.013 ng/ul

RT: 21.674 min Scan# 3160

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:228 Resp: 896521

Ion Ratio Lower Upper

228 100

226 28.8 25.8 38.6

229 18.3 17.5 26.3



#89

Di-n-octyl phthalate

Concen: 8.643 ng/ul




RT: 22.510 min Scan# 3302


Delta R.T. -0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Tgt Ion:149 Resp: 540383





#90

Benzo(b)fluoranthene

Concen: 9.481 ng/ul

RT: 23.386 min Scan# 3451

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:252 Resp: 912397

Ion Ratio Lower Upper

252 100

253 22.0 18.3 27.5

125 11.6 9.5 14.3



#91

Benzo(k)fluoranthene

Concen: 9.713 ng/ul

RT: 23.439 min Scan# 3460

Delta R.T. -0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion:252 Resp: 910535


Ion Ratio Lower Upper

252 100

253 20.0 17.4 26.0

125 11.0 9.0 13.6





#93

Benzo(a)pyrene

Concen: 9.382 ng/ul

RT: 24.039 min Scan# 3

Delta R.T. -0.006 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:252 Resp: 811589

Ion Ratio Lower Upper

252 100

253 23.1 17.0 25.4

125 12.7 9.5 14.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 8.647 ng/ul

RT: 26.815 min Scan# 4034

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50




Tgt Ion:276 Resp: 992047


Ion Ratio Lower Upper

276 100

138 24.5 19.6 29.4

277 22.7 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 8.446 ng/ul

RT: 26.839 min Scan# 410

Delta R.T. 0.000 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Instrument :

BNA_N

ClientSampleId :

SSTD010246





Tgt Ion:278 Resp: 793988

Ion Ratio Lower Upper

278 100

139 17.1 14.4 21.6

279 22.3 18.2 27.2



#96

Benzo(g,h,i)perylene

Concen: 4.480 ng/ul

RT: 27.621 min Scan# 4171

Delta R.T. -0.017 min

Lab File: BN024120.D

Acq: 03 Mar 2023 09:50

Tgt Ion:276 Resp: 424242

Ion Ratio Lower Upper

276 100

138 24.4 18.8 28.2

277 22.7 18.6 27.8

