

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN030620\
 Data File : BN010126.D
 Acq On : 06 Mar 2020 19:20
 Operator : CG/JU
 Sample : PB127346BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK46

Manual Integrations
 APPROVED

mohammad
 3/9/2020 3:05:18 PM

Quant Time: Mar 07 02:56:30 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 05 14:04:18 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.35	152	60508	20.00	ng/ul	0.00
18) Naphthalene-d8	10.10	136	261489	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.00	164	169503	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.76	188	360211	20.00	ng/ul	0.00
78) Chrysene-d12	20.99	240	304728	20.00	ng/ul	0.00
86) Perylene-d12	23.08	264	345789	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.00	96	10130	6.89	ng/uL	0.00
5) Phenol-d5	6.56	99	164841	32.03	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.71	67	121309	34.16	ng/ul	0.00
9) 2-Chlorophenol-d4	6.89	132	135477	34.96	ng/ul	0.00
13) 4-Methylphenol-d8	8.07	113	131773	32.19	ng/ul	0.00
19) Nitrobenzene-d5	8.50	128	64513	33.72	ng/ul	0.00
22) 2-Nitrophenol-d4	9.21	143	71782	34.30	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.75	165	128046m	31.83	ng/ul	0.00
29) 4-Chloroaniline-d4	10.26	131	182524	40.38	ng/ul	0.00
44) Dimethylphthalate-d6	13.43	166	447517	35.35	ng/ul	0.00
47) Acenaphthylene-d8	13.69	160	516079	34.61	ng/ul	0.00
52) 4-Nitrophenol-d4	14.26	143	54380m	23.52	ng/ul	0.00
58) Fluorene-d10	15.00	176	394913	36.13	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.18	200	52853	23.62	ng/ul	0.00
71) Anthracene-d10	16.86	188	600763	36.38	ng/ul	0.00
79) Pyrene-d10	19.17	212	656467	41.22	ng/ul	0.00
90) Benzo(a)pyrene-d12	22.95	264	644910	37.35	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030620\
Data File : BN010126.D
Acq On : 06 Mar 2020 19:20
Operator : CG/JU
Sample : PB127346BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK46

Quant Time: Mar 07 02:56:30 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 05 14:04:18 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

mohammad
3/9/2020 3:05:18 PM

