


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN030620\
 Data File : BN010115.D
 Acq On : 06 Mar 2020 10:14
 Operator : CG/JU
 Sample : DFTPP71
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP71

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN021220MA.M
 Title : SVOA CALIBRATION
 Last Update : Sat Mar 07 22:39:20 2020



AutoFind: Scans 2455, 2456, 2457; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	66.3	111490	PASS
68	69	0.00	2	1.9	1511	PASS
69	198	0.00	100	48.0	80674	PASS
70	69	0.00	2	0.3	279	PASS
127	198	10	80	48.8	82077	PASS
197	198	0.00	2	0.2	366	PASS
198	198	100	100	100.0	168106	PASS
199	198	5	9	6.8	11467	PASS
275	198	10	60	25.7	43170	PASS
365	198	1	100	3.6	6028	PASS
441	198	0.01	100	9.9	16579	PASS
442	198	50	100	66.6	111946	PASS
443	442	15	24	18.4	20576	PASS

m/z	Abundance
36.00	580.0
36.90	407.0
38.00	1334.0
39.00	7123.0
40.00	144512.0
44.00	2911.0
44.90	389.0
49.00	945.0
50.00	23040.0
51.00	85496.0
52.00	4157.0
55.10	334.0
56.00	2203.0
57.00	5623.0
61.00	1007.0
62.00	1358.0
63.00	2970.0
64.00	568.0
64.90	1548.0
68.00	1178.0
69.00	63064.0
73.00	1237.0
74.00	7092.0
75.00	11828.0
76.10	3004.0
77.00	71720.0
78.00	4408.0
79.00	5157.0
80.00	4044.0
81.00	4922.0
82.00	1228.0
83.00	1141.0
85.00	865.0
86.00	1912.0
86.90	804.0
91.00	1075.0
92.00	1291.0
93.00	8654.0
94.00	1070.0
96.00	486.0
98.00	7695.0
99.00	5670.0
99.80	384.0
101.00	2916.0
103.00	1317.0
104.00	1694.0
105.00	1778.0
106.00	664.0
107.00	23176.0
108.00	3158.0
108.90	615.0
110.00	37848.0
111.00	5900.0
112.00	486.0
116.10	1419.0
117.00	22464.0
118.00	1607.0
121.90	941.0
123.00	1862.0
123.90	1009.0
124.90	1035.0
126.10	308.0
127.00	69664.0
128.00	5804.0
129.00	28408.0
130.00	2506.0
130.90	419.0
132.10	363.0
133.90	487.0
135.00	2473.0
136.00	1193.0
137.00	1326.0
139.90	604.0
141.00	4312.0
142.00	964.0
142.80	807.0
146.00	724.0
147.00	2569.0
148.00	5103.0
148.90	1057.0
150.00	431.0
150.90	409.0
153.00	1322.0
154.00	1058.0
155.10	2536.0
156.00	3354.0
157.10	877.0
158.00	962.0
159.00	476.0
160.00	1465.0
160.90	1827.0
161.90	341.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

165.00	1848.0
166.00	1294.0
167.00	9409.0
168.00	4806.0
168.90	688.0
170.80	463.0
171.80	622.0
173.00	1114.0
174.00	1848.0
175.00	3264.0
176.00	713.0
176.90	1429.0
177.90	438.0
179.00	6811.0
180.00	4683.0
181.00	2344.0
183.90	463.0
185.00	3145.0
186.00	23144.0
187.00	6385.0
187.90	496.0
189.00	2262.0
189.90	332.0
191.00	938.0
191.90	1917.0
193.00	2214.0
194.00	319.0
196.00	4466.0
197.10	1098.0
197.90	155840.0
198.90	10642.0
199.90	872.0
201.30	431.0
201.70	489.0
202.90	1085.0
204.00	6726.0
205.00	10531.0
206.00	46392.0
207.00	5555.0
208.00	1818.0
208.90	615.0
210.00	665.0
210.90	1777.0
212.00	364.0
214.90	919.0
216.00	1121.0
216.90	11433.0
217.90	2109.0
219.80	385.0
221.00	6208.0
221.80	1432.0
223.00	2557.0
224.00	22912.0
225.00	6074.0
226.00	1104.0
226.90	10679.0
227.90	2310.0
228.90	2312.0
230.00	366.0
231.00	971.0
233.90	1018.0
235.00	768.0
235.90	766.0
237.00	1087.0
238.90	471.0
240.00	371.0
241.00	599.0
241.90	1467.0
243.00	1261.0
244.00	18856.0
245.00	2387.0
246.00	3890.0
246.90	856.0
248.80	500.0
253.10	586.0
254.00	1082.0
255.00	99064.0
256.00	13985.0
256.90	837.0
258.00	7298.0
258.90	1053.0
264.90	2113.0
265.90	507.0
272.00	357.0
273.00	3189.0
274.00	9065.0
275.00	44536.0
276.00	5127.0
276.90	3650.0
277.90	715.0
282.90	486.0
284.90	681.0
293.00	719.0
294.00	350.0
296.00	12432.0
296.90	2012.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

303.00	1738.0
303.90	335.0
314.10	472.0
315.00	1610.0
316.00	941.0
320.90	434.0
323.00	4827.0
324.00	719.0
326.80	960.0
327.90	399.0
331.90	395.0
332.80	519.0
334.00	2541.0
334.90	684.0
341.00	420.0
345.90	815.0
351.90	1322.0
352.90	629.0
354.00	1091.0
364.90	7065.0
365.90	753.0
370.90	315.0
372.00	2155.0
373.00	560.0
383.00	555.0
401.90	860.0
402.90	1454.0
404.00	475.0
420.90	1181.0
422.00	883.0
422.90	7485.0
424.00	2418.0
424.80	324.0
441.00	19464.0
442.00	133696.0
442.90	25800.0
444.00	2472.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

m/z	Abundance
36.00	622.0
37.10	608.0
38.00	2010.0
39.10	10961.0
40.00	148416.0
41.00	453.0
44.00	2561.0
49.00	1591.0
50.10	31736.0
51.00	133824.0
52.00	6650.0
54.90	839.0
55.90	3122.0
57.00	8348.0
58.00	434.0
61.00	1982.0
61.90	1183.0
63.00	4736.0
63.90	656.0
65.00	2081.0
68.10	2022.0
69.00	94672.0
69.90	536.0
73.00	1112.0
74.00	10211.0
75.00	16154.0
76.00	4852.0
77.00	105688.0
78.00	6677.0
79.00	7837.0
80.00	5322.0
81.00	8431.0
82.00	1701.0
83.00	1234.0
84.90	1359.0
86.00	1962.0
87.00	721.0
88.00	340.0
91.00	1738.0
91.90	2013.0
93.00	14740.0
94.00	923.0
95.00	317.0
95.90	590.0
98.00	11608.0
99.00	7824.0
99.80	404.0
101.00	4267.0
102.90	1324.0
103.90	2884.0
104.90	2500.0
106.00	1337.0
107.00	32384.0
108.00	4779.0
109.00	826.0
110.00	50504.0
111.00	8611.0
112.00	1312.0
113.00	549.0
116.10	1690.0
117.00	32032.0
118.00	1968.0
120.00	410.0
122.00	2126.0
123.00	3271.0
123.90	1802.0
124.90	1471.0
127.00	99176.0
128.00	7078.0
129.00	40512.0
129.90	3388.0
131.00	955.0
134.00	1513.0
134.90	3857.0
136.00	1001.0
137.00	1623.0
138.00	396.0
138.80	333.0
140.00	319.0
141.00	5167.0
141.90	1946.0
142.90	1174.0
144.10	386.0
144.90	333.0
146.00	1110.0
147.00	3265.0
148.00	6549.0
149.00	1443.0
150.00	548.0
151.20	466.0
151.90	356.0
153.10	1697.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

153.90	1339.0
155.00	3112.0
156.00	5049.0
157.00	763.0
157.90	832.0
158.90	838.0
160.00	1861.0
161.00	2386.0
161.90	798.0
162.90	351.0
164.90	1991.0
165.90	2027.0
167.00	10831.0
167.90	6209.0
169.00	944.0
169.90	596.0
171.00	662.0
172.00	1471.0
173.00	1295.0
174.00	2467.0
175.00	4790.0
176.00	1099.0
176.90	2416.0
177.80	764.0
178.90	8177.0
180.00	5868.0
181.00	2715.0
182.10	459.0
183.00	365.0
183.90	771.0
185.00	4196.0
186.00	30624.0
187.00	7486.0
187.90	1106.0
188.90	2543.0
189.90	308.0
190.90	902.0
191.90	2827.0
193.00	2757.0
194.00	649.0
196.00	5986.0
197.90	208832.0
199.00	13752.0
200.00	1493.0
200.90	623.0
201.50	852.0
203.00	1803.0
204.00	7181.0
205.00	12902.0
206.00	56736.0
207.00	8449.0
208.00	2224.0
209.10	922.0
210.00	762.0
211.00	2205.0
214.90	921.0
216.00	1676.0
216.90	15244.0
218.10	1690.0
219.80	329.0
220.20	521.0
221.10	6132.0
222.00	1815.0
223.00	2792.0
224.00	30848.0
225.00	7478.0
226.00	889.0
227.00	13603.0
227.90	2466.0
229.00	2232.0
230.00	556.0
231.00	1009.0
233.90	1147.0
234.90	733.0
235.80	624.0
237.00	1323.0
239.00	803.0
239.90	320.0
240.90	740.0
242.00	1045.0
242.90	1364.0
244.00	22040.0
244.90	3073.0
245.90	3785.0
246.90	1210.0
249.00	787.0
252.90	351.0
254.00	792.0
255.00	118960.0
256.00	17408.0
257.00	1180.0
257.90	7032.0
258.90	1031.0
265.00	2756.0
272.90	3288.0
274.00	7774.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

275.00	49568.0
276.00	7105.0
277.00	4677.0
278.00	1153.0
282.00	336.0
282.90	564.0
284.20	341.0
285.00	651.0
292.90	848.0
295.10	702.0
296.00	16240.0
297.00	1886.0
302.00	383.0
302.90	1862.0
304.00	459.0
314.00	688.0
314.90	1758.0
315.90	914.0
321.10	395.0
322.00	314.0
323.00	4790.0
323.90	882.0
326.90	1136.0
328.10	431.0
331.80	349.0
332.80	546.0
334.10	3381.0
334.80	512.0
341.00	623.0
346.00	1167.0
346.90	417.0
351.90	1393.0
353.00	1271.0
354.00	1372.0
364.90	5990.0
366.00	869.0
370.80	332.0
372.10	1909.0
373.00	396.0
382.80	618.0
390.00	455.0
401.90	1086.0
403.00	1145.0
403.90	678.0
421.00	1465.0
421.90	1217.0
423.00	7511.0
424.00	2557.0
425.00	346.0
440.90	19632.0
442.00	131392.0
443.00	22856.0
444.00	1949.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

m/z	Abundance
35.90	625.0
37.00	544.0
38.00	1377.0
39.00	8534.0
40.00	150208.0
41.00	339.0
44.00	3062.0
45.00	414.0
49.00	1281.0
50.00	29136.0
51.00	115152.0
52.00	5606.0
54.90	424.0
55.90	3341.0
57.00	7668.0
61.00	1481.0
62.00	1557.0
63.00	3875.0
63.90	525.0
65.10	1934.0
68.10	1333.0
69.00	84288.0
70.00	301.0
73.00	1393.0
74.00	8579.0
75.00	13663.0
76.00	3972.0
77.00	88536.0
78.00	5639.0
79.00	5749.0
80.00	5591.0
81.00	6365.0
82.00	1341.0
83.00	1447.0
85.00	1175.0
86.00	2423.0
87.00	977.0
87.90	438.0
91.00	1868.0
92.00	1360.0
93.00	11143.0
93.90	677.0
96.10	476.0
98.00	8608.0
99.00	6999.0
99.90	564.0
101.00	3372.0
103.00	890.0
104.00	2515.0
105.00	1960.0
106.00	618.0
107.00	26288.0
108.00	3882.0
110.00	41232.0
111.00	7597.0
112.00	800.0
113.00	356.0
116.00	1225.0
117.00	24768.0
118.00	2048.0
122.00	2357.0
123.00	2727.0
124.00	1266.0
124.90	997.0
127.00	77392.0
128.00	5733.0
129.00	33568.0
130.00	2823.0
131.00	565.0
131.90	434.0
133.90	938.0
135.00	2911.0
135.90	1079.0
137.10	1447.0
138.00	537.0
139.90	512.0
141.00	4753.0
141.90	1444.0
143.00	1326.0
145.90	743.0
147.00	1993.0
147.90	5698.0
148.90	1241.0
149.90	391.0
151.20	587.0
152.90	1785.0
154.00	1272.0
155.00	2428.0
156.00	3294.0
156.90	578.0
157.80	420.0
158.90	500.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

159.90	1506.0
161.00	2043.0
161.90	430.0
165.00	1340.0
166.00	1113.0
167.00	9342.0
168.00	5037.0
169.00	605.0
171.90	830.0
172.90	1001.0
174.00	1140.0
175.00	3763.0
176.00	453.0
177.10	1246.0
177.90	483.0
178.90	5932.0
180.00	4485.0
180.90	2082.0
182.00	336.0
185.00	2826.0
186.00	20360.0
187.00	6369.0
188.00	1359.0
189.00	1651.0
191.10	879.0
192.00	1925.0
193.00	2474.0
194.10	562.0
196.00	3765.0
196.90	1080.0
197.90	139648.0
198.90	10008.0
200.00	1174.0
201.30	774.0
202.90	1283.0
204.00	6062.0
205.00	9798.0
206.00	39416.0
207.00	5932.0
207.90	1757.0
208.80	375.0
209.90	588.0
210.90	1350.0
215.00	441.0
216.00	914.0
216.90	11821.0
218.00	1386.0
221.00	5856.0
221.90	1033.0
222.90	2289.0
224.00	20672.0
225.00	6328.0
226.00	549.0
227.00	9789.0
227.90	1563.0
228.90	2055.0
230.90	768.0
234.00	674.0
235.00	658.0
235.80	313.0
236.90	1064.0
239.00	674.0
240.90	536.0
242.00	1137.0
242.90	1379.0
244.00	15716.0
245.00	2007.0
245.90	3170.0
246.90	881.0
248.90	344.0
252.90	607.0
255.00	82880.0
256.00	12746.0
256.90	1205.0
258.00	4761.0
258.90	973.0
265.00	2141.0
270.90	329.0
272.90	2093.0
274.00	6091.0
275.00	35408.0
276.00	5058.0
277.00	3077.0
277.90	620.0
283.00	421.0
284.90	529.0
292.90	741.0
295.00	338.0
295.90	10600.0
297.00	1423.0
302.90	1084.0
303.90	412.0
314.10	577.0
314.90	947.0
315.90	532.0
321.00	353.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

323.00	2416.0
323.90	588.0
326.90	862.0
327.80	331.0
331.80	300.0
333.90	1755.0
335.00	603.0
341.00	417.0
345.80	328.0
352.00	876.0
353.00	706.0
354.00	1032.0
364.90	5031.0
365.90	671.0
372.00	1533.0
372.90	456.0
383.00	335.0
401.90	781.0
403.00	839.0
404.00	429.0
420.80	618.0
422.00	834.0
422.90	4400.0
424.00	1866.0
441.00	10643.0
442.00	70752.0
443.00	13074.0
443.90	1259.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71

Scan 2450 (17.198 min): BN010115.D
DFTPP71

m/z	Abundance
36.00	529.0
39.20	420.0
40.00	152768.0
44.00	2380.0
48.90	529.0
73.10	322.0
207.00	696.0
207.90	395.0

Instrument :
BNA_N
ClientSampleId :
DFTPP71