

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN030719\
 Data File : BN004649.D
 Acq On : 07 Mar 2019 22:24
 Operator : JU/SJ
 Sample : K1721-01
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BF8B7

Quant Time: Mar 08 01:22:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 08 00:36:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.72	152	36841	20.00	ng/ul	0.00
18) Naphthalene-d8	10.51	136	171930	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.36	164	107264	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.11	188	259585	20.00	ng/ul	0.00
78) Chrysene-d12	21.30	240	315746	20.00	ng/ul	0.00
86) Perylene-d12	23.56	264	319448	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.23	96	2883	4.41	ng/uL	0.00
5) Phenol-d5	6.88	99	18832	6.19	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.06	67	36512	22.99	ng/ul	0.00
9) 2-Chlorophenol-d4	7.25	132	54096	20.48	ng/ul	0.00
13) 4-Methylphenol-d8	8.42	113	37203	14.05	ng/ul	0.00
19) Nitrobenzene-d5	8.88	128	32038	24.44	ng/ul	0.00
22) 2-Nitrophenol-d4	9.60	143	37278	25.04	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.12	165	67720	23.87	ng/ul	0.00
29) 4-Chloroaniline-d4	10.65	131	3955	1.16	ng/ul	0.00
44) Dimethylphthalate-d6	13.77	166	247285	26.66	ng/ul	0.00
47) Acenaphthylene-d8	14.06	160	292183	26.17	ng/ul	0.00
52) 4-Nitrophenol-d4	14.56	143	9709	5.64	ng/ul	0.00
58) Fluorene-d10	15.36	176	209045	26.52	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.48	200	42783	23.87	ng/ul	0.00
71) Anthracene-d10	17.21	188	344763	27.27	ng/ul	0.00
79) Pyrene-d10	19.51	212	409749	26.56	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.42	264	440538	25.68	ng/ul	0.00

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN030719\
Data File : BN004649.D
Acq On : 07 Mar 2019 22:24
Operator : JU/SJ
Sample : K1721-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BF8B7

Quant Time: Mar 08 01:22:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 08 00:36:35 2019
Response via : Initial Calibration

