

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030722\
 Data File : BN018860.D
 Acq On : 07 Mar 2022 16:08
 Operator : CG/JU
 Sample : N1797-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-MW161S1-GW-20220303

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 03/08/2022
 Supervised By : Jagrut Upadhyay 03/08/2022

Quant Time: Mar 08 02:21:19 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 03 00:44:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.912	152	3410	0.400	ng	0.00
7) Naphthalene-d8	10.722	136	9831	0.400	ng	0.00
13) Acenaphthene-d10	14.538	164	6350	0.400	ng	0.00
19) Phenanthrene-d10	17.277	188	15171	0.400	ng	0.00
29) Chrysene-d12	21.458	240	16449	0.400	ng	# 0.00
35) Perylene-d12	23.855	264	14485	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.464	112	774	0.104	ng	0.00
5) Phenol-d6	7.067	99	601	0.067	ng	0.00
8) Nitrobenzene-d5	9.067	82	1950m	0.272	ng	0.00
11) 2-Methylnaphthalene-d10	12.306	152	4293	0.268	ng	0.00
14) 2,4,6-Tribromophenol	16.025	330	737	0.252	ng	0.00
15) 2-Fluorobiphenyl	13.169	172	7807	0.287	ng	0.00
27) Fluoranthene-d10	19.307	212	15913	0.356	ng	0.00
31) Terphenyl-d14	19.902	244	12971	0.369	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.312	88	992	0.260	ng	# 88
32) Benzo(a)anthracene	21.440	228	1404	0.023	ng	# 88
33) Chrysene	21.494	228	2294	0.034	ng	94
34) Bis(2-ethylhexyl)phtha...	21.369	149	13396	0.351	ng	# 98
36) Indeno(1,2,3-cd)pyrene	26.340	276	3118	0.048	ng	95
37) Benzo(b)fluoranthene	23.124	252	3702	0.055	ng	# 66
38) Benzo(k)fluoranthene	23.173	252	2711	0.041	ng	# 57
40) Dibenzo(a,h)anthracene	26.354	278	2173	0.044	ng	# 69
41) Benzo(g,h,i)perylene	27.100	276	2947	0.055	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030722\
Data File : BN018860.D
Acq On : 07 Mar 2022 16:08
Operator : CG/JU
Sample : N1797-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-TT-MW161S1-GW-20220303

Quant Time: Mar 08 02:21:19 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 03 00:44:34 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 03/08/2022
Supervised By : Jagrut Upadhyay 03/08/2022

