

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030722\
 Data File : BN018861.D
 Acq On : 07 Mar 2022 16:49
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 08 02:21:24 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN030322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 03 00:44:34 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	59	0.00
2	1,4-Dioxane	0.400	0.415	-3.7	60	0.00
3	n-Nitrosodimethylamine	0.400	0.459	-14.8	70	-0.01
4 S	2-Fluorophenol	0.400	0.447	-11.7	68	0.00
5 S	Phenol-d6	0.400	0.470	-17.5	74	0.00
6	bis(2-Chloroethyl)ether	0.400	0.424	-6.0	63	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	64	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	72	0.00
9	Naphthalene	0.400	0.388	3.0	64	0.00
10	Hexachlorobutadiene	0.400	0.377	5.8	61	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	66	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	65	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	68	0.00
14 S	2,4,6-Tribromophenol	0.400	0.410	-2.5	83	0.00
15 S	2-Fluorobiphenyl	0.400	0.381	4.8	66	0.00
16	Acenaphthylene	0.400	0.376	6.0	70	0.00
17	Acenaphthene	0.400	0.389	2.8	69	0.00
18	Fluorene	0.400	0.396	1.0	70	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.438	-9.5	93	0.00
21	4-Bromophenyl-phenylether	0.400	0.366	8.5	69	0.00
22	Hexachlorobenzene	0.400	0.349	12.8	64	0.00
23	Atrazine	0.400	0.397	0.8	83	0.00
24	Pentachlorophenol	0.400	0.532	-33.0#	109	0.00
25	Phenanthrene	0.400	0.393	1.8	73	-0.01
26	Anthracene	0.400	0.384	4.0	76	0.00
27 SURR	Fluoranthene-d10	0.400	0.402	-0.5	77	0.00
28	Fluoranthene	0.400	0.401	-0.3	77	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	93	0.00
30	Pyrene	0.400	0.331	17.3	80	0.00
31 S	Terphenyl-d14	0.400	0.340	15.0	83	0.00
32	Benzo(a)anthracene	0.400	0.391	2.3	97	0.00
33	Chrysene	0.400	0.402	-0.5	97	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.540	-35.0#	46	0.00
35 I	Perylene-d12	0.400	0.400	0.0	113	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.443	-10.7	135	-0.01
37	Benzo(b)fluoranthene	0.400	0.404	-1.0	120	0.00
38	Benzo(k)fluoranthene	0.400	0.388	3.0	114	0.00
39 C	Benzo(a)pyrene	0.400	0.383	4.3	115	0.00
40	Dibenzo(a,h)anthracene	0.400	0.447	-11.7	137	-0.01
41	Benzo(g,h,i)perylene	0.400	0.455	-13.7	138	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4