




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030821\
 Data File : BN013863.D
 Acq On : 08 Mar 2021 09:02
 Operator : CG/JU
 Sample : DFTTP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 08 10:14:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 05 13:12:06 2021
 Response via : Initial Calibration






TIC: BN013863.D\data.ms

(70) Pentachlorophenol (C)**16.745min (-0.006) 20010.47 ng****response 224092**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.47
264.00	61.60	61.44
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030821\
Data File : BN013863.D
Acq On : 08 Mar 2021 09:02
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 08 10:14:02 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 05 13:12:06 2021
Response via : Initial Calibration



TIC: BN013863.D\data.ms

(77) Benzidine**19.339min (-0.006) 0.00 ng****response 1566478**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.11
183.00	13.20	11.57
0.00	0.00	0.00