

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030821\
 Data File : BN013868.D
 Acq On : 08 Mar 2021 13:51
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Manual Integrations
 APPROVED

mohammad
 3/9/2021 12:19:50 PM

Quant Time: Mar 08 15:15:26 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 08 10:32:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.725	152	6826	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	25748	0.40	ng	#-0.01
13) Acenaphthene-d10	14.346	164	13917	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	30021	0.40	ng	# 0.00
29) Chrysene-d12	21.261	240	28636	0.40	ng	0.00
36) Perylene-d12	23.493	264	28157	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.317	112	6426	0.35	ng	0.00
5) Phenol-d6	6.887	99	8472	0.35	ng	0.00
8) Nitrobenzene-d5	8.857	82	6638	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.089	152	16437	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	1718	0.30	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	21781	0.44	ng	0.00
27) Fluoranthene-d10	19.119	212	33009	0.37	ng	0.00
31) Terphenyl-d14	19.720	244	26153	0.41	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.255	88	3550	0.43	ng	98
3) n-Nitrosodimethylamine	3.577	42	2408	0.38	ng	93
6) bis(2-Chloroethyl)ether	7.145	93	7705	0.45	ng	96
9) Naphthalene	10.543	128	26628	0.42	ng	100
10) Hexachlorobutadiene	10.835	225	5035	0.41	ng	# 99
12) 2-Methylnaphthalene	12.164	142	17754	0.40	ng	99
16) Acenaphthylene	14.067	152	21589	0.35	ng	100
17) Acenaphthene	14.410	154	15947	0.41	ng	99
18) Fluorene	15.399	166	19669	0.39	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	1168	0.56	ng	# 84
21) 4-Bromophenyl-phenylether	16.289	248	6419	0.39	ng	95
22) Hexachlorobenzene	16.398	284	7676	0.45	ng	99
23) Atrazine	16.556	200	3749	0.25	ng	96
24) Pentachlorophenol	16.739	266	1929	0.35	ng	97
25) Phenanthrene	17.128	178	33242	0.42	ng	100
26) Anthracene	17.226	178	26545	0.35	ng	99
28) Fluoranthene	19.149	202	36308	0.39	ng	99
30) Pyrene	19.511	202	36949	0.42	ng	100
32) Benzo(a)anthracene	21.250	228	31598	0.36	ng	98
33) Chrysene	21.303	228	37527	0.44	ng	97
34) Bis(2-ethylhexyl)phtha...	21.186	149	9293	0.22	ng	98
35) Indeno(1,2,3-cd)pyrene	25.724	276	44650	0.43	ng	98
37) Benzo(b)fluoranthene	22.825	252	39901	0.44	ng	96
38) Benzo(k)fluoranthene	22.866	252	39032m	0.45	ng	
39) Benzo(a)pyrene	23.393	252	36069	0.45	ng	# 90
40) Dibenzo(a,h)anthracene	25.738	278	36777	0.43	ng	97
41) Benzo(g,h,i)perylene	26.402	276	37496	0.43	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030821\
Data File : BN013868.D
Acq On : 08 Mar 2021 13:51
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM

Quant Time: Mar 08 15:15:26 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 08 10:32:09 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.725 min Scan# 614
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51





Tgt Ion: 152 Resp: 6826
Ion Ratio Lower Upper
152 100
150 151.3 121.8 182.8
115 56.1 46.6 69.8



#2
1,4-Dioxane
Concen: 0.43 ng
RT: 3.255 min Scan# 59
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51




Tgt Ion: 88 Resp: 3550
Ion Ratio Lower Upper
88 100
58 78.0 64.6 96.8
43 33.0 26.6 40.0



Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:50 PM







#3
n-Nitrosodimethylamine
Concen: 0.38 ng
RT: 3.577 min Scan# 99
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion: 42 Resp: 2408
Ion Ratio Lower Upper
42 100
74 111.3 95.0 142.6
44 6.1 5.8 8.8

Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4




Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:50 PM



#4
2-Fluorophenol
Concen: 0.35 ng
RT: 5.317 min Scan# 315
Delta R.T. 0.008 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion: 112 Resp: 6426
Ion Ratio Lower Upper
112 100
64 60.5 44.3 66.5
63 31.4 23.0 34.6









#5
Phenol-d6
Concen: 0.35 ng
RT: 6.887 min Scan# 510
Delta R.T. 0.008 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion: 99	Resp: 8472
Ion Ratio	Lower Upper
99 100	
42 22.7	15.2 22.8
71 33.0	26.1 39.1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.45 ng
RT: 7.145 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51


Tgt Ion: 93	Resp: 7705
Ion Ratio	Lower Upper
93 100	
63 80.4	60.8 91.2
95 33.2	26.0 39.0





#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.012 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:	136	Resp:	25748
Ion Ratio	Lower	Upper	
136	100		
137	11.2	9.0	13.6
54	10.8	6.4	9.6#
68	7.8	4.5	6.7#

Manual Integrations
APPROVED



mohammad
3/9/2021 12:19:50 PM




#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51



Tgt Ion:	82	Resp:	6638
Ion Ratio	Lower	Upper	
82	100		
128	32.5	39.2	58.8#
54	61.0	40.1	60.1#









#9
Naphthalene
Concen: 0.42 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:	128	Resp:	26628
Ion Ratio	Lower	Upper	
128	100		
129	11.2	9.0	13.4
127	13.2	10.4	15.6

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.835 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:	225	Resp:	5035
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	63.4	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.089 min Scan# 1070
Delta R.T. -0.004 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:152 Resp: 16437
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM




#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.164 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:142 Resp: 17754
Ion Ratio Lower Upper
142 100
141 87.2 70.5 105.7
115 34.9 27.4 41.0









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51





Tgt Ion:164 Resp: 13917
Ion Ratio Lower Upper
164 100
162 102.8 79.6 119.4
160 47.9 36.0 54.0



#14
2,4,6-Tribromophenol
Concen: 0.30 ng
RT: 15.843 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51




Tgt Ion:330 Resp: 1718
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.6 26.8 40.2



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:50 PM







#15
2-Fluorobiphenyl
Concen: 0.44 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:172 Resp: 21781
Ion Ratio Lower Upper
172 100
171 33.4 28.6 42.8
170 22.1 18.9 28.3

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#16
Acenaphthylene
Concen: 0.35 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51


Tgt Ion:152 Resp: 21589
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.0 10.5 15.7





#17
Acenaphthene
Concen: 0.41 ng
RT: 14.410 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:154 Resp: 15947
Ion Ratio Lower Upper
154 100
153 111.5 88.4 132.6
152 55.4 43.0 64.4

Manual Integrations
APPROVED



mohammad
3/9/2021 12:19:50 PM




#18
Fluorene
Concen: 0.39 ng
RT: 15.399 min Scan# 1392
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51



Tgt Ion:166 Resp: 19669
Ion Ratio Lower Upper
166 100
165 97.9 77.8 116.8
167 13.4 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Instrument :
BNA_N
ClientSampled :
SSTDCCC0.4

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	7.9	9.0	13.6#




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.56 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion	Ratio	Lower	Upper
198	100		
51	93.7	61.7	92.5#
105	34.9	33.0	49.6









#21
4-Bromophenyl-phenylether
Concen: 0.39 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.1	75.5	113.3
141	58.0	53.8	80.8




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#22
Hexachlorobenzene
Concen: 0.45 ng
RT: 16.398 min Scan# 1472
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51


Tgt Ion	Ratio	Lower	Upper
284	100		
142	34.1	27.9	41.9
249	29.7	24.2	36.2





#23
Atrazine
Concen: 0.25 ng
RT: 16.556 min Scan# 1485
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:200 Resp: 3749
Ion Ratio Lower Upper
200 100
173 26.0 18.6 28.0
215 49.3 41.0 61.6

Manual Integrations
APPROVED



mohammad
3/9/2021 12:19:50 PM




#24
Pentachlorophenol
Concen: 0.35 ng
RT: 16.739 min Scan# 1500
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51



Tgt Ion:266 Resp: 1929
Ion Ratio Lower Upper
266 100
264 62.6 51.6 77.4
268 64.9 54.6 82.0









#25
Phenanthrene
Concen: 0.42 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:178 Resp: 33242
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.6 12.3 18.5




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#26
Anthracene
Concen: 0.35 ng
RT: 17.226 min Scan# 1540
Delta R.T. 0.001 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:178 Resp: 26545
Ion Ratio Lower Upper
178 100
176 18.5 14.6 21.8
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.119 min Scan# 1827
Delta R.T. -0.002 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:212 Resp: 33009
Ion Ratio Lower Upper
212 100
106 12.7 9.7 14.5
104 7.0 5.5 8.3

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.149 min Scan# 1833
Delta R.T. -0.002 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:202 Resp: 36308
Ion Ratio Lower Upper
202 100
101 10.9 8.1 12.1
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.261 min Scan# 2124
Delta R.T. -0.008 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:240 Resp: 28636
Ion Ratio Lower Upper
240 100
120 15.0 10.2 15.4
236 27.4 21.3 31.9

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#30
Pyrene
Concen: 0.42 ng
RT: 19.511 min Scan# 1906
Delta R.T. -0.002 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:202 Resp: 36949
Ion Ratio Lower Upper
202 100
200 20.3 16.3 24.5
203 17.3 14.0 21.0









#31
Terphenyl-d14
Concen: 0.41 ng
RT: 19.720 min Scan# 1948
Delta R.T. -0.002 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:	244	Resp:	26153
Ion Ratio	Lower	Upper	
244	100		
212	7.7	6.1	9.1
122	12.1	9.8	14.8

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM



#32
Benzo(a)anthracene
Concen: 0.36 ng
RT: 21.250 min Scan# 2123
Delta R.T. 0.003 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion:	228	Resp:	31598
Ion Ratio	Lower	Upper	
228	100		
226	26.7	20.5	30.7
229	19.6	16.2	24.4









#33
Chrysene
Concen: 0.44 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.003 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion: 228	Resp: 37527
Ion Ratio	Lower Upper
228 100	
226 28.7	24.3 36.5
229 19.7	15.1 22.7

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.22 ng
RT: 21.186 min Scan# 2117
Delta R.T. -0.008 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion: 149	Resp: 9293
Ion Ratio	Lower Upper
149 100	
167 26.6	22.1 33.1
279 4.0	3.5 5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.43 ng
RT: 25.724 min Scan# 3201
Delta R.T. -0.004 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51





Tgt Ion: 276 Resp: 44650
Ion Ratio Lower Upper
276 100
138 23.2 19.8 29.6
277 24.8 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.493 min Scan# 2549
Delta R.T. -0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51




Tgt Ion: 264 Resp: 28157
Ion Ratio Lower Upper
264 100
260 24.6 19.3 28.9
265 31.6 66.2 99.4#




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED




mohammad
3/9/2021 12:19:50 PM



#37
Benzo(b)fluoranthene
Concen: 0.44 ng
RT: 22.825 min Scan# 2354
Delta R.T. -0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51






Tgt Ion:252 Resp: 39901
Ion Ratio Lower Upper
252 100
253 24.0 20.4 30.6
125 16.6 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.45 ng m
RT: 22.866 min Scan# 2366
Delta R.T. -0.000 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51


Tgt Ion:252 Resp: 39032
Ion Ratio Lower Upper
252 100
253 23.0 20.5 30.7
125 10.9 12.3 18.5#



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:50 PM







#39
Benzo(a)pyrene
Concen: 0.45 ng
RT: 23.393 min Scan# 2520
Delta R.T. -0.004 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion	Ratio	Lower	Upper
252	100		
253	25.4	21.8	32.6
125	27.2	14.6	21.8#

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:50 PM




#40
Dibenzo(a,h)anthracene
Concen: 0.43 ng
RT: 25.738 min Scan# 3205
Delta R.T. -0.004 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.7	13.1	19.7
279	24.8	21.4	32.0







#41
Benzo(g,h,i)perylene
Concen: 0.43 ng
RT: 26.402 min Scan# 3399
Delta R.T. -0.007 min
Lab File: BN013868.D
Acq: 08 Mar 2021 13:51



Tgt Ion:	276	Resp:	37496
Ion Ratio	Lower	Upper	
276	100		
277	24.2	20.0	30.0
138	21.1	17.8	26.8



Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:50 PM