

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030821\
 Data File : BN013870.D
 Acq On : 08 Mar 2021 15:24
 Operator : CG/JU
 Sample : PB135037BS
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB135037BS

Manual Integrations
 APPROVED

mohammad
 3/9/2021 12:19:52 PM

Quant Time: Mar 08 15:58:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 08 10:32:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.716	152	6591	0.40	ng	0.00
7) Naphthalene-d8	10.492	136	24888	0.40	ng	#-0.01
13) Acenaphthene-d10	14.346	164	13804	0.40	ng	0.00
19) Phenanthrene-d10	17.092	188	29989	0.40	ng	# 0.00
29) Chrysene-d12	21.271	240	29148	0.40	ng	# 0.00
36) Perylene-d12	23.493	264	28336	0.40	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.308	112	6077	0.34	ng	0.00
5) Phenol-d6	6.879	99	7981	0.34	ng	0.00
8) Nitrobenzene-d5	8.857	82	6419	0.40	ng	0.00
11) 2-Methylnaphthalene-d10	12.089	152	15979	0.40	ng	0.00
14) 2,4,6-Tribromophenol	15.843	330	1642	0.29	ng	0.00
15) 2-Fluorobiphenyl	12.976	172	21573	0.44	ng	0.00
27) Fluoranthene-d10	19.124	212	32755	0.37	ng	0.00
31) Terphenyl-d14	19.724	244	25926	0.40	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.255	88	3466	0.43	ng	97
3) n-Nitrosodimethylamine	3.577	42	2423	0.39	ng	# 93
6) bis(2-Chloroethyl)ether	7.145	93	7427	0.45	ng	95
9) Naphthalene	10.543	128	25623	0.42	ng	99
10) Hexachlorobutadiene	10.834	225	4787	0.41	ng	# 100
12) 2-Methylnaphthalene	12.164	142	17241	0.40	ng	99
16) Acenaphthylene	14.067	152	21096	0.34	ng	100
17) Acenaphthene	14.410	154	15490	0.40	ng	98
18) Fluorene	15.399	166	19284	0.39	ng	99
20) 4,6-Dinitro-2-methylph...	15.475	198	1162	0.56	ng	89
21) 4-Bromophenyl-phenylether	16.289	248	6271	0.38	ng	99
22) Hexachlorobenzene	16.398	284	7472	0.44	ng	99
23) Atrazine	16.556	200	3690	0.24	ng	93
24) Pentachlorophenol	16.739	266	1758	0.32	ng	100
25) Phenanthrene	17.128	178	33113	0.42	ng	100
26) Anthracene	17.226	178	26387	0.35	ng	99
28) Fluoranthene	19.154	202	36329	0.39	ng	99
30) Pyrene	19.516	202	36758	0.41	ng	100
32) Benzo(a)anthracene	21.250	228	31978	0.36	ng	98
33) Chrysene	21.303	228	36426	0.42	ng	98
34) Bis(2-ethylhexyl)phtha...	21.186	149	8986	0.21	ng	98
35) Indeno(1,2,3-cd)pyrene	25.724	276	44547	0.42	ng	98
37) Benzo(b)fluoranthene	22.825	252	39263	0.43	ng	97
38) Benzo(k)fluoranthene	22.870	252	40543m	0.46	ng	
39) Benzo(a)pyrene	23.393	252	35318	0.43	ng	# 89
40) Dibenzo(a,h)anthracene	25.741	278	36627	0.42	ng	97
41) Benzo(g,h,i)perylene	26.406	276	37243	0.43	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN030821\
Data File : BN013870.D
Acq On : 08 Mar 2021 15:24
Operator : CG/JU
Sample : PB135037BS
Misc :
ALS Vial : 4 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
PB135037BS

Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM

Quant Time: Mar 08 15:58:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN022421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 08 10:32:09 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.40 ng
RT: 7.716 min Scan# 613
Delta R.T. -0.008 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24





Tgt Ion: 152 Resp: 6591
Ion Ratio Lower Upper
152 100
150 157.4 121.8 182.8
115 61.2 46.6 69.8



#2
1,4-Dioxane
Concen: 0.43 ng
RT: 3.255 min Scan# 59
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Tgt Ion: 88 Resp: 3466
Ion Ratio Lower Upper
88 100
58 77.3 64.6 96.8
43 33.2 26.6 40.0



Instrument :
BNA_N
ClientSampleId :
PB135037BS

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:52 PM







#3
n-Nitrosodimethylamine
Concen: 0.39 ng
RT: 3.577 min Scan# 99
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion: 42 Resp: 2423
Ion Ratio Lower Upper
42 100
74 110.9 95.0 142.6
44 9.1 5.8 8.8#

Instrument :
BNA_N
ClientSampled :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#4
2-Fluorophenol
Concen: 0.34 ng
RT: 5.308 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion: 112 Resp: 6077
Ion Ratio Lower Upper
112 100
64 60.4 44.3 66.5
63 32.4 23.0 34.6









#5
Phenol-d6
Concen: 0.34 ng
RT: 6.879 min Scan# 509
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	99	Resp:	7981
Ion	Ratio	Lower	Upper
99	100		
42	22.8	15.2	22.8#
71	33.4	26.1	39.1

Instrument :
BNA_N
ClientSampled :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM




#6
bis(2-Chloroethyl)ether
Concen: 0.45 ng
RT: 7.145 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	93	Resp:	7427
Ion	Ratio	Lower	Upper
93	100		
63	81.4	60.8	91.2
95	32.9	26.0	39.0









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.492 min Scan# 859
Delta R.T. -0.013 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24





Tgt Ion: 136 Resp: 24888
Ion Ratio Lower Upper
136 100
137 11.1 9.0 13.6
54 11.1 6.4 9.6#
68 8.0 4.5 6.7#



#8
Nitrobenzene-d5
Concen: 0.40 ng
RT: 8.857 min Scan# 730
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Tgt Ion: 82 Resp: 6419
Ion Ratio Lower Upper
82 100
128 32.6 39.2 58.8#
54 59.1 40.1 60.1



Instrument :
BNA_N
ClientSampleId :
PB135037BS

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:52 PM







#9
Naphthalene
Concen: 0.42 ng
RT: 10.543 min Scan# 863
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	128	Resp:	25623
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.0	13.4
127	13.4	10.4	15.6

Instrument :
BNA_N
ClientSampleId :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#10
Hexachlorobutadiene
Concen: 0.41 ng
RT: 10.834 min Scan# 886
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	225	Resp:	4787
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.2	51.2	76.8









#11
2-Methylnaphthalene-d10
Concen: 0.40 ng
RT: 12.089 min Scan# 1070
Delta R.T. -0.004 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:152 Resp: 15979
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2

Instrument :
BNA_N
ClientSampleId :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#12
2-Methylnaphthalene
Concen: 0.40 ng
RT: 12.164 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24


Tgt Ion:142 Resp: 17241
Ion Ratio Lower Upper
142 100
141 88.4 70.5 105.7
115 35.5 27.4 41.0





#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.346 min Scan# 1309
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Instrument :
BNA_N
ClientSampleId :
PB135037BS




Tgt Ion:164 Resp: 13804
Ion Ratio Lower Upper
164 100
162 102.9 79.6 119.4
160 47.2 36.0 54.0

Manual Integrations
APPROVED



mohammad
3/9/2021 12:19:52 PM




#14
2,4,6-Tribromophenol
Concen: 0.29 ng
RT: 15.843 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24



Tgt Ion:330 Resp: 1642
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.7 26.8 40.2









#15
2-Fluorobiphenyl
Concen: 0.44 ng
RT: 12.976 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	172	Resp:	21573
Ion Ratio	Lower	Upper	
172	100		
171	33.2	28.6	42.8
170	21.9	18.9	28.3

Instrument :
BNA_N
ClientSampleId :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#16
Acenaphthylene
Concen: 0.34 ng
RT: 14.067 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24


Tgt Ion:	152	Resp:	21096
Ion Ratio	Lower	Upper	
152	100		
151	19.3	15.7	23.5
153	13.1	10.5	15.7





#17
Acenaphthene
Concen: 0.40 ng
RT: 14.410 min Scan# 1314
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Instrument :
BNA_N
ClientSampleId :
PB135037BS



Tgt Ion	Ratio	Lower	Upper
154	100		
153	112.5	88.4	132.6
152	55.7	43.0	64.4




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#18
Fluorene
Concen: 0.39 ng
RT: 15.399 min Scan# 1392
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.2	77.8	116.8
167	13.6	10.6	16.0









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.092 min Scan# 1529
Delta R.T. 0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Instrument :
BNA_N
ClientSampleId :
PB135037BS

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	8.8	9.0	13.6#




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#20
4,6-Dinitro-2-methylphenol
Concen: 0.56 ng
RT: 15.475 min Scan# 1398
Delta R.T. 0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion	Ratio	Lower	Upper
198	100		
51	89.9	61.7	92.5
105	38.5	33.0	49.6









#21
4-Bromophenyl-phenylether
Concen: 0.38 ng
RT: 16.289 min Scan# 1463
Delta R.T. 0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Instrument :
BNA_N
ClientSampleId :
PB135037BS

Tgt Ion:	248	Resp:	6271
Ion Ratio	Lower	Upper	
248	100		
250	92.6	75.5	113.3
141	66.6	53.8	80.8




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#22
Hexachlorobenzene
Concen: 0.44 ng
RT: 16.398 min Scan# 1472
Delta R.T. 0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24


Tgt Ion:	284	Resp:	7472
Ion Ratio	Lower	Upper	
284	100		
142	34.7	27.9	41.9
249	30.8	24.2	36.2





#23
Atrazine
Concen: 0.24 ng
RT: 16.556 min Scan# 1485
Delta R.T. 0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Instrument :
BNA_N
ClientSampleId :
PB135037BS




Tgt Ion:200 Resp: 3690
Ion Ratio Lower Upper
200 100
173 27.7 18.6 28.0
215 46.6 41.0 61.6

Manual Integrations
APPROVED



mohammad
3/9/2021 12:19:52 PM




#24
Pentachlorophenol
Concen: 0.32 ng
RT: 16.739 min Scan# 1500
Delta R.T. 0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Tgt Ion:266 Resp: 1758
Ion Ratio Lower Upper
266 100
264 64.4 51.6 77.4
268 68.5 54.6 82.0





#25
Phenanthrene
Concen: 0.42 ng
RT: 17.128 min Scan# 1532
Delta R.T. 0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Instrument :
BNA_N
ClientSampled :
PB135037BS




Tgt Ion:178 Resp: 33113
Ion Ratio Lower Upper
178 100
176 19.0 15.0 22.4
179 15.3 12.3 18.5

Manual Integrations
APPROVED



mohammad
3/9/2021 12:19:52 PM




#26
Anthracene
Concen: 0.35 ng
RT: 17.226 min Scan# 1540
Delta R.T. 0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Tgt Ion:178 Resp: 26387
Ion Ratio Lower Upper
178 100
176 18.5 14.6 21.8
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.37 ng
RT: 19.124 min Scan# 1828
Delta R.T. 0.003 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24





Tgt Ion:212 Resp: 32755
Ion Ratio Lower Upper
212 100
106 12.7 9.7 14.5
104 7.2 5.5 8.3



#28
Fluoranthene
Concen: 0.39 ng
RT: 19.154 min Scan# 1834
Delta R.T. 0.003 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Tgt Ion:202 Resp: 36329
Ion Ratio Lower Upper
202 100
101 11.0 8.1 12.1
203 17.2 13.9 20.9



Instrument :
BNA_N
ClientSampleId :
PB135037BS

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:52 PM







#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.271 min Scan# 2125
Delta R.T. 0.003 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:240 Resp: 29148
Ion Ratio Lower Upper
240 100
120 5.9 10.2 15.4#
236 25.4 21.3 31.9

Instrument :
BNA_N
ClientSampleId :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM



#30
Pyrene
Concen: 0.41 ng
RT: 19.516 min Scan# 1907
Delta R.T. 0.003 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24


Tgt Ion:202 Resp: 36758
Ion Ratio Lower Upper
202 100
200 20.4 16.3 24.5
203 17.4 14.0 21.0





#31
 Terphenyl-d14
 Concen: 0.40 ng
 RT: 19.724 min Scan# 1949
 Delta R.T. 0.003 min
 Lab File: BN013870.D
 Acq: 08 Mar 2021 15:24




Instrument :
 BNA_N
 ClientSampled :
 PB135037BS




Tgt Ion:244 Resp: 25926
 Ion Ratio Lower Upper
 244 100
 212 7.5 6.1 9.1
 122 10.4 9.8 14.8

Manual Integrations
 APPROVED



mohammad
 3/9/2021 12:19:52 PM




#32
 Benzo(a)anthracene
 Concen: 0.36 ng
 RT: 21.250 min Scan# 2123
 Delta R.T. 0.003 min
 Lab File: BN013870.D
 Acq: 08 Mar 2021 15:24



Tgt Ion:228 Resp: 31978
 Ion Ratio Lower Upper
 228 100
 226 26.9 20.5 30.7
 229 19.4 16.2 24.4









#33
Chrysene
Concen: 0.42 ng
RT: 21.303 min Scan# 2128
Delta R.T. 0.003 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	228	Resp:	36426
Ion Ratio	Lower	Upper	
228	100		
226	29.1	24.3	36.5
229	19.9	15.1	22.7

Instrument :
BNA_N
ClientSampleId :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM




#34
Bis(2-ethylhexyl)phthalate
Concen: 0.21 ng
RT: 21.186 min Scan# 2117
Delta R.T. -0.008 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	149	Resp:	8986
Ion Ratio	Lower	Upper	
149	100		
167	26.5	22.1	33.1
279	4.0	3.5	5.3









#35
Indeno(1,2,3-cd)pyrene
Concen: 0.42 ng
RT: 25.724 min Scan# 3201
Delta R.T. -0.004 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24





Tgt Ion: 276 Resp: 44547
Ion Ratio Lower Upper
276 100
138 23.2 19.8 29.6
277 24.8 19.9 29.9



#36
Perylene-d12
Concen: 0.40 ng
RT: 23.493 min Scan# 2549
Delta R.T. -0.000 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24




Tgt Ion: 264 Resp: 28336
Ion Ratio Lower Upper
264 100
260 24.5 19.3 28.9
265 31.4 66.2 99.4#




Instrument :
BNA_N
ClientSampleId :
PB135037BS

Manual Integrations
APPROVED




mohammad
3/9/2021 12:19:52 PM



#37
Benzo(b)fluoranthene
Concen: 0.43 ng
RT: 22.825 min Scan# 2354
Delta R.T. -0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24






Tgt Ion:252 Resp: 39263
Ion Ratio Lower Upper
252 100
253 24.1 20.4 30.6
125 15.9 11.6 17.4



#38
Benzo(k)fluoranthene
Concen: 0.46 ng m
RT: 22.870 min Scan# 2367
Delta R.T. 0.003 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24


Tgt Ion:252 Resp: 40543
Ion Ratio Lower Upper
252 100
253 23.4 20.5 30.7
125 11.0 12.3 18.5#



Instrument :
BNA_N
ClientSampleId :
PB135037BS

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:52 PM







#39
Benzo(a)pyrene
Concen: 0.43 ng
RT: 23.393 min Scan# 2520
Delta R.T. -0.004 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	252	Resp:	35318
Ion Ratio	Lower	Upper	
252	100		
253	25.5	21.8	32.6
125	28.2	14.6	21.8#

Instrument :
BNA_N
ClientSampleId :
PB135037BS




Manual Integrations
APPROVED


mohammad
3/9/2021 12:19:52 PM




#40
Dibenzo(a,h)anthracene
Concen: 0.42 ng
RT: 25.741 min Scan# 3206
Delta R.T. -0.001 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24

Tgt Ion:	278	Resp:	36627
Ion Ratio	Lower	Upper	
278	100		
139	16.1	13.1	19.7
279	24.7	21.4	32.0







#41
Benzo(g,h,i)perylene
Concen: 0.43 ng
RT: 26.406 min Scan# 3400
Delta R.T. -0.004 min
Lab File: BN013870.D
Acq: 08 Mar 2021 15:24



Tgt Ion:	276	Resp:	37243
Ion Ratio	Lower	Upper	
276	100		
277	24.2	20.0	30.0
138	21.4	17.8	26.8



Instrument :
BNA_N
ClientSampleId :
PB135037BS

Manual Integrations
APPROVED

mohammad
3/9/2021 12:19:52 PM