

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
 Data File : BN010169.D
 Acq On : 10 Mar 2020 20:40
 Operator : CG/JU
 Sample : L1842-11
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D6-20200304

Manual Integrations
 APPROVED

mohammad
 3/11/2020 5:07:59 PM

Quant Time: Mar 11 11:57:45 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	750	0.40	ng	-0.04
7) Naphthalene-d8	10.15	136	2362	0.40	ng	-0.03
13) Acenaphthene-d10	14.00	164	1559	0.40	ng	-0.06
19) Phenanthrene-d10	16.79	188	3128	0.40	ng	-0.02
27) Chrysene-d12	20.99	240	3521	0.40	ng	-0.04
34) Perylene-d12	23.10	264	3567	0.40	ng	-0.05
System Monitoring Compounds						
4) 2-Fluorophenol	5.05	112	240	0.14	ng	-0.03
5) Phenol-d6	6.64	99	143	0.07	ng	0.00
8) Nitrobenzene-d5	8.61	82	579	0.29	ng	0.01
11) 2-Methylnaphthalene-d10	11.76	152	1236	0.27	ng	-0.03
14) 2,4,6-Tribromophenol	15.56	330	219	0.36	ng	-0.02
15) 2-Fluorobiphenyl	12.64	172	1870	0.32	ng	-0.04
25) Fluoranthene-d10	18.83	212	3199	0.30	ng	-0.03
29) Terphenyl-d14	19.43	244	2841	0.34	ng	-0.05
Target Compounds						
2) 1,4-Dioxane	3.03	88	4261m	5.547	ng	Qvalue
18) Fluorene	15.10	166	182	0.030	ng	94
23) Phenanthrene	16.84	178	644	0.069	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
Data File : BN010169.D
Acq On : 10 Mar 2020 20:40
Operator : CG/JU
Sample : L1842-11
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D6-20200304

Manual Integrations
APPROVED

mohammad
3/11/2020 5:07:59 PM

Quant Time: Mar 11 11:57:45 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

