

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
 Data File : BN010188.D
 Acq On : 11 Mar 2020 12:21
 Operator : CG/JU
 Sample : L1842-27
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE139D1-20200306

Manual Integrations
 APPROVED

mohammad
 3/11/2020 5:08:46 PM

Quant Time: Mar 11 15:41:18 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	691	0.40	ng	-0.04
7) Naphthalene-d8	10.16	136	3244m	0.40	ng	-0.01
13) Acenaphthene-d10	14.01	164	1401	0.40	ng	-0.04
19) Phenanthrene-d10	16.80	188	3103	0.40	ng	-0.01
27) Chrysene-d12	21.00	240	3053	0.40	ng	-0.03
34) Perylene-d12	23.10	264	3076	0.40	ng	-0.04
System Monitoring Compounds						
4) 2-Fluorophenol	5.07	112	303	0.20	ng	-0.02
5) Phenol-d6	6.60	99	311	0.17	ng	-0.05
8) Nitrobenzene-d5	8.63	82	760m	0.28	ng	0.04
11) 2-Methylnaphthalene-d10	11.76	152	1394	0.22	ng	-0.02
14) 2,4,6-Tribromophenol	15.57	330	211	0.39	ng	-0.01
15) 2-Fluorobiphenyl	12.64	172	1971	0.37	ng	-0.04
25) Fluoranthene-d10	18.83	212	3394	0.32	ng	-0.03
29) Terphenyl-d14	19.43	244	2911	0.40	ng	-0.05
Target Compounds						
2) 1,4-Dioxane	3.04	88	1077m	1.522	ng	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
Data File : BN010188.D
Acq On : 11 Mar 2020 12:21
Operator : CG/JU
Sample : L1842-27
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE139D1-20200306

Manual Integrations
APPROVED

mohammad
3/11/2020 5:08:46 PM

Quant Time: Mar 11 15:41:18 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

