

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
 Data File : BN010159.D
 Acq On : 10 Mar 2020 13:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 10 18:50:40 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.410	0.462	-12.7	110	0.00
3	n-Nitrosodimethylamine	0.673	0.896	-33.1#	147	0.00
4 S	2-Fluorophenol	0.889	1.019	-14.6	115	0.00
5 S	Phenol-d6	1.082	1.164	-7.6	110	0.00
6	bis(2-Chloroethyl)ether	1.052	1.111	-5.6	106	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
8 S	Nitrobenzene-d5	0.333	0.341	-2.4	109	0.00
9	Naphthalene	1.091	1.059	2.9	100	0.00
10	Hexachlorobutadiene	0.240	0.242	-0.8	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.774	0.738	4.7	99	0.00
12	2-Methylnaphthalene	0.748	0.701	6.3	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	0.00
14 S	2,4,6-Tribromophenol	0.154	0.163	-5.8	112	0.00
15 S	2-Fluorobiphenyl	1.510	1.470	2.6	96	0.00
16	Acenaphthylene	1.705	1.640	3.8	101	0.00
17	Acenaphthene	1.199	1.114	7.1	92	0.00
18	Fluorene	1.582	1.425	9.9	88	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4-Bromophenyl-phenylether	0.279	0.107	61.6#	34#	0.00
21	Hexachlorobenzene	0.249	0.243	2.4	87	0.00
22	Pentachlorophenol	0.064	0.066	-3.1	118	0.00
23	Phenanthrene	1.190	1.075	9.7	83	0.00
24	Anthracene	1.006	0.950	5.6	90	0.00
25 SURR	Fluoranthene-d10	1.384	1.283	7.3	85	0.00
26	Fluoranthene	1.425	1.302	8.6	84	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
28	Pvrene	1.519	1.363	10.3	86	0.00
29 S	Terphenyl-d14	0.956	0.849	11.2	85	0.00
30	Benzo(a)anthracene	1.327	1.206	9.1	95	0.00
31	Chrysene	1.536	1.395	9.2	87	0.00
32	Bis(2-ethylhexyl)phthalate	0.648	0.791	-22.1	126	0.00
33	Indeno(1,2,3-cd)pyrene	1.599	1.709	-6.9	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.462	1.335	8.7	96	0.00
36	Benzo(k)fluoranthene	1.592	1.403	11.9	90	0.00
37 C	Benzo(a)pyrene	1.334	1.294	3.0	96	0.00
38	Dibenzo(a,h)anthracene	1.314	1.243	5.4	97	0.00
39	Benzo(a,h,i)perylene	1.400	1.369	2.2	100	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
Data File : BN010159.D
Acq On : 10 Mar 2020 13:32
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 10 18:50:40 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		