

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
 Data File : BN010172.D
 Acq On : 10 Mar 2020 22:28
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 11 04:09:51 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.410	0.551	-34.4#	100	0.00
3	n-Nitrosodimethylamine	0.673	0.970	-44.1#	121	0.00
4 S	2-Fluorophenol	0.889	1.078	-21.3	93	0.00
5 S	Phenol-d6	1.082	1.166	-7.8	84	0.02
6	bis(2-Chloroethyl)ether	1.052	1.116	-6.1	82	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	78	0.01
8 S	Nitrobenzene-d5	0.333	0.332	0.3	83	0.03
9	Naphthalene	1.091	1.130	-3.6	83	0.01
10	Hexachlorobutadiene	0.240	0.266	-10.8	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.774	0.796	-2.8	83	0.01
12	2-Methylnaphthalene	0.748	0.764	-2.1	83	0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
14 S	2,4,6-Tribromophenol	0.154	0.166	-7.8	93	0.01
15 S	2-Fluorobiphenyl	1.510	1.493	1.1	79	0.01
16	Acenaphthylene	1.705	1.694	0.6	85	0.00
17	Acenaphthene	1.199	1.210	-0.9	81	0.00
18	Fluorene	1.582	1.515	4.2	76	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	70	0.01
20	4-Bromophenyl-phenylether	0.279	0.175	37.3#	46#	0.00
21	Hexachlorobenzene	0.249	0.263	-5.6	78	0.01
22	Pentachlorophenol	0.064	0.062	3.1	91	0.01
23	Phenanthrene	1.190	1.144	3.9	73	0.01
24	Anthracene	1.006	0.976	3.0	76	0.01
25 SURR	Fluoranthene-d10	1.384	1.397	-0.9	76	0.00
26	Fluoranthene	1.425	1.410	1.1	75	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
28	Pvrene	1.519	1.431	5.8	76	0.00
29 S	Terphenyl-d14	0.956	0.902	5.6	76	0.00
30	Benzo(a)anthracene	1.327	1.196	9.9	79	0.00
31	Chrysene	1.536	1.563	-1.8	82	0.01
32	Bis(2-ethylhexyl)phthalate	0.648	0.809	-24.8	109	0.00
33	Indeno(1,2,3-cd)pyrene	1.599	1.639	-2.5	80	0.00
34 I	Perylene-d12	1.000	1.000	0.0	79	0.00
35	Benzo(b)fluoranthene	1.462	1.342	8.2	80	0.00
36	Benzo(k)fluoranthene	1.592	1.584	0.5	83	0.00
37 C	Benzo(a)pyrene	1.334	1.294	3.0	79	0.00
38	Dibenzo(a,h)anthracene	1.314	1.246	5.2	80	0.02
39	Benzo(a,h,i)perylene	1.400	1.396	0.3	84	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
Data File : BN010172.D
Acq On : 10 Mar 2020 22:28
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 11 04:09:51 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		