

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
 Data File : BN010174.D
 Acq On : 11 Mar 2020 03:55
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 11 04:31:09 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 19 15:57:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.410	0.422	-2.9	87	0.00
3	n-Nitrosodimethylamine	0.673	0.889	-32.1#	126	0.02
4 S	2-Fluorophenol	0.889	0.885	0.4	86	0.00
5 S	Phenol-d6	1.082	1.091	-0.8	89	0.02
6	bis(2-Chloroethyl)ether	1.052	1.310	-24.5	108	0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	76	0.01
8 S	Nitrobenzene-d5	0.333	0.315	5.4	76	0.03
9	Naphthalene	1.091	1.129	-3.5	81	0.01
10	Hexachlorobutadiene	0.240	0.269	-12.1	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.774	0.835	-7.9	85	0.01
12	2-Methylnaphthalene	0.748	0.833	-11.4	88	0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	74	0.00
14 S	2,4,6-Tribromophenol	0.154	0.149	3.2	80	0.01
15 S	2-Fluorobiphenyl	1.510	1.668	-10.5	85	0.01
16	Acenaphthylene	1.705	1.668	2.2	81	0.00
17	Acenaphthene	1.199	1.197	0.2	77	0.00
18	Fluorene	1.582	1.498	5.3	72	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.01
20	4-Bromophenyl-phenylether	0.279	0.200	28.3#	50#	0.00
21	Hexachlorobenzene	0.249	0.266	-6.8	73	0.01
22	Pentachlorophenol	0.064	0.057	10.9	79	0.01
23	Phenanthrene	1.190	1.150	3.4	68	0.01
24	Anthracene	1.006	0.972	3.4	70	0.01
25 SURR	Fluoranthene-d10	1.384	1.373	0.8	70	0.00
26	Fluoranthene	1.425	1.419	0.4	70	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	72	0.00
28	Pvrene	1.519	1.641	-8.0	80	0.00
29 S	Terphenyl-d14	0.956	0.969	-1.4	75	0.01
30	Benzo(a)anthracene	1.327	1.176	11.4	72	0.00
31	Chrysene	1.536	1.619	-5.4	78	0.00
32	Bis(2-ethylhexyl)phthalate	0.648	0.726	-12.0	90	-0.01
33	Indeno(1,2,3-cd)pyrene	1.599	1.806	-12.9	81	0.00
34 I	Perylene-d12	1.000	1.000	0.0	74	0.00
35	Benzo(b)fluoranthene	1.462	1.418	3.0	79	0.00
36	Benzo(k)fluoranthene	1.592	1.622	-1.9	80	0.00
37 C	Benzo(a)pyrene	1.334	1.326	0.6	76	0.00
38	Dibenzo(a,h)anthracene	1.314	1.334	-1.5	80	0.01
39	Benzo(a,h,i)perylene	1.400	1.491	-6.5	84	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN031020\
Data File : BN010174.D
Acq On : 11 Mar 2020 03:55
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 11 04:31:09 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN021920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 19 15:57:43 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			