

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031423\
 Data File : BN024285.D
 Acq On : 14 Mar 2023 17:55
 Operator : CG/JU
 Sample : PB151355BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB151355BL

Quant Time: Mar 15 01:56:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 15 01:49:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.047	152	7824	0.400	ng	0.00
7) Naphthalene-d8	10.866	136	22013	0.400	ng	0.00
13) Acenaphthene-d10	14.675	164	10447	0.400	ng	0.00
19) Phenanthrene-d10	17.425	188	21444	0.400	ng	0.00
29) Chrysene-d12	21.619	240	15579	0.400	ng	0.00
35) Perylene-d12	24.117	264	13306	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.584	112	6895	0.398	ng	0.00
5) Phenol-d6	7.195	99	8465	0.380	ng	0.00
8) Nitrobenzene-d5	9.211	82	6031	0.410	ng	0.00
11) 2-Methylnaphthalene-d10	12.444	152	10815	0.400	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	1117	0.209	ng	0.00
15) 2-Fluorobiphenyl	13.307	172	15711	0.398	ng	0.00
27) Fluoranthene-d10	19.450	212	19022	0.418	ng	0.00
31) Terphenyl-d14	20.043	244	11044	0.420	ng	0.00
Target Compounds						
						Qvalue
3) n-Nitrosodimethylamine	3.714	42	46	0.004	ng	# 1
6) bis(2-Chloroethyl)ether	7.469	93	49	0.002	ng	# 66
9) Naphthalene	10.909	128	111	0.002	ng	# 2
10) Hexachlorobutadiene	11.197	225	22	0.002	ng	# 65
12) 2-Methylnaphthalene	12.516	142	65	0.002	ng	# 49
16) Acenaphthylene	14.397	152	81	0.002	ng	# 86
17) Acenaphthene	14.740	154	49	0.002	ng	# 74
18) Fluorene	15.725	166	65	0.002	ng	# 95
20) 4,6-Dinitro-2-methylph...	15.787	198	10	0.181	ng	# 1
21) 4-Bromophenyl-phenylether	16.606	248	19	0.001	ng	# 48
22) Hexachlorobenzene	16.730	284	37	0.002	ng	# 29
23) Atrazine	16.879	200	15	0.002	ng	# 1
24) Pentachlorophenol	17.078	266	52	0.134	ng	90
25) Phenanthrene	17.562	178	147	0.002	ng	95
26) Anthracene	17.562	178	147	0.003	ng	98
28) Fluoranthene	19.480	202	168	0.002	ng	# 83
30) Pyrene	19.842	202	158	0.003	ng	# 87
32) Benzo(a)anthracene	21.601	228	200	0.004	ng	# 52
33) Chrysene	21.655	228	169	0.003	ng	# 50
34) Bis(2-ethylhexyl)phtha...	21.520	149	464	0.019	ng	# 99
36) Indeno(1,2,3-cd)pyrene	26.728	276	111	0.002	ng	# 51
37) Benzo(b)fluoranthene	23.354	252	181	0.003	ng	# 1
38) Benzo(k)fluoranthene	23.404	252	115	0.002	ng	# 1
39) Benzo(a)pyrene	24.003	252	75	0.002	ng	# 1
40) Dibenzo(a,h)anthracene	26.743	278	99	0.003	ng	# 1
41) Benzo(g,h,i)perylene	27.521	276	137	0.003	ng	# 1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031423\
 Data File : BN024285.D
 Acq On : 14 Mar 2023 17:55
 Operator : CG/JU
 Sample : PB151355BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB151355BL


Quant Time: Mar 15 01:56:26 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 15 01:49:05 2023
 Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 8.047 min Scan# 60
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Instrument :
BNA_N
ClientSampleId :
PB151355BL






Tgt Ion:152 Resp: 7824
Ion Ratio Lower Upper
152 100
150 152.5 124.7 187.1
115 59.0 49.1 73.7



#3
n-Nitrosodimethylamine
Concen: 0.004 ng
RT: 3.714 min Scan# 60
Delta R.T. -0.050 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion: 42 Resp: 46
Ion Ratio Lower Upper
42 100
74 0.0 81.6 122.4#
44 280.4 7.3 10.9#









#4
2-Fluorophenol
Concen: 0.398 ng
RT: 5.584 min Scan# 319
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:112 Resp: 6895
Ion Ratio Lower Upper
112 100
64 63.7 49.0 73.6
63 33.3 25.7 38.5



#5
Phenol-d6
Concen: 0.380 ng
RT: 7.195 min Scan# 542
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion: 99 Resp: 8465
Ion Ratio Lower Upper
99 100
42 24.9 20.1 30.1
71 33.4 25.8 38.8









#6
bis(2-Chloroethyl)ether
Concen: 0.002 ng
RT: 7.469 min Scan# 580
Delta R.T. 0.007 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion: 93 Resp: 49
Ion Ratio Lower Upper
93 100
63 53.1 65.4 98.2#
95 53.1 25.8 38.8#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.866 min Scan# 936
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion: 136 Resp: 22013
Ion Ratio Lower Upper
136 100
137 11.1 8.9 13.3
54 8.5 7.1 10.7
68 6.3 5.3 7.9









#8
Nitrobenzene-d5
Concen: 0.410 ng
RT: 9.211 min Scan# 781
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion: 82 Resp: 6031
Ion Ratio Lower Upper
82 100
128 40.6 32.4 48.6
54 55.5 45.2 67.8



#9
Naphthalene
Concen: 0.002 ng
RT: 10.909 min Scan# 940
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion: 128 Resp: 111
Ion Ratio Lower Upper
128 100
129 52.9 9.0 13.6#
127 47.8 10.2 15.4#









#10
Hexachlorobutadiene
Concen: 0.002 ng
RT: 11.197 min Scan# 90
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:225 Resp: 22
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 36.4 51.0 76.4#



#11
2-Methylnaphthalene-d10
Concen: 0.400 ng
RT: 12.444 min Scan# 1244
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:152 Resp: 10815
Ion Ratio Lower Upper
152 100
151 21.1 16.8 25.2









#12
2-Methylnaphthalene
Concen: 0.002 ng
RT: 12.516 min Scan# 1263
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:142 Resp: 65
Ion Ratio Lower Upper
142 100
141 93.8 71.8 107.6
115 134.4 30.0 45.0#



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.675 min Scan# 1549
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:164 Resp: 10447
Ion Ratio Lower Upper
164 100
162 106.0 84.7 127.1
160 56.5 44.6 66.8









#14
2,4,6-Tribromophenol
Concen: 0.209 ng
RT: 16.159 min Scan# 1682
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:330 Resp: 1117
Ion Ratio Lower Upper
330 100
332 95.7 78.0 117.0
141 44.1 30.8 46.2



#15
2-Fluorobiphenyl
Concen: 0.398 ng
RT: 13.307 min Scan# 1421
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:172 Resp: 15711
Ion Ratio Lower Upper
172 100
171 34.7 27.8 41.8
170 23.5 19.0 28.6





#16

Acenaphthylene

Concen: 0.002 ng

RT: 14.397 min Scan# 1523

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

PB151355BL





Tgt Ion:152 Resp: 81

Ion Ratio Lower Upper

152 100

151 25.9 15.4 23.0#

153 18.5 10.4 15.6#



#17

Acenaphthene

Concen: 0.002 ng

RT: 14.740 min Scan# 1555

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55




Tgt Ion:154 Resp: 49


Ion Ratio Lower Upper

154 100

153 126.5 88.4 132.6

152 95.9 48.7 73.1#





#18

Fluorene

Concen: 0.002 ng

RT: 15.725 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

PB151355BL





Tgt Ion:166 Resp: 65

Ion Ratio Lower Upper

166 100

165 96.9 80.2 120.2

167 21.5 10.8 16.2#



#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.425 min Scan# 1784

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55




Tgt Ion:188 Resp: 21444


Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.4 8.9 13.3









#20
4,6-Dinitro-2-methylphenol
Concen: 0.181 ng
RT: 15.787 min Scan# 1652
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:198 Resp: 10
Ion Ratio Lower Upper
198 100
51 534.7 68.2 102.2#
105 467.3 51.6 77.4#



#21
4-Bromophenyl-phenylether
Concen: 0.001 ng
RT: 16.606 min Scan# 1718
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:248 Resp: 19
Ion Ratio Lower Upper
248 100
250 117.9 81.9 122.9
141 128.6 28.2 42.2#









#22
Hexachlorobenzene
Concen: 0.002 ng
RT: 16.730 min Scan# 1728
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:284 Resp: 37
Ion Ratio Lower Upper
284 100
142 67.6 30.2 45.4#
249 83.8 24.5 36.7#



#23
Atrazine
Concen: 0.002 ng
RT: 16.879 min Scan# 1740
Delta R.T. 0.013 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:200 Resp: 15
Ion Ratio Lower Upper
200 100
173 96.2 18.9 28.3#
215 96.2 36.7 55.1#





#24

Pentachlorophenol

Concen: 0.134 ng

RT: 17.078 min Scan# 1795

Delta R.T. 0.013 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

PB151355BL





Tgt Ion:266 Resp: 52

Ion Ratio Lower Upper

266 100

264 69.2 51.1 76.7

268 75.0 52.2 78.4



#25

Phenanthrene

Concen: 0.002 ng

RT: 17.562 min Scan# 1795

Delta R.T. 0.100 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55




Tgt Ion:178 Resp: 147


Ion Ratio Lower Upper

178 100

176 17.0 15.4 23.0

179 17.0 12.2 18.2





#26

Anthracene

Concen: 0.003 ng

RT: 17.562 min Scan# 1795

Delta R.T. 0.013 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

PB151355BL





Tgt Ion:178 Resp: 147

Ion Ratio Lower Upper

178 100

176 17.7 15.1 22.7

179 15.0 12.1 18.1



#27

Fluoranthene-d10

Concen: 0.418 ng

RT: 19.450 min Scan# 2173

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55




Tgt Ion:212 Resp: 19022


Ion Ratio Lower Upper

212 100

106 20.2 15.4 23.2

104 11.6 8.6 13.0





#28

Fluoranthene

Concen: 0.002 ng

RT: 19.480 min Scan# 2180

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

PB151355BL





Tgt Ion:202 Resp: 168

Ion Ratio Lower Upper

202 100

101 0.0 11.3 16.9#

203 16.1 13.6 20.4



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.619 min Scan# 2485

Delta R.T. 0.009 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55




Tgt Ion:240 Resp: 15579


Ion Ratio Lower Upper

240 100

120 14.0 12.8 19.2

236 27.5 22.5 33.7





#30

Pyrene

Concen: 0.003 ng

RT: 19.842 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55

Instrument :

BNA_N

ClientSampleId :

PB151355BL





Tgt Ion:202 Resp: 158

Ion Ratio Lower Upper

202 100

200 26.6 17.0 25.4#

203 24.1 14.2 21.2#



#31

Terphenyl-d14

Concen: 0.420 ng

RT: 20.043 min Scan# 2309

Delta R.T. 0.000 min

Lab File: BN024285.D

Acq: 14 Mar 2023 17:55




Tgt Ion:244 Resp: 11044


Ion Ratio Lower Upper

244 100

212 9.0 7.6 11.4

122 15.2 15.1 22.7









#32
Benzo(a)anthracene
Concen: 0.004 ng
RT: 21.601 min Scan# 2483
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:228 Resp: 200
Ion Ratio Lower Upper
228 100
226 49.7 22.2 33.2#
229 45.2 15.9 23.9#



#33
Chrysene
Concen: 0.003 ng
RT: 21.655 min Scan# 2489
Delta R.T. 0.000 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:228 Resp: 169
Ion Ratio Lower Upper
228 100
226 55.2 24.5 36.7#
229 45.5 15.5 23.3#









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.019 ng
 RT: 21.520 min Scan# 2474
 Delta R.T. 0.009 min
 Lab File: BN024285.D
 Acq: 14 Mar 2023 17:55




Instrument :
 BNA_N
 ClientSampleId :
 PB151355BL


Tgt Ion:	149	Resp:	464
Ion Ratio	Lower	Upper	
149	100		
167	26.7	21.2	31.8
279	6.7	1.8	2.8



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 24.117 min Scan# 3158
 Delta R.T. -0.011 min
 Lab File: BN024285.D
 Acq: 14 Mar 2023 17:55

Tgt Ion:	264	Resp:	13306
Ion Ratio	Lower	Upper	
264	100		
260	27.7	21.4	32.2
265	57.2	35.7	53.5









#36
Indeno(1,2,3-cd)pyrene
Concen: 0.002 ng
RT: 26.728 min Scan# 4050
Delta R.T. -0.035 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55




Instrument :
BNA_N
ClientSampleId :
PB151355BL


Tgt Ion:276 Resp: 111
Ion Ratio Lower Upper
276 100
138 0.0 27.4 41.2#
277 7.2 19.8 29.6#



#37
Benzo(b)fluoranthene
Concen: 0.003 ng
RT: 23.354 min Scan# 2897
Delta R.T. -0.003 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55


Tgt Ion:252 Resp: 181
Ion Ratio Lower Upper
252 100
253 100.0 18.9 28.3#
125 163.8 13.8 20.6#








#38
Benzo(k)fluoranthene
Concen: 0.002 ng
RT: 23.404 min Scan# 2914
Delta R.T. -0.006 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Instrument :
BNA_N
ClientSampleId :
PB151355BL







Tgt Ion:252 Resp: 115
Ion Ratio Lower Upper
252 100
253 99.7 19.0 28.6#
125 160.7 13.5 20.3#



#39
Benzo(a)pyrene
Concen: 0.002 ng
RT: 24.003 min Scan# 3119
Delta R.T. -0.011 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:252 Resp: 75
Ion Ratio Lower Upper
252 100
253 104.9 19.5 29.3#
125 182.4 16.2 24.4#









#40
Dibenzo(a,h)anthracene
Concen: 0.003 ng
RT: 26.743 min Scan# 4056
Delta R.T. -0.044 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Instrument :
BNA_N
ClientSampleId :
PB151355BL

Tgt Ion:278 Resp: 99
Ion Ratio Lower Upper
278 100
139 114.0 19.4 29.0#
279 194.2 20.7 31.1#



#41
Benzo(g,h,i)perylene
Concen: 0.003 ng
RT: 27.521 min Scan# 4322
Delta R.T. -0.055 min
Lab File: BN024285.D
Acq: 14 Mar 2023 17:55

Tgt Ion:276 Resp: 137
Ion Ratio Lower Upper
276 100
277 93.2 19.9 29.9#
138 111.4 24.5 36.7#

