

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031423\
 Data File : BN024275.D
 Acq On : 14 Mar 2023 11:17
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 15 01:53:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 15 01:49:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	95	0.00
2	1,4-Dioxane	0.400	0.409	-2.2	93	0.00
3	n-Nitrosodimethylamine	0.400	0.418	-4.5	97	0.00
4 S	2-Fluorophenol	0.400	0.394	1.5	93	0.00
5 S	Phenol-d6	0.400	0.382	4.5	93	0.00
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	92	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.397	0.8	100	0.00
9	Naphthalene	0.400	0.387	3.3	93	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	91	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.370	7.5	90	0.00
12	2-Methylnaphthalene	0.400	0.369	7.8	90	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.400	0.318	20.5	79	0.00
15 S	2-Fluorobiphenyl	0.400	0.386	3.5	88	0.00
16	Acenaphthylene	0.400	0.359	10.3	85	0.00
17	Acenaphthene	0.400	0.370	7.5	85	0.00
18	Fluorene	0.400	0.364	9.0	84	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.351	12.3	66	0.00
21	4-Bromophenyl-phenylether	0.400	0.354	11.5	79	0.00
22	Hexachlorobenzene	0.400	0.370	7.5	82	0.00
23	Atrazine	0.400	0.341	14.8	83	0.00
24	Pentachlorophenol	0.400	0.363	9.3	77	0.00
25	Phenanthrene	0.400	0.372	7.0	82	0.00
26	Anthracene	0.400	0.349	12.8	80	0.00
27 SURR	Fluoranthene-d10	0.400	0.363	9.3	82	0.00
28	Fluoranthene	0.400	0.362	9.5	81	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	86	0.00
30	Pyrene	0.400	0.382	4.5	82	0.00
31 S	Terphenyl-d14	0.400	0.380	5.0	81	0.00
32	Benzo(a)anthracene	0.400	0.363	9.3	81	0.00
33	Chrysene	0.400	0.380	5.0	81	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.351	12.3	82	0.00
35 I	Perylene-d12	0.400	0.400	0.0	88	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	83	0.00
37	Benzo(b)fluoranthene	0.400	0.359	10.3	84	0.00
38	Benzo(k)fluoranthene	0.400	0.355	11.3	82	0.00
39 C	Benzo(a)pyrene	0.400	0.366	8.5	86	0.00
40	Dibenzo(a,h)anthracene	0.400	0.346	13.5	81	0.00
41	Benzo(g,h,i)perylene	0.400	0.370	7.5	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4