

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031423\
 Data File : BN024286.D
 Acq On : 14 Mar 2023 18:32
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 15 01:56:44 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 15 01:49:05 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	85	0.00
2	1,4-Dioxane	0.400	0.399	0.3	81	0.00
3	n-Nitrosodimethylamine	0.400	0.438	-9.5	92	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	82	0.00
5 S	Phenol-d6	0.400	0.372	7.0	81	0.00
6	bis(2-Chloroethyl)ether	0.400	0.402	-0.5	83	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	88	0.00
9	Naphthalene	0.400	0.389	2.8	82	0.00
10	Hexachlorobutadiene	0.400	0.394	1.5	81	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.360	10.0	77	0.00
12	2-Methylnaphthalene	0.400	0.360	10.0	77	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	74	0.00
14 S	2,4,6-Tribromophenol	0.400	0.297	25.8#	60	0.00
15 S	2-Fluorobiphenyl	0.400	0.391	2.3	73	0.00
16	Acenaphthylene	0.400	0.361	9.8	70	0.00
17	Acenaphthene	0.400	0.370	7.5	70	0.00
18	Fluorene	0.400	0.358	10.5	68	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	71	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.346	13.5	52	0.00
21	4-Bromophenyl-phenylether	0.400	0.338	15.5	62	0.00
22	Hexachlorobenzene	0.400	0.373	6.8	67	0.00
23	Atrazine	0.400	0.319	20.3	64	0.00
24	Pentachlorophenol	0.400	0.341	14.8	57	0.00
25	Phenanthrene	0.400	0.374	6.5	68	0.00
26	Anthracene	0.400	0.343	14.2	65	0.00
27 SURR	Fluoranthene-d10	0.400	0.355	11.3	65	0.00
28	Fluoranthene	0.400	0.357	10.8	65	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	69	0.00
30	Pyrene	0.400	0.388	3.0	66	0.00
31 S	Terphenyl-d14	0.400	0.371	7.3	63	0.00
32	Benzo(a)anthracene	0.400	0.348	13.0	63	0.00
33	Chrysene	0.400	0.386	3.5	66	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.347	13.3	65	0.00
35 I	Perylene-d12	0.400	0.400	0.0	70	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.357	10.8	65	0.00
37	Benzo(b)fluoranthene	0.400	0.361	9.8	67	0.00
38	Benzo(k)fluoranthene	0.400	0.370	7.5	68	0.00
39 C	Benzo(a)pyrene	0.400	0.354	11.5	66	0.00
40	Dibenzo(a,h)anthracene	0.400	0.344	14.0	63	0.00
41	Benzo(g,h,i)perylene	0.400	0.387	3.3	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4