

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031519\
 Data File : BN004784.D
 Acq On : 15 Mar 2019 22:44
 Operator : JU/SJ
 Sample : PB117973BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Mar 16 01:04:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 15 04:01:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	28923	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	126420	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.35	164	79983	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.10	188	195237	20.00	ng/ul	0.00
78) Chrysene-d12	21.30	240	254375	20.00	ng/ul	-0.02
86) Perylene-d12	23.56	264	295058	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.22	96	2256	4.40	ng/uL	0.00
5) Phenol-d5	6.87	99	48889	20.46	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.05	67	25874	20.75	ng/ul	0.00
9) 2-Chlorophenol-d4	7.24	132	45488	21.93	ng/ul	0.00
13) 4-Methylphenol-d8	8.41	113	42976	20.67	ng/ul	0.00
19) Nitrobenzene-d5	8.87	128	22175	23.01	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.59	143	25799	23.57	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.12	165	45212	21.67	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.64	131	65574	26.20	ng/ul	0.00
44) Dimethylphthalate-d6	13.76	166	174339	25.21	ng/ul	0.00
47) Acenaphthylene-d8	14.05	160	198678	23.87	ng/ul	0.00
52) 4-Nitrophenol-d4	14.56	143	22578	17.59	ng/ul	-0.01
58) Fluorene-d10	15.35	176	145154	24.70	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.47	200	26773	19.86	ng/ul	0.00
71) Anthracene-d10	17.20	188	245237	25.79	ng/ul	-0.01
79) Pyrene-d10	19.50	212	292979	23.57	ng/ul	-0.01
90) Benzo(a)pyrene-d12	23.41	264	379065	23.92	ng/ul	-0.02

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031519\
Data File : BN004784.D
Acq On : 15 Mar 2019 22:44
Operator : JU/SJ
Sample : PB117973BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Mar 16 01:04:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 15 04:01:10 2019
Response via : Initial Calibration

