

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031519\
 Data File : BN004794.D
 Acq On : 16 Mar 2019 04:38
 Operator : JU/SJ
 Sample : K1920-02
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Mar 16 05:25:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 15 04:01:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	30787	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	135674	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.35	164	85942	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.10	188	207353	20.00	ng/ul	-0.01
78) Chrysene-d12	21.30	240	269663	20.00	ng/ul	-0.02
86) Perylene-d12	23.55	264	301525	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.22	96	2342	4.29	ng/uL	0.00
5) Phenol-d5	6.88	99	15043	5.91	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.05	67	38446	28.97	ng/ul	0.00
9) 2-Chlorophenol-d4	7.24	132	55782	25.27	ng/ul	0.00
13) 4-Methylphenol-d8	8.41	113	32239	14.57	ng/ul	0.00
19) Nitrobenzene-d5	8.87	128	35064	33.90	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.59	143	40878	34.80	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.12	165	66352	29.64	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.65	131	392	0.15	ng/ul	0.00
44) Dimethylphthalate-d6	13.76	166	252918	34.04	ng/ul	0.00
47) Acenaphthylene-d8	14.05	160	302625	33.83	ng/ul	0.00
52) 4-Nitrophenol-d4	14.57	143	4946	3.59	ng/ul	0.00
58) Fluorene-d10	15.35	176	218016	34.52	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.47	200	43291	30.24	ng/ul	0.00
71) Anthracene-d10	17.20	188	360731	35.72	ng/ul	-0.01
79) Pyrene-d10	19.50	212	432686	32.84	ng/ul	-0.01
90) Benzo(a)pyrene-d12	23.41	264	537324	33.18	ng/ul	-0.02

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031519\
Data File : BN004794.D
Acq On : 16 Mar 2019 04:38
Operator : JU/SJ
Sample : K1920-02
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Mar 16 05:25:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 15 04:01:10 2019
Response via : Initial Calibration

