

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031519\
 Data File : BN004795.D
 Acq On : 16 Mar 2019 05:13
 Operator : JU/SJ
 Sample : K1920-08
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0997

Quant Time: Mar 16 05:55:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 15 04:01:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	27494	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	122001	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.35	164	76593	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.10	188	182908	20.00	ng/ul	-0.01
78) Chrysene-d12	21.30	240	233152	20.00	ng/ul	-0.02
86) Perylene-d12	23.55	264	268474	20.00	ng/ul	-0.02

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.21	96	1709	3.51	ng/uL	-0.01
5) Phenol-d5	6.88	99	10935	4.81	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.05	67	26632	22.47	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.24	132	38974	19.77	ng/ul	0.00
13) 4-Methylphenol-d8	8.41	113	22818	11.55	ng/ul	0.00
19) Nitrobenzene-d5	8.87	128	23443	25.21	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.59	143	27326	25.87	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.12	165	45569	22.64	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.65	131	281	0.12	ng/ul	0.00
44) Dimethylphthalate-d6	13.76	166	181728	27.44	ng/ul	0.00
47) Acenaphthylene-d8	14.05	160	216798	27.20	ng/ul	0.00
52) 4-Nitrophenol-d4	14.58	143	3636	2.96	ng/ul	0.00
58) Fluorene-d10	15.35	176	165950	29.48	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.47	200	32731	25.92	ng/ul	0.00
71) Anthracene-d10	17.20	188	269136	30.21	ng/ul	-0.01
79) Pyrene-d10	19.50	212	324627	28.50	ng/ul	-0.01
90) Benzo(a)pyrene-d12	23.41	264	409276	28.39	ng/ul	-0.02

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031519\
Data File : BN004795.D
Acq On : 16 Mar 2019 05:13
Operator : JU/SJ
Sample : K1920-08
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0997

Quant Time: Mar 16 05:55:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 15 04:01:10 2019
Response via : Initial Calibration

