

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031523\
 Data File : BN024294.D
 Acq On : 15 Mar 2023 15:30
 Operator : CG/JU
 Sample : 01843-06
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE132D2-20230306

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/16/2023
 Supervised By :Jagrut Upadhyay 03/16/2023

Quant Time: Mar 16 01:18:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 16 01:06:57 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.047	152	7278	0.400	ng	0.00
7) Naphthalene-d8	10.855	136	21591	0.400	ng	# 0.00
13) Acenaphthene-d10	14.675	164	10967	0.400	ng	0.00
19) Phenanthrene-d10	17.425	188	23429	0.400	ng	0.00
29) Chrysene-d12	21.610	240	15898	0.400	ng	# 0.00
35) Perylene-d12	24.123	264	12144	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.584	112	2297	0.143	ng	0.00
5) Phenol-d6	7.195	99	1664	0.080	ng	0.00
8) Nitrobenzene-d5	9.200	82	5520	0.383	ng	0.00
11) 2-Methylnaphthalene-d10	12.440	152	9695	0.365	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	1535	0.274	ng	0.00
15) 2-Fluorobiphenyl	13.307	172	15314	0.370	ng	0.00
27) Fluoranthene-d10	19.446	212	22625	0.456	ng	0.00
31) Terphenyl-d14	20.033	244	13559	0.506	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.432	88	52374	6.614	ng	Qvalue 98
25) Phenanthrene	17.462	178	1679m	0.024	ng	
34) Bis(2-ethylhexyl)phtha...	21.511	149	2464	0.099	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031523\
Data File : BN024294.D
Acq On : 15 Mar 2023 15:30
Operator : CG/JU
Sample : 01843-06
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE132D2-20230306

Quant Time: Mar 16 01:18:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 16 01:06:57 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 03/16/2023
Supervised By :Jagrut Upadhyay 03/16/2023

