


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031721\
 Data File : BN013963.D
 Acq On : 17 Mar 2021 08:49
 Operator : CG/JU
 Sample : DFTPP33
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP033

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-BN031721.M
 Title : SVOA CALIBRATION
 Last Update : Wed Mar 17 10:09:09 2021



AutoFind: Scans 2457, 2458, 2459; Background Corrected with Scan 2451

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.9	170178	PASS
68	69	0.00	2	1.6	2693	PASS
69	198	0.00	100	36.7	169496	PASS
70	69	0.00	2	0.6	953	PASS
127	198	10	80	47.0	216789	PASS
197	198	0.00	2	0.2	891	PASS
198	198	100	100	100.0	461589	PASS
199	198	5	9	6.9	31642	PASS
275	198	10	60	24.9	114712	PASS
365	198	1	100	3.3	15252	PASS
441	198	0.01	100	13.3	61205	PASS
442	198	50	100	88.1	406656	PASS
443	442	15	24	19.0	77192	PASS

Scan 2451 (17.504 min): BN013963.D
DFTPP33

m/z	Abundance
39.90	1006.0
43.90	1093.0
49.00	984.0
50.90	398.0
73.00	962.0
83.90	525.0
85.90	393.0
147.00	481.0
207.00	2648.0
208.00	633.0
209.00	329.0
281.00	406.0

Instrument :

BNA_N

ClientSampleId :

DFTPP033

m/z	Abundance
37.00	855.0
38.00	2531.0
39.00	15177.0
40.00	1626.0
41.00	663.0
44.00	1305.0
45.00	449.0
49.00	2017.0
50.00	51504.0
51.00	207360.0
52.00	10479.0
52.90	436.0
55.00	1233.0
56.00	5949.0
57.00	14590.0
57.90	700.0
60.90	2357.0
62.00	2575.0
63.00	7866.0
64.00	989.0
65.00	3915.0
66.10	313.0
68.00	3347.0
69.00	203840.0
70.00	1099.0
73.00	2267.0
74.00	21424.0
75.00	32792.0
76.00	11130.0
77.00	238656.0
78.00	16006.0
79.00	15478.0
80.00	11395.0
81.00	16480.0
82.00	4308.0
83.00	3877.0
83.90	1036.0
85.00	2911.0
86.00	5294.0
87.00	2230.0
87.90	1169.0
88.90	348.0
91.00	3525.0
92.00	3970.0
93.00	26408.0
93.90	1842.0
95.00	602.0
96.00	1242.0
97.10	420.0
98.00	20288.0
99.00	16400.0
100.00	1570.0
101.00	9783.0
102.00	591.0
103.00	2907.0
104.00	5823.0
104.90	5795.0
105.90	1911.0
107.00	72048.0
108.00	10430.0
109.00	1953.0
110.00	128672.0
111.00	20160.0
112.00	2263.0
112.90	799.0
114.90	490.0
116.00	3608.0
117.00	59872.0
118.00	4518.0
118.90	617.0
119.90	868.0
120.90	326.0
122.00	4265.0
123.00	6663.0
124.00	3177.0
125.00	2870.0
126.00	796.0
127.00	246848.0
128.00	18832.0
129.00	93432.0
130.00	7838.0
130.90	1599.0
132.00	842.0
132.80	458.0
133.90	2769.0
135.00	7717.0
136.00	3706.0
137.00	4432.0
138.00	813.0
139.00	620.0
139.90	1233.0
141.00	12325.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

141.90	3545.0
143.00	2714.0
144.00	905.0
144.90	639.0
145.90	2270.0
147.00	6527.0
148.00	14098.0
149.00	3185.0
150.00	870.0
151.00	1619.0
151.90	958.0
152.90	3998.0
153.90	2708.0
155.00	6353.0
156.00	10472.0
157.00	2070.0
157.90	2161.0
158.90	1579.0
160.00	4201.0
160.90	5547.0
161.90	1630.0
162.90	697.0
164.00	589.0
165.00	4844.0
166.00	3712.0
167.00	25384.0
168.00	14443.0
169.00	2374.0
169.90	837.0
170.90	1093.0
172.00	2227.0
173.00	2906.0
174.00	5241.0
175.00	9974.0
176.00	2823.0
177.00	4196.0
177.90	1466.0
179.00	18888.0
180.00	13024.0
181.00	6608.0
182.00	958.0
183.00	467.0
184.00	1211.0
185.00	8596.0
186.00	67576.0
187.00	18800.0
188.00	2221.0
189.00	4450.0
190.00	623.0
191.00	2187.0
192.00	5964.0
193.00	7041.0
194.00	1337.0
195.10	804.0
196.00	13898.0
197.90	499136.0
199.00	35112.0
200.00	2477.0
201.40	2167.0
203.00	3443.0
204.00	17544.0
205.00	29864.0
206.00	129640.0
207.00	20296.0
208.00	4883.0
209.00	1677.0
210.00	1892.0
211.00	4536.0
213.10	403.0
214.90	1552.0
216.00	2823.0
217.00	35328.0
217.90	4536.0
219.00	517.0
219.90	573.0
221.00	20104.0
221.80	5108.0
223.00	7721.0
224.00	71608.0
225.00	18552.0
226.00	2169.0
227.00	30792.0
228.00	4305.0
229.00	5963.0
230.00	919.0
231.00	2618.0
231.90	519.0
233.00	487.0
233.90	2034.0
235.00	2071.0
236.00	1654.0
237.00	3053.0
237.90	430.0
239.00	1280.0
239.90	978.0
240.90	1611.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

242.00	3919.0
243.00	4064.0
244.00	57264.0
245.00	7776.0
246.00	10804.0
246.90	2226.0
248.00	494.0
249.00	1759.0
250.00	406.0
251.00	500.0
252.00	339.0
253.00	1194.0
254.00	1474.0
255.00	280896.0
256.00	41848.0
257.00	3214.0
258.00	16864.0
259.00	2779.0
259.90	479.0
261.00	580.0
264.00	451.0
265.00	6311.0
266.00	950.0
269.90	308.0
271.00	529.0
272.10	711.0
273.00	8258.0
274.00	22656.0
275.00	119832.0
276.00	15864.0
277.00	10103.0
278.00	1702.0
278.90	382.0
281.00	404.0
282.10	358.0
283.00	1180.0
284.00	737.0
285.00	1735.0
285.90	306.0
288.90	362.0
290.00	422.0
291.00	335.0
292.00	369.0
293.00	2641.0
294.00	883.0
295.00	616.0
296.00	32976.0
297.00	5218.0
298.00	367.0
300.90	369.0
302.00	561.0
303.00	4015.0
304.00	1274.0
308.00	438.0
309.10	339.0
310.00	522.0
312.90	330.0
314.00	1414.0
315.00	3690.0
316.00	2122.0
317.00	371.0
320.90	1060.0
321.90	554.0
323.00	10124.0
324.10	2024.0
327.00	2151.0
327.90	1101.0
329.00	300.0
332.00	861.0
333.00	1175.0
334.00	6058.0
335.00	1696.0
341.10	1294.0
342.10	330.0
346.00	2647.0
347.00	474.0
352.00	3209.0
353.00	2001.0
354.00	3076.0
355.00	619.0
365.00	14296.0
365.90	2002.0
370.00	335.0
370.90	687.0
372.00	4849.0
373.00	1197.0
383.00	1282.0
383.90	447.0
390.00	775.0
390.80	414.0
392.00	397.0
400.90	450.0
402.00	2339.0
403.00	3292.0
404.00	1136.0
421.00	2456.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

422.00	2573.0
423.00	14899.0
424.00	5463.0
425.00	890.0
441.00	51176.0
442.00	330688.0
443.00	74176.0
444.00	5566.0
445.00	401.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

m/z	Abundance
37.00	727.0
38.00	2268.0
39.00	14291.0
40.00	1406.0
41.00	810.0
44.00	1251.0
45.00	396.0
49.00	1719.0
50.00	50208.0
51.00	196608.0
52.00	9584.0
53.00	484.0
55.00	1247.0
56.00	5872.0
57.00	13091.0
57.90	636.0
60.90	2287.0
62.00	2726.0
63.00	7844.0
64.00	1204.0
65.00	3789.0
66.90	335.0
67.90	2959.0
69.00	193280.0
70.00	1044.0
73.00	2273.0
74.00	20680.0
75.00	32408.0
76.00	10311.0
77.00	230208.0
78.00	15648.0
79.00	14495.0
80.00	11189.0
81.00	16273.0
82.00	4295.0
83.00	4084.0
84.00	1056.0
84.90	2830.0
86.00	4825.0
87.00	2336.0
87.90	735.0
88.90	313.0
91.00	3666.0
92.00	4158.0
93.00	25624.0
94.00	1659.0
95.00	439.0
96.00	1494.0
97.00	553.0
98.00	20008.0
99.00	15918.0
100.00	1424.0
101.00	9439.0
101.90	528.0
103.00	3026.0
103.90	5514.0
105.00	5103.0
106.00	1752.0
107.00	69648.0
108.00	10729.0
108.90	1982.0
110.00	130744.0
111.00	19696.0
111.90	2611.0
112.90	892.0
115.00	379.0
116.00	4005.0
117.00	59544.0
118.00	4087.0
118.90	695.0
119.90	838.0
120.90	430.0
121.90	4695.0
123.00	7340.0
123.90	3041.0
125.00	3001.0
126.00	964.0
127.00	255232.0
128.00	18464.0
129.00	95416.0
129.90	7668.0
131.00	1500.0
132.00	740.0
132.90	599.0
133.90	2851.0
135.00	8572.0
136.00	3302.0
137.00	4422.0
137.90	1001.0
138.80	532.0
139.90	1024.0
141.00	12431.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

142.00	4227.0
143.00	3102.0
144.00	937.0
144.90	680.0
146.10	2275.0
147.00	6619.0
147.90	16392.0
149.00	3282.0
150.00	793.0
151.10	1487.0
151.70	1020.0
152.90	4118.0
154.00	3170.0
155.00	6888.0
156.00	10581.0
157.00	2169.0
157.90	2574.0
159.00	1770.0
159.90	4163.0
161.00	5970.0
161.90	2017.0
162.90	560.0
163.90	634.0
165.00	5021.0
166.00	3890.0
167.00	25896.0
168.00	14671.0
169.00	2582.0
169.90	855.0
171.00	987.0
171.90	2484.0
173.00	3485.0
174.00	6015.0
175.00	9804.0
176.00	2725.0
177.00	5241.0
177.90	1675.0
179.00	18928.0
180.00	14060.0
181.00	6261.0
182.00	1017.0
182.80	661.0
184.00	1391.0
185.00	9315.0
186.00	74112.0
187.00	22160.0
188.00	2549.0
188.90	4644.0
189.90	824.0
191.00	2448.0
192.00	6832.0
193.00	7394.0
194.00	1273.0
194.90	710.0
196.00	14258.0
196.80	2673.0
197.90	529920.0
199.00	36720.0
199.90	3063.0
201.50	2686.0
203.00	3950.0
204.00	19680.0
205.00	32824.0
206.10	137664.0
207.00	21176.0
208.00	4836.0
209.00	1977.0
210.20	2338.0
211.00	5662.0
212.90	509.0
214.90	1619.0
216.00	3159.0
217.00	36456.0
218.00	4670.0
218.90	553.0
221.10	22312.0
221.80	4741.0
223.00	9079.0
224.00	79200.0
225.00	20000.0
226.00	2294.0
227.00	34912.0
228.00	4664.0
228.90	6969.0
230.00	844.0
231.10	2863.0
231.90	503.0
233.00	580.0
234.00	2368.0
235.00	2640.0
236.00	1546.0
237.00	2972.0
238.00	359.0
238.90	1276.0
240.00	1073.0
240.90	1695.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

242.00	3938.0
243.00	4558.0
244.00	67880.0
245.10	7845.0
246.00	11965.0
247.00	2400.0
247.90	452.0
248.90	2298.0
250.90	497.0
251.90	484.0
253.00	1183.0
254.00	1664.0
255.00	328576.0
256.00	46544.0
257.00	3647.0
258.00	18536.0
259.00	2640.0
260.10	576.0
261.00	478.0
263.90	473.0
265.00	7765.0
266.00	982.0
269.90	374.0
271.00	586.0
271.90	813.0
273.00	9609.0
274.00	26456.0
275.00	134208.0
276.00	20008.0
277.00	11617.0
278.00	2103.0
279.00	396.0
281.00	336.0
282.00	339.0
282.90	1497.0
284.00	808.0
285.00	2188.0
285.90	403.0
289.00	476.0
290.00	382.0
292.00	591.0
293.00	2880.0
294.00	699.0
295.00	561.0
296.00	37592.0
297.00	5538.0
298.00	374.0
300.90	515.0
302.00	507.0
303.00	4831.0
304.10	1133.0
304.90	360.0
308.00	447.0
309.00	419.0
309.90	569.0
313.00	381.0
314.00	1950.0
315.00	4410.0
316.00	2880.0
316.90	366.0
321.00	1388.0
322.00	685.0
323.00	13377.0
324.00	2508.0
327.00	2590.0
328.00	1354.0
332.00	1015.0
333.00	1300.0
334.00	7704.0
335.00	2273.0
341.10	1567.0
341.90	427.0
345.90	2388.0
346.90	659.0
352.10	3742.0
353.00	2754.0
354.00	3535.0
355.00	999.0
365.00	18240.0
365.90	2601.0
370.00	385.0
370.90	893.0
372.00	6775.0
373.10	1775.0
383.00	1974.0
384.00	624.0
390.00	804.0
390.90	595.0
391.90	436.0
402.00	3192.0
403.00	3717.0
404.00	1213.0
421.00	3618.0
422.00	3177.0
423.00	23504.0
424.00	6762.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

425.00	681.0
441.00	72528.0
442.00	541440.0
443.00	88792.0
444.00	9491.0
444.90	644.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

m/z	Abundance
37.00	467.0
38.00	1277.0
39.00	7706.0
40.00	1319.0
41.00	491.0
43.00	334.0
43.90	1111.0
45.00	378.0
49.00	1478.0
50.00	27296.0
51.00	107760.0
52.00	5224.0
55.00	735.0
56.00	2775.0
57.00	7229.0
61.00	1404.0
62.00	1482.0
63.00	4146.0
64.00	733.0
65.00	2141.0
68.00	1775.0
69.00	111368.0
70.00	718.0
72.90	1308.0
74.00	10724.0
75.00	18376.0
76.00	6027.0
77.00	126512.0
78.00	8224.0
79.00	8566.0
80.00	5805.0
81.00	9044.0
81.90	2197.0
83.00	2289.0
83.90	839.0
85.00	1856.0
86.00	3146.0
87.00	1466.0
87.90	524.0
88.90	325.0
91.00	2030.0
91.90	2359.0
93.00	14532.0
94.00	949.0
95.90	949.0
97.10	446.0
98.00	11177.0
99.00	9918.0
100.00	909.0
101.00	5461.0
102.90	1947.0
104.00	2952.0
105.00	3182.0
106.00	795.0
107.00	40840.0
108.00	6358.0
109.00	1241.0
110.00	74072.0
111.00	11574.0
112.00	1567.0
112.90	500.0
116.00	2333.0
117.00	35752.0
117.90	2532.0
119.10	454.0
120.00	640.0
121.90	2627.0
123.00	3994.0
124.00	1699.0
124.90	1630.0
126.00	336.0
127.00	148288.0
128.00	11297.0
129.00	57152.0
129.90	4781.0
131.00	925.0
132.00	581.0
132.90	410.0
133.90	1773.0
135.00	4632.0
135.90	1817.0
137.00	2732.0
137.90	481.0
138.80	458.0
139.90	609.0
141.00	7338.0
141.90	2496.0
142.90	1570.0
143.90	495.0
145.00	517.0
145.90	1523.0
147.00	3978.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

147.90	9399.0
149.00	2100.0
149.90	574.0
151.00	960.0
152.90	2427.0
153.90	1953.0
155.00	4337.0
156.00	6592.0
157.10	1311.0
158.00	1502.0
158.90	1054.0
159.90	2801.0
161.00	3991.0
161.90	1313.0
162.90	334.0
163.90	330.0
165.00	2723.0
166.00	2294.0
167.00	16960.0
168.00	10029.0
169.00	1568.0
169.90	498.0
171.00	671.0
171.90	1412.0
173.00	1803.0
174.00	3421.0
175.00	7035.0
176.00	1918.0
177.00	2971.0
177.90	1121.0
178.90	12096.0
180.00	8646.0
181.00	4394.0
182.00	689.0
183.10	373.0
183.90	871.0
185.00	6214.0
186.00	45464.0
187.00	12393.0
188.00	1572.0
188.90	2760.0
189.90	396.0
191.00	1353.0
192.00	4307.0
193.00	4991.0
194.00	885.0
195.00	624.0
196.00	8624.0
198.00	355712.0
198.90	23096.0
200.00	1523.0
201.40	1736.0
202.90	2605.0
204.00	12408.0
205.00	21568.0
206.00	90624.0
207.00	14357.0
208.00	3093.0
209.00	1353.0
209.90	1246.0
211.00	3698.0
214.90	1010.0
216.00	1904.0
217.00	23168.0
217.90	2972.0
219.00	360.0
221.00	13752.0
221.90	2975.0
223.00	5600.0
224.00	53752.0
225.00	12534.0
226.00	1245.0
227.00	21360.0
228.00	2852.0
229.00	4374.0
230.00	532.0
231.00	2033.0
233.00	459.0
234.00	1250.0
235.00	1788.0
236.00	1272.0
237.00	2188.0
238.90	696.0
239.90	686.0
240.90	1245.0
242.00	2958.0
243.00	2591.0
244.00	39736.0
245.00	5749.0
246.00	8293.0
246.90	1364.0
247.90	580.0
248.90	1458.0
250.00	335.0
251.00	377.0
251.90	331.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033

253.00	876.0
254.00	1177.0
255.00	209792.0
256.00	28144.0
257.00	2096.0
258.00	12445.0
259.00	1895.0
260.90	455.0
263.90	303.0
265.00	4877.0
266.00	673.0
271.00	451.0
271.90	538.0
273.00	6819.0
274.00	17112.0
275.00	90096.0
276.00	10844.0
277.00	7593.0
278.00	1259.0
281.00	415.0
282.00	301.0
283.00	741.0
283.90	664.0
285.00	1516.0
289.00	407.0
292.00	308.0
293.00	1845.0
294.00	498.0
295.00	473.0
296.00	26744.0
297.00	3773.0
301.00	310.0
302.10	488.0
303.00	3549.0
304.00	1032.0
307.90	490.0
308.90	319.0
310.00	346.0
314.00	1266.0
315.00	3170.0
316.00	1655.0
317.00	314.0
321.00	901.0
322.10	465.0
323.00	8943.0
324.00	1705.0
327.00	1604.0
328.00	939.0
332.00	594.0
333.00	986.0
334.00	6285.0
335.00	1406.0
341.00	1106.0
345.90	1857.0
346.90	431.0
352.00	2732.0
353.00	2021.0
354.00	2853.0
355.00	547.0
359.00	374.0
364.90	13221.0
365.90	1855.0
370.10	362.0
371.00	571.0
372.00	4696.0
373.10	974.0
383.00	1526.0
383.90	430.0
390.00	620.0
391.00	544.0
402.00	2147.0
402.90	3048.0
404.00	1094.0
421.00	2354.0
422.10	2276.0
423.00	17344.0
424.00	4383.0
424.90	569.0
441.00	59912.0
442.00	347840.0
443.00	68608.0
444.00	6293.0
444.90	320.0

Instrument :
BNA_N
ClientSampleId :
DFTPP033