

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN031723\  
 Data File : BN024369.D  
 Acq On : 18 Mar 2023 07:05  
 Operator : CG/JU  
 Sample : 01879-03  
 Misc :  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 ClientSampleId :  
 RE109D3-20230308

Quant Time: Mar 20 01:07:40 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN031323.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Sat Mar 18 05:49:08 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-------------------------------|--------|------|----------|-------|-------|----------|
| -----                         |        |      |          |       |       |          |
| Internal Standards            |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 8.040  | 152  | 9025     | 0.400 | ng    | 0.00     |
| 7) Naphthalene-d8             | 10.855 | 136  | 26494    | 0.400 | ng    | 0.00     |
| 13) Acenaphthene-d10          | 14.665 | 164  | 12106    | 0.400 | ng    | 0.00     |
| 19) Phenanthrene-d10          | 17.413 | 188  | 24554    | 0.400 | ng    | # 0.00   |
| 29) Chrysene-d12              | 21.601 | 240  | 15349    | 0.400 | ng    | # 0.00   |
| 35) Perylene-d12              | 24.094 | 264  | 11072    | 0.400 | ng    | # 0.00   |
| System Monitoring Compounds   |        |      |          |       |       |          |
| 4) 2-Fluorophenol             | 5.577  | 112  | 2477     | 0.124 | ng    | 0.00     |
| 5) Phenol-d6                  | 7.188  | 99   | 1827     | 0.071 | ng    | 0.00     |
| 8) Nitrobenzene-d5            | 9.201  | 82   | 6469     | 0.365 | ng    | 0.00     |
| 11) 2-Methylnaphthalene-d10   | 12.433 | 152  | 10309    | 0.317 | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol      | 16.159 | 330  | 1565     | 0.253 | ng    | 0.00     |
| 15) 2-Fluorobiphenyl          | 13.296 | 172  | 17578    | 0.384 | ng    | 0.00     |
| 27) Fluoranthene-d10          | 19.442 | 212  | 19827    | 0.381 | ng    | 0.00     |
| 31) Terphenyl-d14             | 20.034 | 244  | 10729    | 0.415 | ng    | 0.00     |
| Target Compounds              |        |      |          |       |       |          |
| 2) 1,4-Dioxane                | 3.432  | 88   | 28430    | 2.895 | ng    | 94       |
| 34) Bis(2-ethylhexyl)phtha... | 21.502 | 149  | 2427     | 0.101 | ng    | 98       |
| -----                         |        |      |          |       |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_N\Data\BN031723\  
Data File : BN024369.D  
Acq On : 18 Mar 2023 07:05  
Operator : CG/JU  
Sample : 01879-03  
Misc :  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
BNA\_N  
ClientSampleId :  
RE109D3-20230308

Quant Time: Mar 20 01:07:40 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_N\Methods\8270-SIM-BN031323.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Sat Mar 18 05:49:08 2023  
Response via : Initial Calibration

