

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031723\
 Data File : BN024373.D
 Acq On : 18 Mar 2023 09:32
 Operator : CG/JU
 Sample : 01879-18
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RE134D3-20230309

Quant Time: Mar 20 01:08:30 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 18 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	8350	0.400	ng	0.00
7) Naphthalene-d8	10.855	136	24941	0.400	ng	0.00
13) Acenaphthene-d10	14.665	164	11539	0.400	ng	0.00
19) Phenanthrene-d10	17.413	188	23364	0.400	ng	# 0.00
29) Chrysene-d12	21.610	240	13759	0.400	ng	0.00
35) Perylene-d12	24.117	264	9569	0.400	ng	# 0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.577	112	3077	0.166	ng	0.00
5) Phenol-d6	7.188	99	2267	0.095	ng	0.00
8) Nitrobenzene-d5	9.200	82	7189	0.431	ng	0.00
11) 2-Methylnaphthalene-d10	12.433	152	11577	0.378	ng	0.00
14) 2,4,6-Tribromophenol	16.159	330	1561	0.265	ng	0.00
15) 2-Fluorobiphenyl	13.296	172	20582	0.472	ng	0.00
27) Fluoranthene-d10	19.442	212	18803	0.380	ng	0.00
31) Terphenyl-d14	20.034	244	11492	0.495	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.432	88	82687	9.101	ng	96
34) Bis(2-ethylhexyl)phtha...	21.502	149	2372	0.110	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031723\
Data File : BN024373.D
Acq On : 18 Mar 2023 09:32
Operator : CG/JU
Sample : 01879-18
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RE134D3-20230309

Quant Time: Mar 20 01:08:30 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 18 05:49:08 2023
Response via : Initial Calibration

