



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031723\
Data File : BN024384.D
Acq On : 18 Mar 2023 16:49
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN031323.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Sat Mar 18 05:49:08 2023






AutoFind: Scans 2475, 2476, 2477; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	59.1	250155	PASS
68	69	0.00	2	1.7	3618	PASS
69	198	0.00	100	49.9	210923	PASS
70	69	0.00	2	0.5	1094	PASS
127	198	10	80	54.8	231765	PASS
197	198	0.00	2	0.3	1277	PASS
198	198	100	100	100.0	423104	PASS
199	198	5	9	6.8	28973	PASS
275	198	10	60	23.5	99395	PASS
365	198	1	100	3.2	13548	PASS
441	198	0.01	100	9.3	39203	PASS
442	442	50	100	100.0	258176	PASS
443	442	15	24	20.3	52419	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031723\
 Data File : BN024384.D
 Acq On : 18 Mar 2023 16:49
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Mar 20 02:11:34 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Mar 20 02:11:29 2023
 Response via : Initial Calibration



TIC: BN024384.D\data.ms

(70) Pentachlorophenol (C)

17.063min (0.000) 21258.01 ng




response 464281

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.37
264.00	61.60	63.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031723\
Data File : BN024384.D
Acq On : 18 Mar 2023 16:49
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Mar 20 02:11:34 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Mar 20 02:11:29 2023
Response via : Initial Calibration



TIC: BN024384.D\data.ms

(77) Benzidine

19.657min (0.000) 0.00 ng

response 2303421

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.03
183.00	13.20	11.35
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
3/17/2023	BNA_N	<u>BN024384.D</u>
Compound Name	Response	Retention Time
DDT	1305664	20.886
DDD	24738	20.445
DDE	1850	19.939
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
26588	1332252	2.00

Instrument :
BNA_N
ClientSampleId :
DFTPP