

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031819\
 Data File : BN004807.D
 Acq On : 18 Mar 2019 20:53
 Operator : JU/SJ
 Sample : K1920-13
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C09B4

Quant Time: Mar 19 00:40:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Mar 15 04:01:10 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	25491	20.00	ng/ul	0.00
18) Naphthalene-d8	10.49	136	111472	20.00	ng/ul	-0.01
36) Acenaphthene-d10	14.35	164	68390	20.00	ng/ul	-0.01
62) Phenanthrene-d10	17.10	188	165243	20.00	ng/ul	-0.01
78) Chrysene-d12	21.30	240	213514	20.00	ng/ul	-0.02
86) Perylene-d12	23.56	264	239419	20.00	ng/ul	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.22	96	1861	4.12	ng/uL	0.00
5) Phenol-d5	6.88	99	12116	5.75	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.05	67	30982	28.20	ng/ul	-0.01
9) 2-Chlorophenol-d4	7.24	132	44771	24.50	ng/ul	0.00
13) 4-Methylphenol-d8	8.41	113	25670	14.01	ng/ul	0.00
19) Nitrobenzene-d5	8.87	128	26999	31.77	ng/ul	-0.01
22) 2-Nitrophenol-d4	9.59	143	31196	32.33	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.12	165	51635	28.07	ng/ul	-0.01
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
44) Dimethylphthalate-d6	13.76	166	191597	32.40	ng/ul	0.00
47) Acenaphthylene-d8	14.05	160	227943	32.02	ng/ul	0.00
52) 4-Nitrophenol-d4	14.57	143	3649	3.32	ng/ul	0.00
58) Fluorene-d10	15.35	176	168697	33.57	ng/ul	-0.01
63) 4,6-Dinitro-2-methylphenol	15.47	200	30062	26.35	ng/ul	0.00
71) Anthracene-d10	17.20	188	279282	34.70	ng/ul	-0.01
79) Pyrene-d10	19.50	212	337181	32.32	ng/ul	-0.02
90) Benzo(a)pyrene-d12	23.41	264	415904	32.35	ng/ul	-0.02

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN031819\
Data File : BN004807.D
Acq On : 18 Mar 2019 20:53
Operator : JU/SJ
Sample : K1920-13
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C09B4

Quant Time: Mar 19 00:40:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Mar 15 04:01:10 2019
Response via : Initial Calibration

