


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN031819\
 Data File : BN004802.D
 Acq On : 18 Mar 2019 17:25
 Operator : JU/SJ
 Sample : DFTPP21
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP21

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN030719MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Mar 15 04:01:10 2019



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	23.3	11486	PASS
68	69	0.00	2	0.8	107	PASS
69	198	0.00	100	27.8	13685	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.1	20261	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	49266	PASS
199	198	5	9	6.7	3318	PASS
275	198	10	60	32.8	16170	PASS
365	198	1	100	4.4	2189	PASS
441	198	0.01	100	14.8	7276	PASS
442	198	50	100	94.1	46338	PASS
443	442	15	24	19.3	8940	PASS

m/z	Abundance
39.00	795.0
40.00	1272.0
50.00	3017.0
51.00	11258.0
52.00	573.0
56.00	342.0
57.00	871.0
63.00	492.0
69.00	13008.0
74.00	1371.0
75.00	2080.0
76.00	685.0
77.00	14469.0
78.00	1079.0
79.00	1064.0
80.00	803.0
81.00	1085.0
85.90	342.0
91.90	333.0
93.00	1881.0
98.00	1403.0
99.00	1245.0
100.90	638.0
103.90	376.0
105.00	377.0
107.00	5076.0
108.00	722.0
110.00	9193.0
111.00	1337.0
116.00	323.0
117.00	4097.0
121.90	371.0
123.00	542.0
127.00	18016.0
128.00	1408.0
129.00	7103.0
130.00	643.0
135.00	575.0
137.00	339.0
141.00	991.0
142.00	328.0
147.00	494.0
148.00	1048.0
153.00	328.0
155.00	508.0
156.00	862.0
160.00	353.0
160.90	440.0
165.00	386.0
167.00	1781.0
168.00	1308.0
174.00	375.0
175.00	820.0
177.00	411.0
178.90	1491.0
180.00	1150.0
181.00	478.0
185.00	732.0
186.00	5705.0
187.00	1584.0
188.90	399.0
191.90	446.0
193.00	484.0
196.00	1116.0
198.00	41304.0
199.00	2765.0
202.90	309.0
204.00	1427.0
205.00	2571.0
206.00	10835.0
207.00	1601.0
208.00	421.0
211.00	412.0
217.00	2789.0
217.90	388.0
221.00	1720.0
223.00	643.0
224.00	6002.0
225.00	1589.0
227.00	2581.0
228.00	419.0
229.00	554.0
237.00	326.0
243.10	373.0
244.00	4835.0
245.00	655.0
246.00	920.0
255.00	25200.0
256.00	3837.0
257.00	384.0
258.00	1388.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

265.00	601.0
273.00	727.0
274.00	2084.0
275.00	12303.0
276.00	1676.0
277.00	1006.0
296.00	3080.0
297.00	458.0
303.00	353.0
315.00	339.0
323.00	1081.0
334.00	667.0
352.00	317.0
354.00	353.0
365.00	1438.0
372.00	542.0
403.00	321.0
423.00	1860.0
424.00	700.0
441.00	4231.0
442.10	27168.0
443.00	5706.0
444.00	499.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

m/z	Abundance
39.00	1046.0
40.00	1172.0
50.00	3783.0
51.00	13886.0
52.00	702.0
56.00	406.0
57.00	974.0
63.00	612.0
65.00	338.0
68.10	322.0
69.00	16608.0
74.00	1719.0
75.00	2731.0
76.00	969.0
77.10	18480.0
78.00	1277.0
79.00	1401.0
80.00	1046.0
81.00	1453.0
82.00	342.0
83.00	334.0
85.00	332.0
86.00	450.0
90.90	349.0
92.00	407.0
93.00	2483.0
98.00	1869.0
99.00	1532.0
101.00	876.0
103.90	592.0
105.00	507.0
107.00	6647.0
108.00	1004.0
110.00	12542.0
111.00	1851.0
116.00	408.0
117.00	5543.0
117.90	386.0
122.00	512.0
123.00	743.0
124.00	365.0
127.00	25008.0
128.00	1890.0
129.00	9987.0
130.00	882.0
135.00	836.0
136.00	358.0
137.00	454.0
141.00	1394.0
142.00	445.0
147.00	746.0
148.00	1541.0
149.00	333.0
152.90	457.0
154.00	347.0
155.00	737.0
156.00	1080.0
159.90	465.0
161.00	623.0
164.90	567.0
166.00	425.0
167.00	2680.0
168.00	1533.0
173.00	305.0
174.00	626.0
175.00	1082.0
176.90	542.0
179.00	1963.0
180.00	1574.0
181.00	645.0
185.00	1036.0
186.00	8139.0
187.00	2440.0
189.00	503.0
191.90	659.0
193.00	753.0
196.00	1697.0
198.00	60112.0
199.00	4144.0
200.00	332.0
201.40	314.0
203.00	439.0
204.00	2084.0
205.00	3813.0
206.00	16100.0
207.00	2203.0
208.00	648.0
211.00	646.0
216.00	373.0
217.00	4232.0
218.00	556.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

221.00	2655.0
223.00	976.0
224.00	9111.0
225.00	2544.0
227.00	3979.0
228.00	533.0
229.00	873.0
230.90	370.0
237.00	370.0
242.00	510.0
243.00	583.0
244.00	7611.0
245.00	1032.0
246.00	1484.0
255.00	40784.0
256.00	5834.0
256.90	540.0
258.00	2229.0
258.90	351.0
265.00	841.0
273.00	1252.0
274.00	3199.0
275.00	19552.0
276.00	2651.0
277.00	1574.0
292.90	422.0
296.00	5092.0
297.00	724.0
303.00	683.0
314.90	633.0
316.00	368.0
323.00	1820.0
327.00	311.0
334.00	1268.0
334.90	314.0
345.90	425.0
352.00	583.0
353.00	381.0
354.00	673.0
365.00	2649.0
365.90	354.0
372.00	899.0
401.90	410.0
403.00	703.0
421.00	481.0
422.00	464.0
423.00	3795.0
424.00	1196.0
441.10	8633.0
442.00	53656.0
443.00	10550.0
444.10	880.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

m/z	Abundance
39.00	637.0
40.00	1154.0
50.00	2312.0
51.00	9316.0
52.00	524.0
57.00	701.0
63.00	427.0
69.00	11439.0
74.00	1213.0
75.00	1912.0
76.00	618.0
77.00	12621.0
78.00	908.0
79.00	1010.0
79.90	722.0
81.00	994.0
86.00	310.0
93.00	1653.0
98.00	1383.0
99.00	1104.0
100.90	552.0
104.00	432.0
105.00	370.0
107.00	4790.0
108.00	688.0
110.00	8701.0
111.00	1381.0
117.00	3870.0
121.90	344.0
123.00	504.0
127.00	17760.0
128.00	1382.0
129.00	7185.0
130.00	633.0
135.00	613.0
141.00	1020.0
147.00	541.0
148.00	1063.0
152.90	338.0
155.00	568.0
156.00	858.0
160.00	321.0
160.90	477.0
164.90	364.0
166.00	323.0
167.00	2044.0
168.00	1027.0
173.90	519.0
175.00	882.0
177.00	418.0
179.00	1564.0
180.00	1122.0
181.00	515.0
185.00	754.0
186.00	6051.0
187.00	1799.0
189.00	435.0
192.00	474.0
193.00	649.0
196.00	1381.0
198.00	46384.0
199.00	3046.0
203.00	338.0
204.00	1701.0
205.00	2812.0
206.00	12671.0
207.00	1651.0
208.00	468.0
211.00	502.0
217.00	3510.0
217.90	370.0
221.00	2184.0
223.00	779.0
224.00	7618.0
225.00	1985.0
227.00	3041.0
228.00	457.0
229.00	684.0
236.90	313.0
242.00	374.0
243.00	434.0
244.00	6307.0
245.00	830.0
246.00	1267.0
255.00	33368.0
256.00	4706.0
257.00	328.0
258.00	1846.0
258.90	349.0
265.00	766.0
273.00	1110.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

274.00	2779.0
275.00	16656.0
276.00	2192.0
277.00	1332.0
292.90	337.0
296.00	4524.0
296.90	645.0
303.00	596.0
315.00	499.0
315.90	332.0
323.00	1611.0
324.00	302.0
334.00	1084.0
345.90	373.0
352.00	503.0
352.90	363.0
354.00	566.0
365.00	2480.0
365.90	317.0
372.00	984.0
402.00	478.0
403.00	602.0
421.00	509.0
422.10	544.0
423.00	3741.0
424.00	984.0
441.10	8965.0
442.00	58192.0
443.10	10565.0
444.00	1000.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21

m/z	Abundance
40.00	1389.0

Instrument :
BNA_N
ClientSampleId :
DFTPP21