

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032019\
 Data File : BN004836.D
 Acq On : 21 Mar 2019 01:16
 Operator : JU/SJ
 Sample : K1842-07
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 FB-030719

Quant Time: Mar 21 09:18:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	1343	0.40	ng	0.00
7) Naphthalene-d8	10.49	136	5694	0.40	ng	0.00
13) Acenaphthene-d10	14.34	164	3616	0.40	ng	0.00
19) Phenanthrene-d10	17.10	188	10075	0.40	ng	0.01
27) Chrysene-d12	21.29	240	13938	0.40	ng	0.00
34) Perylene-d12	23.54	264	15762	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.30	112	628	0.17	ng	0.00
5) Phenol-d6	6.90	99	464	0.10	ng	0.02
8) Nitrobenzene-d5	8.87	82	1315	0.35	ng	0.00
11) 2-Methylnaphthalene-d10	12.09	152	3573	0.36	ng	0.00
14) 2,4,6-Tribromophenol	15.86	330	842	0.34	ng	0.01
15) 2-Fluorobiphenyl	12.96	172	6140	0.41	ng	0.00
25) Fluoranthene-d10	19.13	212	13039	0.39	ng	0.00
29) Terphenyl-d14	19.73	244	13056	0.46	ng	0.00
Target Compounds						Qvalue
32) Bis(2-ethylhexyl)phthalate	21.18	149	902	0.049	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032019\
Data File : BN004836.D
Acq On : 21 Mar 2019 01:16
Operator : JU/SJ
Sample : K1842-07
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
FB-030719

Quant Time: Mar 21 09:18:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

