

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032019\
 Data File : BN004824.D
 Acq On : 20 Mar 2019 18:09
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 09:00:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.455	0.440	3.3	95	0.00
3	n-Nitrosodimethylamine	0.269	0.240	10.8	96	0.00
4 S	2-Fluorophenol	1.087	1.051	3.3	94	0.00
5 S	Phenol-d6	1.354	1.318	2.7	98	0.00
6	bis(2-Chloroethyl)ether	1.221	1.212	0.7	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.265	0.250	5.7	97	0.00
9	Naphthalene	1.122	1.125	-0.3	101	-0.01
10	Hexachlorobutadiene	0.214	0.221	-3.3	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.698	0.701	-0.4	102	0.00
12	2-Methylnaphthalene	0.833	0.834	-0.1	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.276	0.246	10.9	97	0.00
15 S	2-Fluorobiphenyl	1.653	1.671	-1.1	101	0.00
16	Acenaphthylene	2.060	1.959	4.9	102	0.00
17	Acenaphthene	1.349	1.366	-1.3	102	0.00
18	Fluorene	1.883	1.903	-1.1	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.261	0.257	1.5	102	0.00
21	Hexachlorobenzene	0.299	0.301	-0.7	102	0.00
22	Pentachlorophenol	0.058	0.054	6.9	109	0.00
23	Phenanthrene	1.292	1.279	1.0	101	0.00
24	Anthracene	1.154	1.097	4.9	100	0.00
25 SURR	Fluoranthene-d10	1.339	1.315	1.8	101	0.00
26	Fluoranthene	1.660	1.644	1.0	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
28	Pvrene	1.220	1.210	0.8	100	0.00
29 S	Terphenyl-d14	0.819	0.814	0.6	101	0.00
30	Benzo(a)anthracene	1.312	1.320	-0.6	105	0.00
31	Chrysene	1.382	1.380	0.1	100	0.00
32	Bis(2-ethylhexyl)phthalate	0.524	0.495	5.5	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.999	2.166	-8.4	108	0.00
34 I	Perylene-d12	1.000	1.000	0.0	107	0.00
35	Benzo(b)fluoranthene	1.447	1.461	-1.0	106	0.00
36	Benzo(k)fluoranthene	1.424	1.438	-1.0	110	0.00
37 C	Benzo(a)pyrene	1.327	1.332	-0.4	108	0.00
38	Dibenzo(a,h)anthracene	1.537	1.561	-1.6	107	0.00
39	Benzo(a,h,i)perylene	1.566	1.590	-1.5	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032019\
Data File : BN004824.D
Acq On : 20 Mar 2019 18:09
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 21 09:00:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			