

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032019\
 Data File : BN004841.D
 Acq On : 21 Mar 2019 04:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 21 09:22:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	0.00
2	1,4-Dioxane	0.455	0.410	9.9	113	0.00
3	n-Nitrosodimethylamine	0.269	0.201	25.3#	103	0.00
4 S	2-Fluorophenol	1.087	1.054	3.0	121	0.00
5 S	Phenol-d6	1.354	1.316	2.8	124	0.00
6	bis(2-Chloroethyl)ether	1.221	1.196	2.0	128	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	130	0.00
8 S	Nitrobenzene-d5	0.265	0.245	7.5	121	0.00
9	Naphthalene	1.122	1.141	-1.7	131	0.00
10	Hexachlorobutadiene	0.214	0.220	-2.8	133	0.00
11 SURR	2-Methylnaphthalene-d10	0.698	0.708	-1.4	132	0.00
12	2-Methylnaphthalene	0.833	0.844	-1.3	131	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	129	0.00
14 S	2,4,6-Tribromophenol	0.276	0.242	12.3	120	0.00
15 S	2-Fluorobiphenyl	1.653	1.702	-3.0	131	0.00
16	Acenaphthylene	2.060	1.987	3.5	131	0.00
17	Acenaphthene	1.349	1.371	-1.6	130	0.00
18	Fluorene	1.883	1.910	-1.4	129	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	130	0.00
20	4-Bromophenyl-phenylether	0.261	0.255	2.3	128	0.00
21	Hexachlorobenzene	0.299	0.302	-1.0	130	0.00
22	Pentachlorophenol	0.058	0.045	22.4	117	0.01
23	Phenanthrene	1.292	1.286	0.5	129	0.00
24	Anthracene	1.154	1.114	3.5	129	0.00
25 SURR	Fluoranthene-d10	1.339	1.311	2.1	129	0.00
26	Fluoranthene	1.660	1.643	1.0	128	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	127	0.00
28	Pvrene	1.220	1.249	-2.4	128	0.00
29 S	Terphenyl-d14	0.819	0.831	-1.5	127	0.00
30	Benzo(a)anthracene	1.312	1.298	1.1	127	0.00
31	Chrysene	1.382	1.444	-4.5	129	-0.01
32	Bis(2-ethylhexyl)phthalate	0.524	0.508	3.1	134	0.00
33	Indeno(1,2,3-cd)pyrene	1.999	2.055	-2.8	126	0.00
34 I	Perylene-d12	1.000	1.000	0.0	130	0.00
35	Benzo(b)fluoranthene	1.447	1.458	-0.8	129	0.00
36	Benzo(k)fluoranthene	1.424	1.410	1.0	130	0.00
37 C	Benzo(a)pyrene	1.327	1.329	-0.2	131	0.00
38	Dibenzo(a,h)anthracene	1.537	1.502	2.3	125	0.00
39	Benzo(a,h,i)perylene	1.566	1.517	3.1	124	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032019\
Data File : BN004841.D
Acq On : 21 Mar 2019 04:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 21 09:22:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			