

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032119\
 Data File : BN004871.D
 Acq On : 22 Mar 2019 03:38
 Operator : JU/SJ
 Sample : K2013-07
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PST-026-B027-031819

Manual Integrations
 APPROVED

Sohil
 3/22/2019 7:16:37 PM

Quant Time: Mar 22 04:57:02 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Mar 20 16:06:20 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	1854	0.40	ng	0.00
7) Naphthalene-d8	10.49	136	8573	0.40	ng	0.00
13) Acenaphthene-d10	14.33	164	5490	0.40	ng	-0.01
19) Phenanthrene-d10	17.09	188	14340	0.40	ng	0.00
27) Chrysene-d12	21.29	240	19437	0.40	ng	0.00
34) Perylene-d12	23.54	264	20884	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.29	112	2044	0.41	ng	0.00
5) Phenol-d6	6.87	99	2740	0.44	ng	0.00
8) Nitrobenzene-d5	8.87	82	1821	0.32	ng	0.00
11) 2-Methylnaphthalene-d10	12.08	152	4859	0.32	ng	0.00
14) 2,4,6-Tribromophenol	15.85	330	1781	0.47	ng	0.00
15) 2-Fluorobiphenyl	12.96	172	7745	0.34	ng	0.00
25) Fluoranthene-d10	19.13	212	14147	0.29	ng	0.00
29) Terphenyl-d14	19.72	244	14410	0.36	ng	0.00
Target Compounds						Qvalue
16) Acenaphthylene	14.07	152	602	0.021	ng	96
23) Phenanthrene	17.13	178	4080	0.088	ng	99
26) Fluoranthene	19.16	202	7688	0.129	ng	98
28) Pyrene	19.52	202	8094	0.137	ng	97
30) Benzo(a)anthracene	21.27	228	3937	0.062	ng	# 91
31) Chrysene	21.32	228	5113	0.076	ng	94
32) Bis(2-ethylhexyl)phthalate	21.18	149	2076	0.082	ng	98
33) Indeno(1,2,3-cd)pyrene	25.82	276	3128	0.032	ng	96
35) Benzo(b)fluoranthene	22.87	252	6170	0.082	ng	# 64
36) Benzo(k)fluoranthene	22.91	252	1973m	0.027	ng	
37) Benzo(a)pyrene	23.44	252	4282	0.062	ng	# 71
39) Benzo(a,h,i)perylene	26.52	276	3245	0.040	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032119\
Data File : BN004871.D
Acq On : 22 Mar 2019 03:38
Operator : JU/SJ
Sample : K2013-07
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PST-026-B027-031819

Manual Integrations
APPROVED

Sohil
3/22/2019 7:16:37 PM

Quant Time: Mar 22 04:57:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

