

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032119\
 Data File : BN004879.D
 Acq On : 22 Mar 2019 08:20
 Operator : JU/SJ
 Sample : K2013-10
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PST-026-B118-031819

Manual Integrations
 APPROVED

Sohil
 3/25/2019 8:52:32 AM

Quant Time: Mar 22 22:25:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 20 16:06:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.70	152	2358	0.40	ng	0.00
7) Naphthalene-d8	10.48	136	10866	0.40	ng	-0.01
13) Acenaphthene-d10	14.33	164	6931	0.40	ng	-0.01
19) Phenanthrene-d10	17.09	188	17657	0.40	ng	0.00
27) Chrysene-d12	21.29	240	23389	0.40	ng	0.00
34) Perylene-d12	23.54	264	25545	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.29	112	3180	0.50	ng	0.00
5) Phenol-d6	6.87	99	4381	0.55	ng	0.00
8) Nitrobenzene-d5	8.86	82	2885	0.40	ng	-0.01
11) 2-Methylnaphthalene-d10	12.08	152	7486	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.84	330	2760	0.58	ng	-0.01
15) 2-Fluorobiphenyl	12.95	172	11927	0.42	ng	-0.01
25) Fluoranthene-d10	19.13	212	22359	0.38	ng	0.00
29) Terphenyl-d14	19.72	244	23805	0.50	ng	0.00
Target Compounds						Qvalue
16) Acenaphthylene	14.07	152	1166	0.033	ng	99
23) Phenanthrene	17.13	178	7369	0.129	ng	99
26) Fluoranthene	19.16	202	11468	0.157	ng	# 96
28) Pyrene	19.52	202	13600	0.191	ng	97
30) Benzo(a)anthracene	21.28	228	6475	0.084	ng	91
31) Chrysene	21.33	228	7306	0.090	ng	95
32) Bis(2-ethylhexyl)phthalate	21.18	149	4788	0.156	ng	98
33) Indeno(1,2,3-cd)pyrene	25.82	276	4423	0.038	ng	99
35) Benzo(b)fluoranthene	22.87	252	8577	0.093	ng	# 72
36) Benzo(k)fluoranthene	22.91	252	2847m	0.031	ng	
37) Benzo(a)pyrene	23.45	252	6455	0.076	ng	# 81
39) Benzo(a,h,i)perylene	26.52	276	4767	0.048	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN032119\
Data File : BN004879.D
Acq On : 22 Mar 2019 08:20
Operator : JU/SJ
Sample : K2013-10
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PST-026-B118-031819

Manual Integrations
APPROVED

Sohil
3/25/2019 8:52:32 AM

Quant Time: Mar 22 22:25:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN032019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 20 16:06:20 2019
Response via : Initial Calibration

